

Amendment No. 4

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry No. HITES /PCD/PMSSY-IV/05/RTP/19-20 dated 06-07-2019; Amendment No. 1 dated 08-07-2019, Amendment No. 2 dated 13-08-2019 and Amendment No. 3 dated 06-09-2019.

Following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION – VII
TECHNICAL SPECIFICATIONS

Schedule No. 01 -Advanced High Energy Linear Accelerator		
Para	TENDER SPECIFICATION	READ AS
24 Pg 62	Server (backup/restore) – latest HW/SW and upgradable for next 10 yrs.	Server (backup/restore) – latest HW/SW and software & hardware update during warranty & CMC period
	RADIATION THERAPY BEAM ANALYZER	
Pg 63	All components comply with national and international regulations and safety rules. All components of the system; all available options are controlled by the same software that runs under Microsoft Windows of latest version of Windows 2000 and Windows XP. The system should suitable to measure pulsed radiation with fluctuation dose rate Ion Chamber:	All components comply with national and international regulations and safety rules. All components of the system; all available options are controlled by the same software that runs under Microsoft Windows of latest version of Windows 10 . The system should suitable to measure pulsed radiation with fluctuation dose rate Ion Chamber:
Pg 65	The communication between the control unit and the computer should performed by a standard RS23; interface.	The communication between the control unit and the computer should performed by a standard RS232 or USB or Ethernet
K	EQUIPMENT CERTIFICATION	
	AERB type approval: please enclose certification. US FDA / CE approvals: please enclose certification(s). Enclose certification of calibration and inspection Log book with instruction for daily, weekly, monthly, quarterly and yearly maintenance check list.	AERB type approval certificate or AERB NOC for the quoted model should be submitted in the bid. US FDA / CE approvals: please enclose certification(s). Enclose certification of calibration and inspection Log book with instruction for daily, weekly, monthly, quarterly and yearly maintenance check list.
N Pg 78	Software and Hardware Five Years warranty of all software and hardware and networking and software service agreement till ten years. All software and hardware update should be free of cost for the purchased functions and there shall be PC hardware refresh done in 4th and 8th year respectively.	Software and Hardware Five Years warranty of all software and hardware and networking and software service agreement till ten years. All software and hardware update should be free of cost for the purchased functions during warranty & CMC period .There shall be PC hardware & software upgrade done in 4th and 8th year respectively.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.