

Amendment No. 1**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/02/18-19 dated: 18.04.2018**

The pre bid meeting of the above referred tender enquiry was held on 25.04.2018. Based on pre-bid discussions following amendments are being incorporated in the tender enquiry document.

Section VII
Technical Specifications

Schedule No. 04 - Electronic Stimulator/Electronic muscle stimulator (Rfx no. 3000002814)		
Tender Page & Para	TENDER SPECIFICATION	READ AS
		Added Para: EMG recording (2 channels) should be available.

Schedule No. 07 -Thermal aesthesiometer - Digital (Rfx no. 3000002817)		
Tender Page & Para	TENDER SPECIFICATION	READ AS
		Added Para: Suitable software should be provided.
		Added Para: This item is for Human application

Schedule No. 09 - Fat Extraction Instrument (Rfx no. 3000002819)		
Tender Page & Para	TENDER SPECIFICATION	READ AS
Page 45 Point no. 1	Extraction Apparatus, fat, complete	Extraction Apparatus, fat, complete (Six samples)
Page 45 Point no. 7	Red pilot light	Red pilot light / Heater Indication : LED
Page 45 Point no. 8	Metal condensers with Type 304 stainless steel heads with automatic pressure-release valves	Metal/Borosilicate condensers with Type 304 stainless steel heads / Teflon heads with automatic pressure-release valves

Page 45 Point no. 12 vii	Accessories 9 sample tubes	Accessories 9 sample tubes / beakers
Page 45 Point no. 12 viii	Accessories 6 support stirrups and	Accessories 6 support stirrups or 6 Thimble Holder stand to hold 6 thimbles
Page 45 Point no. 12 ix	Accessories 9 borosilicate glass reclaiming tubes.	Accessories 9 borosilicate glass reclaiming tubes or flask

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**