

**Amendment No. 1**

**Date: 11/11/2014**

**Sub: Amendment to the Tender Enquiry Document**

**Ref: NIT No.: HLL/PCD/GNCTD/18/JSSH/14-15 dated 03/11/2014**

The following changes have been incorporated in the referred NIT.

**SECTION VII  
Technical Specification**

The following specification (enclosed in the next page) is to be read in continuation to the specification for **Modular Operation Theatres** bearing tender ID: 2014\_HFWD\_69293\_1. This page was inadvertently omitted in the tender enquiry document.

	attached as required at site.		
8.	Supplying, Installation, Testing & Commissioning of <b>imported HD Monitor with Arm</b> . complete as per technical specification attached as required at site.	2	Nos.
9.	<b>Imported Ceiling suspended horizontal Bridge</b> : - - as per technical specification attached as required at site.	2	Nos.
10.	Supplying, Installation, Testing & Commissioning of <b>imported clean room luminaries with RGB with frame</b> - 6 nos. for each OT etc. complete as per technical specification attached as required at site.	12	Nos.
11.	Supplying, Installation, Testing & Commissioning of <b>conductive Tile Flooring inside OT</b> complete as per technical specification attached as required at site.	80	Sqm
12.	Supplying, Installation, Testing & Commissioning of <b>self levelling compound</b> complete as per technical specification attached as required at site.	80	Sqm
13.	Supplying, Installation, Testing & Commissioning of <b>pressure relief dampers inside OT</b> complete as per technical specification attached as required at site.	2	Nos.
14.	Supplying, Installation, Testing & Commissioning of <b>X-Ray viewing Screen inside each OT</b> complete as per technical specification attached as required at site.	2	Nos.
15.	Supplying, Installation, Testing & Commissioning of <b>Distribution board</b> outside OT. complete as per technical specification attached as required at site.	2	Nos.
16.	Supplying, Installation, Testing & Commissioning of <b>Surgeon Control Panel inside OT</b>	2	Nos.
17.	Supplying, Installation, Testing & Commissioning of <b>Hatch Box inside</b>	2	Nos.
18.	Supplying, Installation, Testing & Commissioning of <b>Stainless Steel Scrub Sink Two Bay</b> outside OT	2	Nos.
20.	<b>Electrical wiring, conducting with fixtures inside the OT only.</b>	2	set
21.	<b>Isolation Transformer 25KVA</b>	2	Nos
22.	<b>Pre-Operation Room and Post Operation Room-</b>	01	Set
23.	<b>AHU: - (Optional)</b>	01	Set
24.	<b>50 KVA On - Line UPS</b>	01	No.

  
 Dr. RAJEEV UPPAL  
 Director-Professor & Head  
 Department of Anaesthesiology  
 & Intensive Care  
 G.B. Pant Hospital, N.Delhi-02

All consumables of fixtures  
 shall cover under warranty and cmc  
  
 H. Dabhi

All other contents of the tender enquiry including terms & conditions remain unaltered.

**Note:** Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids