

Amendment No.5**Date: 12/10/2016****Sub: Amendment No.5 to the Tender Enquiry Document****Ref: NIT No.: HLL/PCD/PMSSY-II/11/16-17 dated: 22.08.2016.**

The following changes are made to the Tender enquiry document.

SECTION I**NOTICE INVITING GLOBAL e-TENDERS (NIT)****(i)The due dates for closing and opening of the items (i.e. Sch. 4, 8, 9, 10, 15, 16, 21, 23, 24, 25, 28, 29, 30, 31 and 32) are as under:****Table: 1**

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
4	3000001467	Computerized cardiopulmonary exercise testing system with treadmill for humans	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
8	3000001471	Fiberoptic Phototherapy Lamp	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
9	3000001472	Labour Bed cum Labour Table	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
10	3000001473	Non-invasive ventilator	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
15	3000001480	Defibrillator	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
16	3000001481	Phaco Emulsification System	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
21	3000001487	Vascular Doppler	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
23	3000001489	Surgical Operating Microscope for Neurosurgery	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
24	3000001491	E.N.T. Operating Microscope & Video Camera Unit	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
25	3000001492	Neonatal Incubator	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
28	3000001495	800mA X-Ray unit with Single Detector (U/C Arm)	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
29	3000001496	Video Endoscope unit with NBI/HD+Video with Upper GI	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
30	3000001510	HPLC System With Chromatographic Workstation	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
31	3000001511	Vacuum Assisted Tissue Processor	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00
32	3000001512	IABP (Intra Aortic Balloon Pump) - High End	20.10.2016 18:00	21.10.2016 11:30	21.10.2016 14:00

Note: All the bidders are required to refer amendment No: 3 (regarding startups) & 4 (Regarding Changes in Technical specifications after prebid meeting. There are no changes in technical specifications for Sch: 4, 8, 9, 10, 15, 16, 21, 28, 30, 31, and 32)

(ii) The revised due dates for closing and opening of the item (i.e. Sch. 26) is as under:

Table: 2

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
26	3000001493	Vats Set	25.10.2016 18:00	26.10.2016 11:30	26.10.2016 14:00

Note: The changes in technical Specifications of Sch No. 26 are at Pg. No. 4 of this document

(iii)The revised due dates for closing and opening of the rest of items are as under:

Table: 3

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
1	3000001464	Photo – slit lamp with applanation tonometer	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
2	3000001465	Ophthalmic ND: YAG Laser- 1064 nm	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
3	3000001466	Pneumatic Drill Machine for Neurosurgery	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
5	3000001468	Flexible Cysto-Nephroscope (High End)	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
6	3000001469	PCNL Set	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
7	3000001470	Uretrorenoscope	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
11	3000001474	ICP Monitor	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
12	3000001475	Stretcher Trolley	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
13	3000001476	Colonoscope with Videoscope	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
14	3000001477	Upper GI Endoscope	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
17	3000001482	High Frequency X Ray Unit 800mA	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
18	3000001483	ENT Treatment Unit	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
19	3000001484	Operating Hysteroscope with accessories	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
20	3000001485	Complete Micro Motor System for Trauma Care	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00
22	3000001488	Mobile Surgical C-Arm	09.11.2016 18:00	10.11.2016 11:30	10.11.2016 14:00

Note: The changes after pre-bid meeting with respect to technical specifications of items in Table: 3 will be published in due course.

**Section – VII
Technical Specifications**

**Item SI No. 26
VATS Equipment specifications**

1. Existing Para 3.1.1: Road Lens full HD Telescope 30° , enlarged view, diameter 10mm, length 31 cm, autoclavable, fiber optic light transmission incorporated- Qty 1.

Read as: Road Lens full HD Telescope 30° , enlarged view, diameter 10mm, length 31 cm, autoclavable, fiber optic light transmission incorporated **compactible to HD Camera** - Qty 1

2. Existing Para 3.1.3: Telescope 10 mm, length 32 cm, variable direction of view from 0° - 120° , twisting controller to select the desired view of direction, fiber optic light transmission incorporated- Qty 1 Direction of view should be zero degree.

Read as: Deleted

3. Existing Para 3.2.4: Controls to freeze images enhance a portion of frozen image (zoom & post-processing).

Read as: Deleted.

4. Existing Para 3.2.5: Operates on Xenon lamp **with battery backup of at least 45 minutes.**

Read as: Operates on Xenon lamp.

5. Existing Para 3.3: Light outlets- **3.**

Read as: Light outlets- **1**

6. Added Para: Complete system should be European CE or USFDA approved.

7. Added Para: Suitable online UPS for complete system with at least 45 min backup.

All other contents of the tender enquiry including terms & conditions remain unaltered.