

PREBID QUERIES/CLARIFICATIONS ASKED BY BIDDERS AND ITS RECOMMENDATIONS/REPLIES

NE ref: HITES/PCD/NCI-AIIMS/33/18-19 dated 26.09.2018
 Item: ETO Sterilizer (Rfx 3000003411)
 Prebid Meeting held on: 11/10/2018

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Bidder	Queries/clarifications asked by Bidders	Recommendations/ replies by the Committee
	Technical Terms				
1	Point 2 Page 49	The bidder should have supplied and installed at least 1 (one) similar equipment meeting major parameters of technical specification during last Five years from the date of Bid Opening.	Med fresh Pvt Ltd	The bidder should have supplied and installed at least 1 (one) similar equipment (ETO/Plasma /LTSF) meeting major parameters of technical specification during last Five years from the date of Bid Opening.	No change
2	Point 16, Page 49	Gas cartridges should be EPA or European CE certified.			To be amended as:- DELETED
3	Point 25, Page 46	Should be US FDA/European CE certified.			To be amended as:- DELETED
4	Point 30, Page 46	A observation glass window should be present in ETO room to monitor the process.			To be amended as:- DELETED

