

Prebidqueries/clarifications asked by Firms and its recommendations/replies (ref: prebid meeting dated: 07/08/2018)
NIB ref: HITES/PCD/NCI-AIIMS/23/18-19 dated 26.07.2018


Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
TECHNICAL ISSUES					
Rfx No. 3000003180, Abdominal Instruments Set					
1	Page 46, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
2			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
3			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
4	Page 46 ,point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure Surgical	Delete	Deleted
5	Page 46 ,point 4	4. All needle holders &Metzenbaum scissors should be Tungston Carbide	Cure Surgical	All needle holders &Metzenbaum scissors should be Tungston Carbide Inserted	To be Amended as: All needle holders &Metzenbaum scissors should be Tungston Carbide Insert
6	Page 46 ,point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory	Cure Surgical	Need the list of instrument and 6 weeks time	Added Para :- Items mentioned at S..No 11,34a,34f,34m,34o are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
		performance will be taken into account for technical approval.			into account for technical approval.
7	Page 46 , BOQ point 1	SS tray large	Cure Surgical	Delete	Deleted
	Page 46 , BOQ point 2	Towel Clips Large	As per pre bid discussion		Deleted
	Page 46 , BOQ point 3	Towel Clips medium	As per pre bid discussion		Deleted
	Page 46 , BOQ point 4	Towel Clips small	As per pre bid discussion		Deleted
8	Page 47 , BOQ point 20	SS bowl large	Cure Surgical	Delete	Deleted
9	Page 47 , BOQ point 21	SS bowl medium	Cure Surgical	Delete	Deleted
10	Page 47 , BOQ point 22	Sponge holder	Cure Surgical	229mm	Deleted
9	Page 47 , BOQ point 23	Suction tip	Cure Surgical	180mm	To be Amended as: Suction tip 18-20cm
	Page 47,BOQ point 24	Self retaining retractor with 3rd blade ong	Link Medical	More detailed specifications are required	To be Amended as: Gosset or Similar type three blade self retaining abdominal wall retractor 17-20cm








Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	Page 47,BOQ point 25	Self retaining retractor with 3rd blade small	Link Medical	More detailed specifications are required	To be Amended as: Gosset or Similar type three blade self retaining abdominal wall retractor 12-15cm
	Page 47,BOQ point 29	Deaver's retractor large	Link Medical	More detailed specifications are required	To be Amended as: Deaver's retractor 33-36cm
	Page 47,BOQ point 30	Deaver's retractor medium	Link Medical	More detailed specifications are required	To be Amended as: Deaver's retractor 24-26cm
10	Page 47 , BOQ point 34	In Kidney Tray (Big)	Cure Surgical	Delete	Deleted
11	Page 47 , BOQ point 34.d	Debaky's vascular forceps	Cure Surgical	180mm	To be Amended as: Debaky's vascular forceps 16-18cm
			Link Medical	More detailed specifications are required	
Rfx No. 3000003181, Thoracotomy Instruments Set					
1	Page 49,point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified 
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified.	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 49 ,point	The Instr. ments should be of	Cure Surgical	Delete	Deleted





Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	3	medical grade stainless steel (SS 304/SS 316), rust free and have aniglare property			
3	Page 49 ,point 4	4. All needle holders &Metzenbaum scissors should be Tungston Carbide	Cure Surgical	All needle holders &Metzenbaum scissors should be Tungston Carbide Inserted	To be Amended as: All needle holders &Metzenbaum scissors should be Tungston Carbide Insert
4	Page 49 ,point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure Surgical	Need the list of instrument and 6 weeks time	Added para :- Items mentioned at S.No 10,13,23,31,34,37,44,50 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval
5	Page 49, BOQ point 1	SS tray large	Cure Surgical	Delete	Deleted
	Page 49, BOQ point 2	Towel clips large	As per prebid discussion		Deleted
	Page 49, BOQ point 3	Towel clips medium	As per prebid discussion		Deleted
	Page 49, BOQ point 4	Towel clips small	As per prebid discussion		Deleted
6	Page 49, BOQ point 16	SS bowl large	Cure Surgical	Delete	Deleted

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
7	Page 50, BOQ point.17	SS bowl medium	Cure Surgical	Delete	Deleted
8	Page 50, BOQ point.18	Sponge holder	Cure Surgical	229mm	Deleted
9	Page 50, BOQ point.19	Suction tip	Cure Surgical	180mm	To be Amended as: Suction tip 18-20cm
	Page 50, BOQ point.22	Chest wall retractor with key large blade	Link Medical	More detailed specifications of items are required.	Amended as: Caspar or Similar large blade 15-17cm chest wall retractor with key
	Page 50, BOQ point.23	Chest wall retractor with key normal blade	Link medical	More detailed specifications of items are required	Amended as: Caspar or Similar medium size blade 11-13cm chest wall retractor with key
	Page 50, BOQ point.24	Chest wall retractor with key double hook	Link medical	More detailed specifications of items are required	Caspar or Similar Chest wall retractor with key double hook with 11-13cm blade
10	Page 50, BOQ point.25		Cure Surgical	254mm	To be Amended as: Rib shear 24-26cm
			Link Medical	More detailed specifications of items are required	
	Page 50, BOQ point.26	Sternal and rib punch	Link medical	More detailed specifications of items are required	No Change
	Page 50, BOQ point.28	Rib raspatory left	Link medical	More detailed specifications of items are required	To be Amended as: Rib raspatory left 20-22cm

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	Page 50, BOQ point.29	Rib raspatory right	Lirk medical	More detailed specifications of items are required	To be Amended as: Rib raspatory right 20-22cm
11	Page 50, BOQ point.30	Periosteum elevator	Cure Surgical	165mm	To be Amended as: Periosteum elevator 16-18cm
			Lirk medical	More detailed specifications of items are required	
15	Page 50, BOQ point.31	Lung holding forceps	Cure Surgical	197mm	To be Amended as: Lung holding forceps 19-21cm
			Lirk Medical	More detailed specifications of items are required	
16	Page 50, BOQ point.32	Semb dissecting forceps	Cure Surgical	241mm	To be Amended as: Semb dissecting forceps 24-26cm
			Lirk Medical	More detailed specifications of items are required	
	Page 50, BOQ point.33	Nelson lung dissecting forceps	Lirk Medical	More detailed specifications of items are required	To be Amended as: Nelson lung grasping forceps 20-22cm
17	Page 50, BOQ point.35	Debakey's traumatic non toothed dissecting forceps	Cure Surgical	180mm	To be Amended as: Debakey's traumatic non toothed dissecting forceps 17-19cm
			Lirk Medical	More detailed specifications of items are required	
18	Page 50, BOQ point.36	Davidson scapular retractor	Cure Surgical	152mm	To be Amended as: Davidson scapular retractor 15-17cm
			Lirk Medical	More detailed specifications of items are	

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
				required	
19	Page 50, BOQ point..42	Bozemom needle holder	Cure Surgical	200mm	To be Amended as: Bozemom needle holder 19-21cm
			Link Medical	More detailed specifications of items are required	
20	Page 51, BOQ point..53	Sternal needle holder	Cure Surgical	150mm	To be Amended as: Sternal needle holder 14-16cm
			Link Medical	More detailed specifications of items are required	
21	Page 51, BOQ point..54	Pliar	Cure Surgical	140mm	To be Amended as: Pliar 14-16cm
			Link Medical	More detailed specifications of items are required	
22	Page 51, BOQ point..55	Wire cutter	Cure Surgical	180mm	To be Amended as: Wire cutter 17-19cm
			Link Medical	More detailed specifications of items are required	
	Page 51, BOQ point..56	Needle guard	Link Medical	More detailed specifications of items are required	No Change
Rfx No. 3000003182, Large Bone Set					
1	page 52 , point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA &	

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
				European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 52, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	Deleted
3	Page 52, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	Page 52, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	Added para :- Items mentioned at S..No 9,12 19,22,32,34 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 52 , BOQ point 1	SS tray big	Cure surgical	Delete	Deleted
	Page 52 , BOQ point 2	Amputation knife medium	Link Medical	More detailed specifications of items are required	To be Amended as: Amputation knife liston/catlin 22-24cm

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	Page 52 , BOQ point 3	Amputation knife small	Link Medical	More detailed specifications of items are required	To be Amended as: Amputation knife liston/catlin/langenbeck 13-16cm
6	Point 52,BOQ point 4	Bone cutter angled	Cure surgical	150mm	To be Amended as: Bone cutter angled 14-16cm
			Link Medical	More detailed specifications of items are required	
7	Point 52,BOQ point 5	Bone cutter double action	Cure surgical	260mm	To be Amended as: Bone cutter double action 25-27cm
			Link Medical	More detailed specifications of items are required	
	Point 52,BOQ point 6	Bone file big	Link Medical	More detailed specifications of items are required	To be Amended as: Bone file 22-24cm
	Point 52,BOQ point 7	Bone file small	Link Medical	More detailed specifications of items are required	To be Amended as: Bone file 16-18cm
8	Point 53,BOQ point 13	Bone lever lancer type plain	Cure surgical	160mm	To be Amended as: Bone lever lancer type plain 15-17cm
			Link Medical	More detailed specifications of items are required	
9	Point 53,BOQ point 14	Bone lever lancer type serrated	Cure surgical	160mm	To be Amended as: Bone lever lancer type serrated 15-17cm
			Link Medical	More detailed specifications of items are required	

JA

Adhal

SP

See

AB

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
10	Point 53,BOQ point 15	Bone nibbler double action curved	Cure surgical	230mm	To be Amended as: Bone nibbler double action curved 22-24cm
			Link Medical	More detailed specifications of items are required	
11	Point 53,BOQ point 16	Bone nibbler double action straight	Cure surgical	230mm	To be Amended as: Bone nibbler double action straight 22-24cm
			Link Medical	More detailed specifications of items are required	
	Point 53,BOQ point 17	Boyle mouth gag with 5 plates	Link Medical	More detailed specifications of items are required	To be Amended as: MOLT or Similar Mouth gas 10-12cm & 14-16cm – 1 each size
	Point 53,BOQ point 21	Curette medium	Link Medical	More detailed specifications of items are required	To be Amended as: Curette medium 12-15cm
	Point 53,BOQ point 22	Curette with handle big	Link Medical	More detailed specifications of items are required	To be Amended as: Curette with handle big 17-20
	Point 53,BOQ point 24	Hammer/Mallet medium	Link Medical	More detailed specifications of items are required	To be Amended as: Hammer/Mallet small 22-24cm length, head dia 25-27 on one side & 13-15 on other side
	Point 53,BOQ point 25	Hammer/Mallet large	Link Medical	More detailed specifications of items are required	To be Amended as: Hammer/Mallet small 22-24cm length, head dia 38-40 on one side & 28-30cm on other side







Sl. No.	Ref. to the Bidding Document	Specification/classes in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	Point 53, BOQ point 26	Hand saw medium	Link Medical	More detailed specifications of items are required	To be Amended as: Charriere or Similar had saw length 24-27cm
	Point 53, BOQ point 27	Hand saw large	Link Medical	More detailed specifications of items are required	To be Amended as: Satterlee or Similar had saw length 30-33cm
12	Point 53, BOQ point 28	HM wire cutter	Cure surgical	178mm	To be Amended as: HM wire cutter 16-18cm
			Link Medical	More detailed specifications of items are required	
	Point 53, BOQ point 29	Mouth gag (15 cm)	Link Medical	More detailed specifications of items are required	To be Amended as: MOLT or Similar Mouth gas 10-14-16cm – 1 no.
13	Point 53, BOQ point 34	Periosteum elevator	Cure surgical	165mm	To be Amended as: Periosteum elevator 16-18cm
			Link Medical	More detailed specifications of items are required	
Rfx No. 3000003183, Head & Neck and Orbitotomy Bone Set					
1	page 54, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the	To be Amended as: The Instruments Should be US-FDA or

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
				specification in addition to FDA/CE/BIS.	European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 54, point 3	The Instruments should be of medical grade stainless steel (SS 304/CS 316), rust free and have antiglare property	Cure surgical	pl delete	Deleted
3	Page 54, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	Page 54, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as: Items mentioned at S.No 8,11,15,19,29,60 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 54 ,	SS tray big	cure surgical	Delete	Deleted

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	BOQ point 1				
	Page 54 , BOQ point 2	Amutation krifa medium	Link Medical	More detailed specifications of items are required	To be Amended as: Amputation knife liston/catlin 22-24cm
	Page 54 , BOQ point 3	Amutation krifa small	Link Medical	More detailed specifications of items are required	To be Amended as: Amputation knife liston/catlin/langenbeck 13-16cm
	Page 54 , BOQ point 4	Bone cutter angled	Link Medical	More detailed specifications of items are required	To be Amended as: Bone cutter angled 14-16cm
	Page 54 , BOQ point 5	Bone cutter double action	Link Medical	More detailed specifications of items are required	To be Amended as: Bone cutter double action 25-27cm
	Page 54 , BOQ point 6	Bone file big	Link Medical	More detailed specifications of items are required	To be Amended as: Bone file 22-24cm
	Page 54 , BOQ point 7	Bone file small	Link Medical	More detailed specifications of items are required	To be Amended as: Bone file 16-18cm
6	Page 55 , BOQ point 12	Bone lever lencer type plain	cure surgical	160mm	To be Amended as: Bone lever lencer type plain 15-17cm.
			Link Medical	More detailed specifications of items are required	
7	Page 55 ,	Bone lever lencer type serrated	cure surgical	160mm	To be Amended as: Bone lever lencer

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	BOQ point 13		Link Medical	More detailed specifications of items are required	type serrated 15-17cm
8	Page 55 , BOQ point 14	Bone nibbler double action curved	cure surgical	230mm	To be Amended as: Bone nibbler double action curved 22-24cm
			Link Medical	More detailed specifications of items are required	
9	Page 55 , BOQ point 15	Bone nibbler double action straight	cure surgical	230mm	To be Amended as: Bone nibbler double action straight 22-24cm
			Link Medical	More detailed specifications of items are required	
	Page 55 , BOQ point 16	Boyle mouth gag with 5 plates	Link Medical	More detailed specifications of items are required	MOLT or Similar Mouth gas 10-12cm & 14-16cm – 1 each
	Page 55 , BOQ point 20	Curette sma	Link Medical	More detailed specifications of items are required	To be Amended as: Curette medium 12-15cm
	Page 55 , BOQ point 21	Curette with handle medium	Link Medical	More detailed specifications of items are required	To be Amended as: Curette with handle big 17-20
	Page 55 , BOQ point 23	Hammer/Mallet medium	Link Medical	More detailed specifications of items are required	To be Amended as: Hammer/Mallet small 22-24cm length, head dia 25-27 on one side & 13-15 on other side
	Page 55 , BOQ point 24	Hammer/Mallet small	Link Medical	More detailed specifications of items are required	To be Amended as: Hammer/Mallet small 22-24cm length, head dia 38-40 on one side & 28-30cm

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
					on other side
	Page 55 , BOQ point 25	Hand saw medium	Link Medical	More detailed specifications of items are required	To be Amended as: Charriere or Similar had saw length 24-27cm
	Page 55 , BOQ point 26	Hand saw small	Link Medical	More detailed specifications of items are required	To be Amended as: Langenbeck or Similar had saw length 20-22 cm
10	Page 55 , BOQ point 27	HM wire cutter	cure surgical	178mm	To be Amended as: HM wire cutter 17-19cm
			Link Medical	More detailed specifications of items are required	
	Page 55 , BOQ point 28	Mouth gag (15 cm)	Link Medical	More detailed specifications of items are required	To be Amended as: MOLT or Similar Mouth gas 14-16cm – 1 no.
11	Page 55 , BOQ point 33	Periosteum elevator	cure surgical	165mm	To be Amended as: Periosteum elevator 16-18cm
			Link Medical	More detailed specifications of items are required	
	Page 55 , BOQ point 36	Tooth extractor forceps	Link Medical	More detailed specifications of items are required	To be Amended as: Tooth extractor set for different teeth
15	Page 55 ,	Malliable retractors set	cure surgical	Malliable retractors set of 5 200mm	Deleted

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	BOQ point 37		Link Medical	More detailed specifications of items are required	
	Page 55, BOQ point 38	Dental Extractor	Link Medical	More detailed specifications of items are required	Deleted
	Page 55, BOQ point 39	Mouth gag	Link Medical	More detailed specifications of items are required	To be Amended as: MOLT or Similar Mouth gas 10-12cm & 14-16cm – 1 each
	Page 55, BOQ point 40	Bone - saw, file, nibbler	Link Medical	More detailed specifications of items are required	To be Amended as: Bone saw Bone file 16-18cm Bone nibbler double action curved 14-16cm – 1 each
	Page 55, BOQ point 41	Drill set	Link Medical	More detailed specifications of items are required	Deleted
	Page 55, BOQ point 42	Malleable retractors Small	Link Medical	More detailed specifications of items are required	Deleted
	Page 55, BOQ point 43	Malleable retractors medium	Link Medical	More detailed specifications of items are required	Deleted
	Page 55, BOQ point 44	Malleable retractors large	Link Medical	More detailed specifications of items are required	Deleted
16	Page 55, BOQ point 45	Hammer mallet DF H4	cure surgical	Hammer mallet 212gm 200mm	To be Amended as: Hammer Mallet 200-250g (total weight)
			Link Medical	More detailed specifications of items are	

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
				required	18-20cm
17	Page 56 , BOQ point 46	kockers clamp BH 614	cure surgical	kockers clamp 140mm	To be Amended as: kockers clamp 14-16cm
			Link Medical	More detailed specifications of items are required	
18	Page 56 , BOQ point 47	Cheasle FK 318	cure surgical	Cheasle 190mm	Deleted
			Link Medical	More detailed specifications of items are required	
19	Page 56 , BOQ point 48	Bone Gauge straight	cure surgical	9mm 20cm	To be Amended as: Bone Gauge straight 8-10mm, 20cm
			Link Medical	More detailed specifications of items are required	
20	Page 56 , BOQ point 49	Bone Gauge curved	cure surgical	9mm 20cm	To be Amended as: Bone Gauge curved 8-10mm, 20cm
			Link Medical	More detailed specifications of items are required	
21	Page 56 , BOQ point 50	Sharp two prong retractor BT 242	cure surgical	Sharp two prong retractor 22cm	To be Amended as: Sharp two prong retractor 20-22cm
			Link Medical	More detailed specifications of items are required	
22	Page 56 , BOQ point 51	Freer bone elevator OL-165	cure surgical	Freer bone elevator 19cm	To be Amended as: Freer bone elevator 16-19cm
			Link Medical	More detailed specifications of items are required	

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
23	Page 56 , BOQ point 52	periosteal knife OF-700	cure surgical	periosteal knife	To be Amended as: periosteal knife
			Link Medical	More detailed specifications of items are required	
24	Page 56 , BOQ point 53	Perkin antrum self retaining retractor	cure surgical	Perkin antrum self retaining retractor 125mm left & right 1each	To be Amended as: Perkin antrum self retaining retractor 125mm
			Link Medical	More detailed specifications of items are required	
25	Page 56 , BOQ point 54	Langenbeck retractor	cure surgical	10 x 41 mm 8 1/4 in 210 mm	To be Amended as: Langenbeck retractor 20-22cm, 10x40mm
			Link Medical	More detailed specifications of items are required	
26	Page 56 , BOQ point 55	Hartman dressing forceps OK 242 nasal	cure surgical	Hartman dressing forceps 90mm	To be Amended as: Hartman dressing forceps 8-10cm
			Link Medical	More detailed specifications of items are required	
27	Page 56 , BOQ point 56	Butterfly probe	cure surgical	130mm	To be Amended as: Butterfly probe 12-14cm
			Link Medical	More detailed specifications of items are required	
28	Page 56 ,	Retractor small	cure surgical	160mm	To be Amended as: Retractor small 15-

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	BOQ point 57		Link Medical	More detailed specifications of items are required	17cm
29	Page 56 , BOQ point 58	Elevator big FK-133	cure surgical	Elevator 190mm	To be Amended as: Elevator big 18-20cm
			Link Medical	More detailed specifications of items are required	
30	Page 56 , BOQ point 59	Elevator small OM-307	cure surgical	Elevator small 140mm	To be Amended as: Elevator small 14-16cm
			Link Medical	More detailed specifications of items are required	
31	Page 56 , BOQ point 60	Malleable retractors (set of 5)	cure surgical	200mm	No Change
			Link Medical	More detailed specifications of items are required	
32	Page 56 , BOQ point 61	S.S.Tray (medium)	cure surgical	Delete	Deleted
			Link Medical	More detailed specifications of items are required	
Rfx No. 3000003184,Gynae Instrument Set					
1	page 57, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body	

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
				certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 57, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have anti glare property	Cure surgical	pl delete	Deleted
3	Page 57, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	Page 57, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	Added para :- Items mentioned at S..No 3,9,10,14,15 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 57 , BOQ point 1	SE tray, medium	cure surgical	Delete	To be Amended as: Deleted
6	Page 57 , BOQ point 2	Sims' vaginal speculum big 409	cure surgical	Sims' vaginal speculum size1	To be Amended as: Sims' vaginal speculum size1
7	Page 57 , BOQ point 3	Sims' vaginal speculum medium 403	cure surgical	Sims' vaginal speculum size 2	To be Amended as: Sims' vaginal speculum size 2

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
8	Page 57 , BOQ point 4	Sims' vaginal speculum small 407	cure surgical	Sims' vaginal speculum size 3	To be Amended as: Sims' vaginal speculum size 3
9	Page 57 , BOQ point 5	Sponge holder	cure surgical	240mm	Deleted
10	Page 57 , BOQ point 6	Sims uterine probe malleable curved	cure surgical	32cm	To be Amended as: Sims uterine probe malleable curved 30-32cm
			Link Medical	More detailed specifications of items are required	
	Page 57 , BOQ point 7	Bladder sound	Link Medical	More detailed specifications of items are required	To be amended as: Bladder sound 3mm, 4mm, 5mm, 6mm & 7mm – 1 no. each
11	Page 57 , BOQ point 8	Endometrial biopsy curette	cure surgical	6mm 24 cm	To be Amended as: Endometrial biopsy curette 22-24cm
			Link Medical	More detailed specifications of items are required	
12	Page 57 , BOQ point 9	Hysterectomy clamp St	cure surgical	21cm	To be Amended as: Hysterectomy clamp St 20-22cm
			Link Medical	More detailed specifications of items are required	
13	Page 57 , BOQ point 10	Cervical punch biopsy forceps shaft length 31	cure surgical	Cervical punch biopsy forceps shaft length 22cm	To be Amended as: Cervical punch biopsy forceps shaft length 20-22cm

Handwritten signatures and initials in blue ink, including a large signature on the left, a signature in the center, and several initials on the right.

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
			Link Medical	More detailed specifications of items are required	
		Female pelvic Set			
14	Page 58, BOQ point 14	Mangeshkar retractor	Cure surgical	not clear	To be Amended as: Mangeshkar Or Similar retractor
			Link Medical	More detailed specifications of items are required	
Rfx No. 3000003185,Thyroidectomy Instruments set					
1	page 59, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 59, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	To be Amended as: Deleted
3	Page 59, point	All needle holders & Metzenbaum scissors should	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide	To be Amended as: All needle holders & Metzenbaum scissors should be







Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	4	be Tungsten Carbide coated		Insert	Tungsten Carbide Insert
4	Page 59, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	Added para :- items mentioned at S..No 10,15,27g,27h,27k are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval
5	Page 59 , BOQ point 1	SS tray medium	cure surgical	Delete	Deleted
	Page 59 , BOQ point 2	Towel clips large	As per prebid discussion		Deleted
	Page 59 , BOQ point 3	Towel clips medium			Deleted
	Page 59 , BOQ point 4	Towel clips small			Deleted
6	Page 60 , BOQ point 14	Kelly's clamp	Cure surgical	140mm	To be Amended as: Kelly's clamp 12-14cm
	Page 60 , BOQ point 15	Kocher's clamp curved (medium)	Link Medical	More detailed specifications of items are required	To be Amended as: kockers clamp curved 14-16cm
	Page 60 , BOQ point 16	Kocher's curved	Link Medical	More detailed specifications of items are required	Deleted

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	Page 60 , BOQ point 19	Langenbeck retractor medium	Link Medical	More detailed specifications of items are required	To be Amended as: Langenbeck retractor 20-22cm, 10-11x40mm
	Page 60 , BOQ point 20	Langenbeck retractor small	Link Medical	More detailed specifications of items are required	To be Amended as: Langenbeck retractor 20-22cm, 15-16x30mm
	Page 60 , BOQ point 21	S.S.Bowl (large)	As per bid discussion		Deleted
	Page 60 , BOQ point 22	S.S.Bowl (medium)	As per bid discussion		Deleted
7	Page 60 , BOQ point 23	Sponge holder		240mm	To be Amended as: Sponge holder 22-24cm
8	Page 60 , BOQ point 27	In Kidney tray (big)		Delete	Deleted
	Page 60 , BOQ point	Adson vascular forceps	Link Medical	More detailed specifications of items are required	To be Amended as: Adson vascular forceps medium





[Handwritten signatures and initials in blue ink]



Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	27.g				
	Page 60 , BOQ point 27.k	Metzenbaum scissors small			To be Amended as: Metzenbaum scissors medium
Rfx No. 3000003186, General Surgery Instruments Set					
1	page 61, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 61, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	To be Amended as: Deleted
3	Page 61, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	Page 61, point	Demonstration of sample Instruments is mandatory and	Cure surgical	need instrument list and six week time	Added para :- items mentioned at S.No 11,27b,27e,27g,27k,27n, are to be

Handwritten signatures and initials in blue ink, including a large signature on the left and several smaller ones on the right.


Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	8	feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.			demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 61, BOQ point 1	SS tray large	cure surgical	Delete	Deleted
	Page 61, BOQ point 2	Towel clips large			Deleted
	Page 61, BOQ point 3	Towel clips medium			Deleted
	Page 61, BOQ point 4	Towel clips small			Deleted
6	Page 62 , BOQ point 20	SS bowl large	cure surgical	Delete	Deleted
7	Page 62 , BOQ point 21	SS bowl medium	cure surgical	Delete	Deleted
8	Page 62 , BOQ point 22	Sponge holder	cure surgical	240mm	Deleted
9	Page 62 , BOQ point 23	Suction tip	cure surgical	180mm	To be Amended as: Suction tip 16-18cm
10	Page 62 ,	In Kidney Tray (Big)	cure surgical	Delete	Deleted

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	BOQ point 27				
11	Page 62 , BOQ point 27.n	Debakeys's vascular forceps	cure surgical	200mm	To be Amended as: Debakeys's vascular forceps medium
			Link Medical	More detailed specifications of items are required	
Rfx No. 3000003187, Mastectomy Instruments Set					
1	page 63, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	Page 63, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antistatic property	Cure surgical	pl delete	Deleted
3	Page 63, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
4	Page 63, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as :- Items mentioned at S..No 11,17,25b,25d,25f,25l are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 63, BOQ point 1	SS tray medium	cure surgical	Delete	Deleted
	Page 63, BOQ point 2	Towel clips large	As per prebid discussion		Deleted
	Page 63, BOQ point 3	Towel clips medium	As per prebid discussion		Deleted
	Page 63, BOQ point 4	Towel clips small	As per prebid discussion		Deleted
6	Page 64 , BOQ point 22	S S Bowl medium	cure surgical	Delete	Deleted
7	Page 64 , BOQ point 23	S S Bowl small	cure surgical	Delete	Deleted
8	Page 64 , BOQ point 24	Sponge holder	cure surgical	240mm	Deleted
9	Page 64 , BOQ point 25	In kidney tray (big)	cure surgical	Delete	Deleted



Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
10	Page 64 , BOQ point 25.h	Debakey's vascular forceps	cure surgical	200mm	To be Amended as: Debakey's vascular forceps 18-20cm
			Link Medical	More detailed specifications of items are required	
11	Page 64 , BOQ point 25.o	Skin hooks	cure surgical	150mm	To be Amended as: Skin hooks 14-16cm
			Link Medical	More detailed specifications of items are required	
Rfx No. 3000003188, Clip Applicator Set					
1	page 65, point 2	The instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be Amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	page 65, point 3	The instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have ant glare property	Cure surgical	pl delete	Deleted
3	page 65,	All needle holders & Metzenbaum scissors should	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide	To be Amended as: All needle holders & Metzenbaum scissors should be

Handwritten signatures in blue ink, including a large signature on the left, a signature in the middle, and several smaller signatures on the right.

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
	point 4	be Tungsten Carbide coated		Insert	Tungsten Carbide Insert
4	Page 65, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as: Demonstration of medium & medium large clip applicators to be given and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 65, BOQ point 1	SS tray large	cure surgical	Delete	Deleted
	Page 65, BOQ point 2	Towel clips large			Deleted
	Page 65, BOQ point 3	Towel clips medium			Deleted
	Page 65, BOQ point 4	Towel clips small			Deleted
	Page 65, BOQ point 5	Allis tissue holding forceps large			Deleted
	Page 65, BOQ point 6	Allis tissue holding forceps medium			Deleted
	Added Para			Open Surgery clip applicator for use with	Added Para:

Handwritten signatures and initials in blue ink are present at the bottom of the page, including a large signature on the left, several smaller signatures in the center and right, and a large 'L' on the far right.

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
				small size clips	Open Surgery clip applicator for use with small size clips – 1no.
	Added Para			Open Surgery clip applicator for use with small size clips	Added Para: Open Surgery clip applicator for use with medium size clips – 1no.
	Added Para			Open Surgery clip applicator for use with small size clips	Added Para: Open Surgery clip applicator for use with medium large size clips – 1no.
	Added Para			Open Surgery clip applicator for use with small size clips	Added Para: Open Surgery clip applicator for use with large size clips – 1no.
	Added Para			Titanium clips small, medium, medium large and large size for use with clip applicators mentioned	Added Para: Titanium clips small, medium, medium large and large size for use with clip applicators mentioned

Handwritten signatures in blue ink, including a large signature on the left and several smaller ones on the right.

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
Rfx No. 3000003189, Manual Bone Drill Set					
1	page 66, point 2	The Instruments Should be US-FDA or European CE with 4 digit notified body certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	To be amended as: The Instruments Should be US-FDA or European CE certified
			Cure surgical	The Instruments Should be US-FDA & European CE with 4 digit notified body certified	
			Ultimate Medical Devices	We request you to please remove 4 digit certified body from the specifications & change this to European CE or US (FDA)	
2	page 66, point 3	The Instruments should be of medica grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	Deleted
3	page 66, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	Page 66, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	Added para :- Items mentioned at S..No 2,3 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.

Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
5	Page 66, BOQ point 1	S.S Tray (medium)	Cure surgical	Delete	Deleted

Commercial issues					
Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	Page no. 22, Point No. 39.2		Volksmann India Pvt Ltd.	Suggested by bidder :- EMD should be waived off for MSME & Start Up's.	It is advised to refer GIB clause no.19.2 (pg.16) of the bidding document for MSME
	Page no. 33, Point No. 15.1		I.G. Instrument LLP	Suggested by bidder :- 2. Qualification criteria relaxation for msme and startup who manufacture instruments , as per the procurement policy .. This should be waived off who meet quality parameters .	No change considered
	Page no. 33, Point No. 15.4		Volksmann India Pvt Ltd.	Suggested by bidder :- 3. Qualification Criteria should be waived off for start ups	No change considered

