

Amendment No.12

Date: 17/11/2016

Sub: Amendment No.12 to the Tender Enquiry Document**Ref: NIT No.: HLL/PCD/PMSSY-II/11/16-17 dated: 22.08.2016.**

The following changes are made to the Tender enquiry document.

SECTION I**NOTICE INVITING GLOBAL e-TENDERS (NIT)****(i)The due dates for closing and opening of the items (i.e. Sch.13, 14 & 20) are as under:**

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
13	3000001476	Colonoscope with Videoscope	30.11.2016 18:00	01.12.2016 11:30	01.12.2016 14:00
14	3000001477	Upper GI Endoscope	30.11.2016 18:00	01.12.2016 11:30	01.12.2016 14:00
20	3000001485	Complete Micro Motor System for Trauma Care	Deleted.		

**Section – VII
Technical Specifications****Item Sl. No. 13
Colonoscope with Videoscope****1. Existing Para 1.a:** Outer diameter - 11-13 mm.**Read as:** Outer diameter - 11-**13.2** mm**2. Existing Para 1.c:** Depth of field - 3-100 mm.**Read as:** Depth of field: **4**-100 mm**3. Existing Para 2.a:** Xenon light source 300 watts or more.**Read as:** Xenon light source at least **300watts** or more **OR equivalent LED Light source.****4. Existing Para 2.b:** Emergency backup lamp- Halogen 100 watt or more.**Read as:** Deleted

Item Sl. No. 14
Upper GI Endoscope

1. Existing Para 1.j.c: Over tube - 1 No.

Read as: Deleted

2. Existing Para 2: Video processor and light source for endoscopy - 1 Nos (**Optional**-Price to be quoted separately).

Read as: Video processor and light source for endoscopy - 1 Nos (Price to be quoted separately)

3. Existing Para 2.a: Xenon light source 300 watts or more.

Read as: Xenon light source at least **300watts or more OR equivalent LED Light source.**

4. Existing Para 2.b: Emergency backup lamp- Halogen 100 watt or more.

Read as: Deleted

5. Existing Para 3: Video Cart (Imported) 1 NOS (**Optional**-Price to be quoted separately).

Read as: Video Cart (Imported) 1 NOS (Price to be quoted separately)

6. Existing Para 4: Color HD Video Monitor for video processor- 1 No. (**Optional**-Price to be quoted separately).

Read as: Color HD Video Monitor for video processor- 1 No. (Price to be quoted separately)

Item Sl. No. 20
Complete Micro Motor System for Trauma Care

Item is deleted.

All other contents of the tender enquiry including terms & conditions remain unaltered.