

Amendment No. 2

26-10-2017

Sub: Amendment to the Bidding Document.**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/03/17-18 dated: 22.09.2017 read with its Amendment No. 1 dated 25.10.2017.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION - VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item sl. no. 01****16-Slice SPECT / CT Scanner**

Existing	Page 51. Para xiii. Accessories and QC Utility: Automatic collimator change may be included as essential.
Amended as	Page 51. Para xiii. Accessories and QC Utility: Semi-Automatic / Automatic collimator change may be included as essential.

Item sl. no. 02**128-Slice PET / CT Scanner**

Existing	Page 58. 8) Peripherals / Accessories d): Supply of Germanium-68 pin source for a period of 5 years. - Specification of Ge-68 Pin source: 55 MBq (1.49mCi) Ge-68 uniformly dispersed in a ceramic medium with an outer stainless tube and permanently sealed end caps.
Amended as	Page 58. 8) Peripherals / Accessories d): Supply of Germanium-68 / Na-22pin source for a period of 5 years. - Specification of Ge-68 Pin source: 55 MBq (1.49mCi) Ge-68 uniformly dispersed in a ceramic medium with an outer stainless tube and permanently sealed end caps.

All other contents of the Bidding Document including terms & conditions remain unaltered.