

Amendment No. 4

Date: 29/09/2016

Sub: Amendment No.4 to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/10/16-17 dated: 25.07.2016

The following changes are made to the Tender enquiry document.

Section – VII Technical Specifications

Item Sl. No. 1

128 SLICE MDCT WITH INDEPENDENT 64 OR MORE ROWS OF DETECTOR

1. Existing Para 2: Single Source dual energy capability as standard.

Read as: Single Source/**dual source** dual energy capability as standard.

2. Added Para as 2.j under Single Source/dual source dual energy capability as standard: Monochromatic/mono-energetic imaging should be possible.

3. Existing Para 3.c: Tube Voltage: Tube Voltage: Minimum range **80-135 kV**.

Read as: Tube Voltage: Minimum range **80-140 kV or more**.

4. Existing Para 10: Workstation: A multimodality client server architecture based solution with minimum concurrent **30000 slices** rendering capacity (Intellispace Portal 6/ Dexux-AW server 2/ Syngo Via 30A etc), with **64GB RAM** with storage of minimum **2TB** and Additional storage of 10 TB on the server, having following client hardware specification- Workstation:Z820 or equivalent CPU, dual quad core processor, 16 GB RAM, 1TB hard drive, DVD Writing with clinical grade monitor of minimum 3 MP (braw or equivalent) / & 5 button mouse (logitech or equivalent). A reputed Anti- Virus Solution for Server should be in place. The Server should be **Apple or Equivalent** and should be with minimum five concurrent user (five Hardware) facility.

Read as: A multimodality client server architecture based solution with minimum concurrent **20000 slices** rendering capacity (Intellispace Portal 6/ Dexux-AW server 2/ Syngo Via 30A etc), with **32GB RAM** with storage of **minimum 1TB** and Additional storage of 10 TB on the server, having following client hardware specification- Workstation:Z820 or equivalent CPU, dual quad core processor, 16 GB RAM, 1TB hard drive, DVD Writing with clinical grade monitor of minimum 3 MP (braw or equivalent) / & 5 button mouse (logitech or equivalent). A reputed Anti- Virus Solution for Server should be in place. The Server should be with minimum five concurrent user (five Hardware) facility.

5. Existing Para; Partial Turnkey: - Point d (i): Premium executive office Chairs.

Read as: Partial Turnkey: - Point d (i): Premium executive office Chairs- **16 nos**.

All other contents of the tender enquiry including terms & conditions remain unaltered.