

MINUTES OF THE MEETING

TENDER FOR SUPPLY, INSTALLATION, COMMISSIONING AND VALIDATION OF DRY FOG GENERATOR AT PII, COONOOR

Document No. : NPI/110831/EQP/TD/10

Venue : HLL Biotech Limited, Chennai

Date : 10th March 2016

Project : Revival of DPT Vaccines manufacturing facility

Attendees : See attached list of attendees

Issued by : Mr. Suresh S, (Senior Manager . Procurement)


Issued on : 18th March 2016

Issued from : HLL Biotech Limited, Chennai

Agenda	
1.	Pre-bid Meeting for Dry Fog Generator at PII, Coonoor

S. No.	Clarifications on queries	
	Tender for Supply , Installation, commissioning and Validation of Dry Fog Generator at PII, Coonoor	
A	Discussion on Tender Enquiry Document: NPI/110831/EQP/TD/10	
	General Discussion Points	
1.	The Purchaser reserves the right to assess the Tenderer's competency to perform the contract satisfactorily by inspecting their facility. Such assessment shall be done for each technically responsive bidder before opening the Price Bid. The Purchaser reserves the right to ask for a free demonstration of the quoted equipment to similar/identical specification at a pre-determined place acceptable to the purchaser for determining technical responsiveness, before the opening of the Price Bid.	
2.	a. The last date and time for bid submission : 05.04.2016, 15:00 Hrs b. The technical bid opening date & time : 05.04.2016, 15:30 Hrs	
B	Clarifications on Data sheet : Dry Fog Generator (DS/FOG 01)	
	Specific revision in the URS	
	URS Point number and excerpt* / description of the specification *	Point modified as / Comment
1.	Point No. 2.13 Operating Relative humidity < 75% RH	Operating Relative humidity < 80% RH
2.	Point No. 3.1 Spray Nozzle: SS 304 or Clean room Compliant (Vendor to specify)	Spray Nozzle: SS 316L / non-corrosive / Clean room Compliant
3.	Point No. 3.2 Disinfectant solution tank: SS 304	Disinfectant solution tank: SS 316L
4.	Point No. 4.5 Vendor to demonstrate and validate the six log reduction of microbial with suitable indicators.	Vendor to demonstrate and validate the min 3 log reduction of microbial with suitable chemical/biological indicators. (Indicators will be in vendor scope, whereas testing and lab equipments /chemicals for testing will be in PIIC scope)
5.	Point No. 4.6 Portable hydrogen peroxide sensors to be provided inside the room.	Point No. 4.6 Deleted
6.	Vendor to decide the Number of equipment required for decontamination according to the room area mentioned as per annexure-I [DS Pg no. 3 of 6]. i.e., for each block[Diphtheria, Pertussis, Tetanus, Formulation and Microbiology block]	

For HLL Biotech Limited



18/3/16

Authorized Signatory

