

Prebidqueries/clarifications asked by Firms and its recommendations/replies (ref: prebid meeting dated: 07/08/2018)
NIB ref: HITES/PCD/NCI-AIIMS/22/18-19 dated 26.07.2018

Sl. No	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the PrebidCommittee
TECHNICAL ISSUES					
Rfx No. 3000003176, Micro Vascular Instruments Set					
1	page 46, point 2	The Instruments Should be US-FDA certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	No Change
2	page 46, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	Please delete	Deleted
3	page 46, point 4	All needle holders &Metzenbaum scissors should be Tungsten Carbide coated	Cure surgical	All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert

JA

See
Abdul
AB
R

4	page 46, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Indian Surgical Equipment Company Pvt. Ltd.	This point may please be deleted.	To be Amended as: Dedicated Slotted sterilization container container with lock and silicon mat
5	page 46, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as: Items mentioned at BOQ S.No 3,6,11,13,15 and 19 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
6	Page,46 point 4	BOQ, Vein hook	Cure surgical	127mm	To be Amended as: Vein hook:- 11-13cm
7	Page,47 point 13	BOQ, Debakey vascular scissors	Cure surgical	Angled 25°/45°/60° & 165mm	To be Amended as: Debakey vascular scissors 25deg, 15-17cm & 45 deg, 15-17cm: 1 each
			Indian Surgical Equipment Company Pvt. Ltd.	7", 16 mm blade, 45 degree angle	
8	Page,47 point 17	BOQ, Satinsky vascular clamps medium	Indian Surgical Equipment Company Pvt.	10" 4 cm jaw	To be Amended as: Satinsky vascular clamps medium : 24-26cm

			Ltd.		
9	Page,47 point 18	BOQ,	Satinsky vascular clamps small	Indian Surgical Equipment Company Pvt. Ltd.	7 3/4" 2.3 cm jaw To be Amended as: Satinsky vascular clamps small :-18-22cm
10	Page,47 point 19	BOQ,	Debakey's vascular clapms medium	Indian Surgical Equipment Company Pvt. Ltd.	18.2 cm long, angled 6.5 cm jaw To be Amended as: Debakey's vascular clapms medium 16-18cm
11	Page,47 point 20	BOQ,	Debakey vascular clamp small	Indian Surgical Equipment Company Pvt. Ltd.	10.5 cm length jaw 3cm, angled 45 degree. To be Amended as: Debakey vascular clamp small 10-12cm
12	Page,47 point 21	BOQ,	Bull dog clamps 5.5 cm	Cure surgical	Bull dog clamps straght 7.5 cm To be Amended as: Bull dog clamps 5.5-7.5cm
Rfx No. 3000003177, Major Vascular Instruments Set					
1	page 46, point 2		The Instruments Should be US-FDA certified	Volksmann India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. No Change
2	Page 48, point 3		The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete Deleted




3	Page 48, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
4	page 48, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Indian Surgical Equipment Company Pvt. Ltd.	This point may please be deleted.	Deleted
5	Page 48, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as :- Items mentioned at S..No 2,5,9,15,21,40,41,52,58,66 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
6	Page 48, BOQ point 1	Adson non-toothed micro dressing forceps	Cure surgical	127mm	To be Amended as: Adson non-toothed micro dressing forceps: 12-14cm
7	Page 48, BOQ point 3	Artery forceps spencer wells	Cure surgical	140mm	To be Amended as: Artery forceps spencer wells 14-16cm
8	Page 48, BOQ point 5	Artery forceps-Halsted(Adson) curved	Cure surgical	178mm	To be Amended as: Artery forceps-Halsted(Adson) curved 17-19cm
9	Page 48, BOQ point 6	Artery forceps-Halsted(Adson) straight	Cure surgical	178mm	To be Amended as: Artery forceps-Halsted(Adson) straight 17-19cm







10	Page 48, BOQ point 7	Baby satinsky	Cure surgical	127mm	To be Amended as: Baby satinsky 14-18cm
			indian Surgical Equipment Company Pvt. Ltd.	73/4" 2.3 cm jaw	
11	Page 49, BOQ point 9	Bozmann Needle holder serrated	Cure surgical	200mm	To be Amended as: Bozmann Needle holder serrated 20-22cm
			Indian Surgical Equipment Company Pvt. Ltd.	Please provide Sizes/detailed specifications	
12	Page 49, BOQ point 10	Cooley atraumatic forceps	Cure surgical	125mm	To be Amended as: Cooley atraumatic forceps 12-14cm
13	Page 49, BOQ point 11	Cooley vascular clamps	Cure surgical	165mm	To be Amended as: Cooley vascular clamps 16-18cm
			Indian Surgical Equipment Company Pvt. Ltd.	7"	
14	Page 49, BOQ point 12	Crile wood Needle holder serrated	Cure surgical	150/180mm	To be Amended as: Crile wood Needle holder serrated 14-16cm
15	Page 49, BOQ point 13	Cushing dressing forceps (curved)	Cure surgical	180mm	To be Amended as: Cushing dressing forceps (curved) 17-19cm
16	Page 49, BOQ point 14	Cushing nerve hook	Cure surgical	190mm	To be Amended as: Cushing nerve hook 18-20cm
17	Page 49, BOQ point 15	Debakey atraumatic vascular clamp	Cure surgical	165mm	To be Amended as: Debakey atraumatic vascular clamp 15-17cm
18	Page 49, BOQ	Debakey multipurpose	Cure surgical	60mm and 150mm	To be Amended as: Debakey multipurpose

	point 16		vascular clamps			vascular clamps 14-16cm
19	Page 49, point 17	BOQ	Debakey vascular scissors	Cure surgical Indian Surgical Equipment Company Pvt. Ltd.	Angled 25mm 7", 16 mm blade, 45 degree angle	To be Amended as: Debakey vascular scissors 25 deg & 45 deg & length 15-17cm – 1 each
20	Page 49, point 23	BOQ	Debaky atraumatic forceps (angled)	Cure surgical	24cm	To be Amended as: Debaky atraumatic forceps (angled) 23-25cm
21	Page 49, point 31	BOQ	Debaky Needle holder serrated	Cure surgical	180mm	To be Amended as: Debaky Needle holder serrated 17-19cm
				indian Surgical Equipment Company Pvt. Ltd.	Sizes available: 7", 9", 10"	
22	Page 49, point 32	BOQ	Debakysatinsky	Cure surgical indian Surgical Equipment Company Pvt. Ltd.	127mm 7 3/4", 3.5 cm jaw 10 1/8", 4.8 cm jaw	To be Amended as: Debaky satinsky 14-18cm
23	Page 49, point 33	BOQ	Debaky vascular scissors angled	Cure surgical Indian Surgical Equipment Company Pvt. Ltd.	Angled mm 7", 16 mm blade, 45 degree angle	To be Amended as: Debaky vascular scissors angled 45 deg 15-17cm
24	Page 49,	BOQ	Golden Handle small	Indian Surgical Equipment	6 1/4" jaw 16x2mm	To be Amended as: Golden Handle small needle holder 6-8"

	point 34		needle holder	Company Pvt. Ltd.		
25	Page 49, point 35	BOQ	Halsted mosquito hemostatic forceps	Cure surgical	120mm	To be Amended as: Halsted mosquito hemostatic forceps 12cm
26	Page 49, point 37	BOQ	Kelley clamp long curved	Indian Surgical Equipment Company Pvt. Ltd.	6 1/4." curved	To be Amended as: Kelley clamp long curved 14-16cm
27	Page 49, point 38	BOQ	Lahey gall duct forceps	Cure surgical	190mm	To be Amended as: Lahey gall duct forceps 18-20cm
				Indian Surgical Equipment Company Pvt. Ltd.	8"	
28	Page 49, point 41	BOQ	metzenbaum dissecting scissors	Cure surgical	150mm	To be Amended as: Metzenbaum dissecting scissors 15-17cm
				Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 7", 8"	
29	Page 49, point 42	BOQ	metzenbaum dissecting scissors curved	Cure surgical	150mm	To be Amended as: Metzenbaum dissecting scissors curved 14-16cm
				Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 7", 9"	
30	Page 50, point 46	BOQ	Mosquito Artery forceps(curved)	Cure surgical	127mm	To be Amended as: Mosquito Artery forceps(curved) 12-14cm
31	Page 50,	BOQ	Mosquito Artery	Cure surgical	127mm	To be Amended as: Mosquito Artery forceps(straight) 12-14cm

	point 47		forceps(straight)			
32	Page 50, point 52	BOQ	Pott's Smith dissecting forceps	Cure surgical	184mm	To be Amended as: Pott's Smith dissecting forceps 17-19cm
				Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 7", 8"	
33	Page 50, point 53	BOQ	Reynolds scissors fine curved	Cure surgical	177mm	To be Amended as: Reynolds scissors fine curved 17-19cm
34	Page 50, point 57	BOQ	S.S Tray (Big)	Cure surgical	delete	Deleted
				indian Surgical Equipment Company Pvt. Ltd.	May kindly be deleted.	
35	Page 50, point 58	BOQ	Satinsky blood vessel clamp	Cure surgical	64 x 10 mm 9 1/2 in 241 mm	To be Amended as: Satinsky blood vessel clamp medium 22-24cm
				Indian Surgical Equipment Company Pvt. Ltd.	8 1/2" jaw 3.5 cm	
36	Page 50, point 59	BOQ	Satinsky blood vessel clamp (Big)	Cure surgical	70 x 13 mm 9 1/2 in 241 mm	To be Amended as: Satinsky blood vessel clamp (Big) 26-28cm
				Indian Surgical Equipment Company Pvt. Ltd.	10 3/8" jaw 5 cm	

37	Page 50, point 60	BOQ	Satinsky blood vessel clamp (Small)	Cure surgical	57 x 6.0 mm 9 1/2 in 241 mm	To be Amended as: Satinsky blood vessel clamp (Small) 20-24cm
				Indian Surgical Equipment Company Pvt. Ltd.	7 3/4" jaw 2.3 cm	
38	Page 50, point 64	BOQ	Vein Hook	Cure surgical	127mm	To be Amended as: Vein Hook 12-14cm
39	Page 50, point 68	BOQ	Eyelid retractor	Cure surgical	140mm	To be Amended as: Eyelid retractor 12-14cm
				Indian Surgical Equipment Company Pvt. Ltd.	Please provide Sizes/detailed specifications	
Rfx No. 300003178, Thoracic vascular Instruments Set						
1	page 46, point 2		The Instruments Should be US-FDA certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	No Change
2	Page 51, point 3		The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	Deleted
3	Page 51, point 4		All needle holders &Metzenbaum scissors should be	cure surgical	All needle holders &Metzenbaum scissors should be Tungsten	To be Amended as: All needle holders &Metzenbaum scissors should be Tungsten carbide insert



		Tungsten Carbide coated		Carbide Insert	
4	page 48, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Indian Surgical Equipment Company Pvt. Ltd.	This point may please be deleted.	Deleted
5	Page 51, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	To be Amended as :- Items mentioned at BOQ S.No 4,11,17,21,23,29 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval
6	Page 51, BOQ point 1	S.S Tray Large	Cure surgical	Delete	Deleted
			Indian Surgical Equipment Company Pvt. Ltd.	May kindly be deleted	
7	Page 51, BOQ point 9	Satinsky blood vessel clamp big	Cure surgical	70 x 13 mm 9 1/2 in 241 mm	To be Amended as: Satinsky blood vessel clamp big 24-27
			Indian Surgical Equipment Company Pvt.	10 3/8" jaw 5 cm	




			Ltd.		
8	Page 52, BOQ point 14	Adson hemostat forceps straight	Cure surgical	150mm	To be Amended as: Adson hemostat forceps straight 14-16cm
			Indian Surgical Equipment Company Pvt. Ltd.	6 1/4"	
9	Page 52, BOQ point 15	Adson hemostat forceps curved	Cure surgical	150mm	To be Amended as: Adson hemostat forceps curved 14-16cm
			Indian Surgical Equipment Company Pvt. Ltd.	6 1/4"	
10	Page 52, BOQ point 29	Pott's smith dressing forceps TC	Cure surgical	180mm	To be Amended as: Pott's smith dressing forceps TC 17-19cm
			Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 7", 8"	
11	Page 52, BOQ point 32	Cushing nerve hook medium	Cure surgical	197mm	To be Amended as: Cushing nerve hook medium 18-20cm
Rfx No. 3000003179, Plastic Surgical Instruments Set					
1	page 46, point 2	The Instruments Should be US-FDA certified	Volksman n India Pvt Ltd.	ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS.	No Change



2	Page 53, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Cure surgical	pl delete	To be Amended as: Deleted
3	Page 53, point 4	All needle holders &Metzenbaum scissors should be Tungsten Carbide coated	cure surgical	All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert	To be Amended as: All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert
4	page 53, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Indian Surgical Equipment Company Pvt. Ltd.	This point may please be deleted.	Deleted
5	Page 53, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Cure surgical	need instrument list and six week time	Added Para : - Items mentioned at S..No 4, 9,12,16,19 are to be demonstrated. and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
6	Page 53, BOQ	Non toothed	Cure surgical	178mm	To be Amended as: Non toothed dissecting



	point 5		dissecting forceps golden TC tip	Indian Surgical Equipment Company Pvt. Ltd.	7"	forceps golden TC tip 17-19cm
7	Page 53, point 8	BOQ	Toothed dissecting forceps	Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 6" 7" 8"	To be Amended as: Toothed dissecting forceps 17-19cm
8	Page 54, point 12	BOQ	Iris scissor angled golden	Cure surgical	115mm	To be Amended as: Iris scissor angled golden 10-12cm
9	Page 54, point 14	BOQ	Castro-vejio needle holder golden	Cure surgical	146mm	To be Amended as: Castro-vejio needle holder golden 14-16cm
				Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 6" 7" 8"	
10	Page 54, point 15	BOQ	Needle holder golden	Cure surgical	Medium	To be Amended as: Needle holder golden medium
				Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 6" 7" 8"	
11	Page 54, point 16	BOQ	Needle holder small	Indian Surgical Equipment Company Pvt. Ltd.	Sizes available 6" 7" 8"	To be Amended as: Needle holder small 6-8"




12	Page 54, point 19	BOQ	Muller retractor	Cure surgical	160mm	To be Amended as: Muller retractor 15-17cm
				Indian Surgical Equipment Company Pvt. Ltd.	Please provide Sizes/detailed specifications	
13	Page 54, point 20	BOQ	Adson retractor	Cure surgical	240mm	To be Amended as: Adson retractor 23-25cm
				Indian Surgical Equipment Company Pvt. Ltd.	6 3/4" 4x4 blunt, 7 1/4" 4x4 sharp	
14	Page 54, point 21	BOQ	S.S Tray (medium)	Cure surgical	delete	Deleted
				Indian Surgical Equipment Company Pvt. Ltd.	May kindly be deleted	
15	Page 54, point 22	BOQ	Humby's Knife	Cure surgical	delete	No Change
	As per pre bid discussion for Sh 1. To 4.				Demonstration of sample Instruments to be marked	



Commercial issues					
Sl. No.	Ref. to the Bidding Document	Specification/clauses in the Bidding Document	Name of Firms	Queries/clarifications asked by Firms	Recommendations/ replies by the Prebid Committee
	Page no. 22, Point No. 39.2		Volsman n India Pvt Ltd.	Suggested by bidder :- EMD should be waived off for MSME & Start Up's.	It is advised to refer GIB clause no.19.2 (pg.16) of the bidding document for MSME
	Page no. 33, Point No. 15.1		I.G. Instrument LLP	Suggested by bidder :- 2. Qualification criteria relaxation for msme and startup who manufacture instruments , as per the procurement policy .. This should be waived off who meet quality parameters .	No change considered
	Page no. 33, Point No. 15.4		Volsman n India Pvt Ltd.	Suggested by bidder :- 3. Qualification Criteria should be waived off for start ups	No change considered





 (Chairman)