

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 01- Mobile O.T light (Rfx/Event number 3000002600)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
	Page 44, Para 3	There should be single dome equipped with not more than 15 LEDs and minimum 50,000 lux (measured at 1 meter distance).	as per pre-bid discussion	request to increase LED's to 20 and minimum lux to 1,20,000	Ammended as: There should be single dome equipped with not more than 20 LEDs and minimum 50,000 lux (measured at 1 meter distance).
	Page 44, Para 7	Energy irradiated at maximum level from one dome should not be more than 120 W/m2.	as per pre-bid discussion	request to delete this point	Ammended as: Deleted
	Page 44, Para 9	Optional battery backup with batte	as per pre-bid discussion	request to make batery back up as standard and inbuilt for 30 minutes	Ammended as: inbuilt 30 min battery backup should be offered.
				as per NCI AIIMS Committee recommendation	Added Para: BOQ Mobile OT Light - 1 No.
					Total Quantity to be changed to: 9 MobileOT Lights, Estimated cost per unit:

Chitank

See

Shah

g

Sunil

(Signature)





(Signature)

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 02- Ultrasonic Bone Scalpel (Rfx/Event number 3000002602)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
	Page 45, Para 5	Hand piece with LED mounted for illumination of	as per pre-bid discussion	request to delete this point	No Change
				as per NCI AIIMS Committee recommendation	Added Para: BOQ (per each machine)
				as per NCI AIIMS Committee recommendation	Ultrasonic Bone Scalpel - 1 No.
				as per NCI AIIMS Committee recommendation	Hand Piece - 1 No.
				as per NCI AIIMS Committee recommendation	Angled Right Tips - 5 Nos.
				as per NCI AIIMS Committee recommendation	Angled Left Tips - 5 Nos.
				as per NCI AIIMS Committee recommendation	Osteotomy saw blades - 5 Nos.
				as per NCI AIIMS Committee recommendation	Diamond finishing blades of different shapes - 5 Nos.
				as per NCI AIIMS Committee recommendation	Irrigation sets - 10 nos.
				as per NCI AIIMS Committee recommendation	Sterilization tray - 1 No.
				as per NCI AIIMS Committee recommendation	Foot Pedal - 1 No.

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 03 - Multipurpose Pneumatic Saw and Drill system - (Rfx no. 3000002603)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
1	Page.46, Para.1	Multipurpose pneumatic saw and drill system for fine and large bone work including craniotomy & sternotomy.	M/s Jupiter Healthcare Company, Delhi, INDIA	Multipurpose pneumatic saw and drill system for fine and large bone work including craniotomy. Reason: As Craniotomy & Sternotomy are from different specialisation, Sternotomy in the specification is restrictive	No change
2	Page.46, Para.4	Power should be in the range of 140 Watts at 120 psi	M/s Jupiter Healthcare Company, Delhi, INDIA	Power should be 90W or more @ 8.3 Bar Reason: 140 Watts will restrict us from participating. 90W power provided by Xmax does all Cranial & Maxillofacial drilling & cutting of bone	Amended as: Power should be in the range of 90 Watts or more at 120 psi
3			JNJ	power should be at-least 90 watt at 8 bar.	
4	Page.46, Para.5	Should operate between 2 to 8 bar (20 - 120 psi)	M/s Jupiter Healthcare Company, Delhi, INDIA	Should operate between 2.75 to 8 bar (40 - 120 psi)	No change
5			JNJ	should operate between 3-8 bar or 40-120 psi	
6	Page.46, Para.7	Should have open Foot pedal design for easy access and repositioning and should have automatic port cover to protect the motor port when not in use	M/s Jupiter Healthcare Company, Delhi, INDIA	Should have open Foot pedal design for easy access and repositioning and should have automatic port cover to protect the motor port when not in use	No change
7	Page.46, Para.9	Foot pedal should have option for connection of Saw hand pieces	M/s Jupiter Healthcare Company, Delhi, INDIA	Common Foot pedal for all attachments including Saw hand pieces	No change
8	Page.46, Para.10	Regulator hose should be detachable at both regulator side and foot pedal side	M/s Jupiter Healthcare Company, Delhi, INDIA	Regulator hose should be detachable at both regulator side and foot pedal side or only at regulator side	Amended as: Regulator hose should be detachable at both regulator side and/or foot pedal side
9	Page.46, Para.12	Sound level should not be more than 70dB @ 72000rpm @ 120psi / 8 bar	M/s Jupiter Healthcare Company, Delhi, INDIA	<= 76dB	Amended as: Sound level should be ≤ 76dB @ 72000rpm @ 120psi / 8 bar
10			JNJ	sound level should not be more than 76 db	

Delish

Ashwini

[Signature]

[Signature]

[Signature]

[Signature]

Schedule No. 03 - Multipurpose Pneumatic Saw and Drill system - (Rfx no. 3000002603)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
11	Page.46, Para.17	Finger control with variable speed should be available	M/s Jupiter Healthcare Company, Delhi, INDIA	Please delete this, as we will not be able to participate	No change
12			JNJ	Finger control or foot-switch with variable speed should be available	
13	Page.46, Para.18	Should quote for Micro Oscillating saw, Reciprocating saw which can be detachable from the gas hose	M/s Jupiter Healthcare Company, Delhi, INDIA	Should quote for Micro Oscillating saw, Reciprocating saw which can be detachable from the gas hose or the motor	Amended as: Should quote for Micro Oscillating saw, Reciprocating saw which can be detachable from the gas hose or the motor
14	Page.46, Para.20	Oscillating saw should operate between 20000 to 23,000 cpm speed and Reciprocating saw should operate between 10000 to 13,500 cpm	M/s Jupiter Healthcare Company, Delhi, INDIA	Oscillating & Reciprocating saw should operate @12000 cpm speed	Amended as: Oscillating saw should operate between 10000 cpm or more speed and Reciprocating saw should operate between 10000 cpm or more
15			JNJ	oscillating saw to operate between 10,000-13,500 cpm	
16	Page.46, Para.21	Blades should be of stainless steel & Blades should be of 0.4mm thickness with fine cutting design	M/s Jupiter Healthcare Company, Delhi, INDIA	Blades should be of stainless steel, Single Use, Sterile & be of appropriate thickness with fine cutting design	Amended as: Blades should be of stainless steel & Blades should be of 0.4mm-0.6mm thickness with fine cutting design
17	Page.46, Para.21	Micro and Macro drill attachment with drill bits	as per pre-bid discussion	specify quantity of drill bits	Amended as: Micro and Macro drill attachment with 5 nos. of drill bits
18	Page.46, Para.23	Hand piece - Short Straight Attachment (7cm & 8cm): 1 No. each.	M/s Jupiter Healthcare Company, Delhi, INDIA	Hand piece - Short Straight & Angled Attachment : 1 No. each. Reason: Each drill brand has length optimised to their respective construct & length in specification restricts participation in tender	Amended as: Hand piece - Short Straight & Angled Attachment : 1 No. each.
19	Page.46, Para.24	Hand piece - Medium Straight & Angled Attachment 9cm: 1 No each.	M/s Jupiter Healthcare Company, Delhi, INDIA	Hand piece - Medium Straight & Angled Attachment 1 No each. Reason: Each drill brand has length optimised to their respective construct & length in specification restricts participation in tender	Amended as: Hand piece - Medium Straight & Angled Attachment: 1 No each.

(Handwritten signature)

(Handwritten signature)

(Handwritten signature)

(Handwritten signature)

(Handwritten signature)

Schedule No. 03 - Multipurpose Pneumatic Saw and Drill system - (Rfx no. 3000002603)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
20	Page.46, Para.26	Sternotomy handpiece with guard - 1	M/s Jupiter Healthcare Company, Delhi, INDIA	Request this specification to be deleted Reason: As Sternotomy & Craniotomy are from different specialities, this specification restricts the no of companies that can participate in tender	Amended as: Sternotomy/sternal handpiece with guard - 1
21			JNJ	reciprocating saw attachment system to be provided for sternotomy	
22	Page.46, Para.32-6	Diffuser for Lubrication of Motor	M/s Jupiter Healthcare Company, Delhi, INDIA	In line lubrication of motor with diffuser or Oil receptacle as per specification 16	Amended as: In line lubrication of motor with diffuser or Oil receptacle
23	Page.46, Para.32	Should quote separately for the following consumables : 1. Blades for oscillating and reciprocating saws – Various sizes & shapes 2. Drill bits 3. Fast Cutting Burrs (3 mm) 4. Diamond Burrs (3 mm) 5. Bone Cutter Drill bits for craniotomy 6. Diffuser for Lubrication of Motor	as per pre-bid discussion	specify quantity of each item	Should quote separately for the following consumables : 1. Blades for oscillating and reciprocating saws – Various sizes & shapes - 5 Nos each 2. Drill bits - 5 Nos 3. Fast Cutting Burrs (3 mm) - 5 Nos 4. Diamond Burrs (3 mm) - 5 Nos 5. Bone Cutter Drill bits for craniotomy - 5 Nos 6. Diffuser for Lubrication of Motor - 5 Nos
24				as per NCI AIIMS Committee recommendation	Added Para: BOQ
25					Micro Oscillating Saw - 1 No.
26					Reciprocating Saw - 1 No.
27					Drill Hand piece with Micro and Macro drill attachment - 1 No. each
28					Hand Piece - Short Straight attachment - 1 No.
29					Hand Piece - Medium Straight attachment - 1 No.
30					Hand Piece - Medium Angled attachment - 1 No.
31					Craniotomy burr hole attachment with duraguard (Adult) - 1 No.

Signature

Signature

Signature

Signature

Signature

Signature

Schedule No. 03 - Multipurpose Pneumatic Saw and Drill system - (Rfx no. 3000002603)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
32					Sternotomy/Sternal Hand Piece with guard - 1 No.
33					Perforator driver - 1 No.
34					Cleaning brushes of different sizes - 1 No. each
35					Blades for oscillating and reciprocating saws - various sizes & shapes - 5 nos. each
36					Drill bits - 5 Nos
37					Fast cutting burrs (3 mm) - 5 Nos.
38					Diamond burrs (3 mm) - 5 Nos.
39					Bone cutter drill bits for craniotomy - 5 Nos.
40					In line lubrication of motor with diffuser or Oil receptacle - 5 Nos.
41					Connecting Hose - 1 No.

Delivered

There

✓

for

✓

✓

Sundia

for

for

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIMS/12/17-18 Dated: 05.02.2018

Schedule No. 04 - MULTIPURPOSE ELECTRIC DRILL & SAW SYSTEM (Rfx/Event number 3000002604)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
1	Pg 47 ,Para A	Touchscreen controls with colored Graphic display	M/s Jupiter Healthcare Company , Delhi , INDIA	Touchscreen controls with colored graphic display or integral touch pad face panel	No change
2			JNJ	Touchscreen controls with colored graphic display or integral touch pad face panel	No change
3	Pg 47 ,Para A	Footswitch and the hand-piece speed	M/s Jupiter Healthcare Company , Delhi , INDIA	hand-piece speed in LCD or touch pad face panel	No change
4	Pg 47 ,Para A	Capability to accept and use at least two hand-pieces simultaneously	M/s Medtronic	two hand pieces simultaneously cannot be operated. Simultaneous should be deleted	Amended as: Capability to accept at least two hand-pieces.
5	Pg 47 ,Para B	Accelerator as well as fixed speed running mode for reliable and accurate speed control	M/s Jupiter Healthcare Company , Delhi , INDIA	Accelerator as well as fixed speed in increments of 10000 RPM for reliable and accurate speed control using the footswitch	No change
6			JNJ	Accelerator as well as fixed speed in increments of 10000 RPM for reliable and accurate speed control using the footswitch	No change
7	Pg 47 ,Para B	Control for forward, reverse and oscillate modes	M/s Jupiter Healthcare Company , Delhi , INDIA	Control for forward, reverse revolution of the motor	No change
8	Pg 47 ,Para C	Flexible, Autoclavable electrical cord, with push-pull type connectors,	as per pre-bid discussion	request to delete the word push-pull	Amended as: Flexible, Autoclavable electrical cord, with connectors,
9	Pg 47 ,Para D .1	Sterilisable through Steam, ETO & Flash hand-piece,	M/s Jupiter Healthcare Company , Delhi , INDIA	Sterilisable through Steam hand-piece, Rationale for changes: ETO is not safe & regular flash damages the equipment	Amended as: Sterilisable through Steam or ETO or Flash hand-piece,
10			JNJ	Sterilisable through Steam hand-piece,	
11	Pg 47 ,Para D 1	Feedback circuitry to maintain speed and torque, as per the surgeon's selection.	M/s Jupiter Healthcare Company , Delhi , INDIA	Feedback circuitry to maintain speed , as per the surgeon's selection.	No change

Signature

Signature

Page 1 of 5

Signature

Signature

Signature

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018
Schedule No. 04 - MULTIPURPOSE ELECTRIC DRILL & SAW SYSTEM (Rfx/Event number 3000002604)

12		JNJ	Feedback circuitry to maintain speed, as per the surgeon's selection.	
13	Pg 47, Para D 1(v)	as per pre-bid discussion	specify quantity of each bur	Amended as: Burs: 1 mm, 2 mm, 3 mm and 4 mm - 5 nos each.
14	Pg 47, Para D 1(vi)	as per pre-bid discussion	specify quantity of each bur	Amended as: Diamond Burs: 2 mm, 3 mm, 4 mm - 5 nos each
15	Pg 47, Para D 2	M/s Jupiter Healthcare Company, Delhi, INDIA	Sterilisable through Steam hand-piece, Rationale for changes: ETO is not safe & regular flash damages the equipment	Amended as: Sterilisable through Steam or ETO or Flash hand-piece,
16		JNJ	Sterilisable through Steam hand-piece,	
17	Pg 47, Para D 2 a.	M/s Jupiter Healthcare Company, Delhi, INDIA	Straight attachment : approx dimension 7-8 cm	No change
18	Pg 47, Para D 2 b.	M/s Jupiter Healthcare Company, Delhi, INDIA	Straight attachment : Approximate dimension 11-12 cm	No change
19	Pg 47, Para D 2 c.	M/s Jupiter Healthcare Company, Delhi, INDIA	Angled attachment : Approximate dimension 7-8 cm	No change
20	Pg 47, Para D 2 d.	M/s Jupiter Healthcare Company, Delhi, INDIA	Angled attachment : Approximate dimension 11-12 cm	No change
21	Pg 47, Para D 2 f.	M/s Jupiter Healthcare Company, Delhi, INDIA	Footed Attachment for Craniotomy – non rotating, Rationale for changes: Dimensions of each brand is optimised & hence 16mm should be removed	No change
22		JNJ	Footed Attachment for Craniotomy – non rotating	
23	Pg 47, Para D 2 g.	M/s Jupiter Healthcare Company, Delhi, INDIA	Single use sterile perforators, Adult & Paediatric	Amended as: Perforator Bits - 5 nos
24		JNJ	Single use sterile perforators, Adult & Paediatric	
25	Pg 47, Para D 2 h.	M/s Jupiter Healthcare Company, Delhi, INDIA	Single use sterileCutter for Footed attachment	Amended as: Cutter for Footed attachment - 5 nos

(Handwritten signatures and initials)

Schedule No. 04 - MULTIPURPOSE ELECTRIC DRILL & SAW SYSTEM (Rfx/Event number 3000002604)

26		JNJ	Single use sterileCutter for Footec attachment	
27	Pg 47 ,Para D 2 i.	M/s Jupiter Healthcare Company , Delhi , INDIA JNJ	Single use, sterileStainless Steel Burs: 1 mm, 2 mm, 3 mm and 4 mm Single use, sterileStainless Steel Burs: 1 mm, 2 mm, 3 mm and 4 mm	Ammeded as: Burs: 1 mm, 2 mm, 3 mm and 4 mm - 5 nos each
28		JNJ	Single use, sterile Diamond Burs: 2 mm, 3 mm, 4 mm	Ammeded as: Diamond Burs: 2 mm, 3 mm, 4 mm - 5 Nos each
29	Pg 47 ,Para D 2 j.	M/s Jupiter Healthcare Company , Delhi , INDIA JNJ	Single use, sterile Diamond Burs: 2 mm, 3 mm, 4 mm	
30		JNJ	Single use, sterile Diamond Burs: 2 mm, 3 mm, 4 mm	
31	Pg 48 ,Para D 3	as per pre-bid discussion	please amend speed as 15000 - 20000 cpm or more	Ammeded as: Micro oscillator saw handpiece: Maximum speed of 15000 cpm or more
32	Pg 48 ,Para D 3	as per pre-bid discussion	please specify the quantity of blades	Ammeded as: Choice of Blades - 20 nos. (as per user requirement)
33	Pg 48 ,Para D 4	as per pre-bid discussion	please amend speed as 15000 - 20000 cpm or more	Ammeded as: Micro sagittal saw hand-piece: Maximum speed of 15000 cpm or more
34	Pg 48 ,Para D 4	as per pre-bid discussion	please specify the quantity of blades	Ammeded as: Choice of Blades - 20 nos.(as per user requirement)
35	Pg 48 ,Para D 5	as per pre-bid discussion	please amend speed as 10000 - 15000 cpm or more	Ammeded as: Micro reciprocating saw hand-piece: Maximum speed of 10000 cpm or more
36	Pg 48 ,Para D 5	as per pre-bid discussion	please specify the quantity of blades	Ammeded as:Choice of Blades - 20 nos.(as per user requirement)
37	Pg 48 ,Para D 6	as per pre-bid discussion	please amend speed as 1000 rpm or more	Ammeded as: Universal driver: Should have variable speed control, with a Maximum speed of 1000 rpm or more
38			as per NCI AIIMS Committee recommendation	Added Para: BOQ
39				Microprocessor based control unit - 1 No.
40				Foot switch - 1 No.
41				Connecting card - 1 No.

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Schedule No. 04 - MULTIPURPOSE ELECTRIC DRILL & SAW SYSTEM (Rfx/Event number 3000002604)

42		Micro Drill handpiece for onco-plastic & reconstructive surgeries - 1 No.
43		Straight short micro drill attachment - 1 No.
44		Straight long micro drill attachment - 1 No.
45		Angled short micro drill attachment - 1 no.
46		Angled long micro drill attachment: - 1 no.
47		Cutting burrs for micro drill: 1 mm, 2 mm, 3 mm and 4 mm - 5 Nos each
48		Diamond burrs for micro drill: 2 mm, 3 mm and 4 mm - 5 Nos each
49		High torque micro drill Hand piece for onco-skull, spine & head and neck surgeries - 1 No.
50		Straight high torque micro drill attachment 7-8 cm - 1 No.
51		Straight high torque micro drill attachment 11-12 cm - 1 No.
52		Angled high torque micro drill attachment 7-8 cm - 1 No.
53		Angled high torque micro drill attachment 11-12 cm - 1 No.
54		Perforator chuck - 1 No.
55		Footed attachment for craniotomy - non rotating, 16 mm - 1 No.
56		Perforator bits for high torque micro drill - 5 Nos.
57		Cutter for footed attachment - 5 Nos.
58		cutting burrs for high torque micro drill: 1 mm, 2 mm, 3 mm and 4 mm - 5 Nos each
59		Diamond burrs for high torque micro drill: 2 mm, 3 mm and 4 mm - 5 Nos each
60		Micro oscillating saw handpiece - 1 No.
61		Blades for oscillating saw - 20 nos. (as per user requirement)
62		Micro sagittal saw handpiece - 1 No.
63		Blades for sagittal saw - 20 nos. (as per user requirement)
64		Micro Reciprocating saw handpiece - 1 No.








NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 04 - MULTIPURPOSE ELECTRIC DRILL & SAW SYSTEM (Rfx/Event number 3000002604)

65		Blades for reciprocating saw - 20 nos. (as per user requirement)
66		Universal driver - 1 No.
67		Jacob's chuck - 1 No.
68		Wire chuck - 1 No.
69		Pin chuck - 1 no.
70		Micro-debrider - 1 No.
71		Blades for micro-debrider:
72		2.5 mm Aggressive, 11 cm - 1 box
73		4.0 mm Aggressive, 11 cm - 1 box
74		4.0 mm Aggressive, Serrated, 11 cm - 1 box
75		4.0 mm Angled Aggressive, 40 degree, 11 cm - 1 box
76		4.0 mm Angled Aggressive, 60 degree, 11 cm - 1 box
77		4.0 mm Angled, convex, 11 cm - 1 box

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 05 - Biological Fluid Automated Aspirator and Disinfectant (Rfx/Event number 3000002605)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
				as per NCI AIIMS committee recommendation	Added Para: BOQ
				as per NCI AIIMS committee recommendation	Biological fluid automated aspirator and disinfectant - 1 No.
				as per NCI AIIMS committee recommendation	consumables including filters and certified disinfectant packs/bottles sufficient for 100 uses

Joe

Neha

Dr. R. S. Singh

Srinivas

Dr

Dr. Sree

Dr. Srinivas

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 07 - Molecular Resonance Lancet (Rfx/Event number 3000002607)

SI. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
				as per NCI AIIMS Committee recommendation	Added Para: BOQ
				as per NCI AIIMS Committee recommendation	Molecular Resonance Lancet - 1 No.
				as per NCI AIIMS Committee recommendation	Consumables:
				as per NCI AIIMS Committee recommendation	Needles/loops/blades for coagulation, cutting & fulguration functions - 10 Sets
				as per NCI AIIMS Committee recommendation	Double finger Switch (cut & coagulation, three pin type) - 10 Nos.
				as per NCI AIIMS Committee recommendation	One each of Bipolar Forceps Straight, 45 degree, Bayonet Shaped (stainless steel) user selectable with 3 Nos. cable.
				as per NCI AIIMS Committee recommendation	Reusable patient plate - 5 Nos

See

for

Dr. [Signature]

Sankar

[Signature]

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 08 -OT Table (High End) (RfxEvent number 3000002608)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
1	Pg 52 ,Para I	Multipurpose Electro Mechanical/Electro Hydraulic with manual/electric override mobile Table with divided leg section suitable for all major surgical procedures, complete with 5cm or more foam mattress and corded handset.	M/s Medex India(P) Ltd. New Delhi	Multipurpose Electro Mechanical/Electro Hydraulic/ powered with manual/electric override mobile Table with divided leg section suitable for all major surgical procedures complete with 5 cm or More foam mattress and corded handset. Reasoning: Powered and Electro-mechanical are same technology with different names.Both are 100% oil free and latest technology for infection free operation in a sterile zone.	amended as: Multipurpose Electro Mechanical/Electro Hydraulic/ Powered with manual/electric override mobile Table with divided leg section suitable for all major surgical procedures, complete with 5cm or more foam mattress and corded handset.
2	Pg 52 ,Para II	Electrical specifications:	as per pre-bid discussion	request to amend as Technical Specifications	amended as: Technical Specifications
3	Pg 52 ,Para II 5.	Width: 550-600 mm	as per pre-bid discussion	request to amend as Width: 550-600 mm with side rails	amended as: Width: 550-600 mm (with side rails)
4	Pg 52 ,Para II 7.	Maximum height (without mattress) : More than 1050 mm.	as per pre-bid discussion	request to amend as Maximum height (without mattress) : More than 1000 mm.	amended as: Maximum height (without mattress) : More than 1000 mm.
5	Pg 52 ,Para II(8)	Maximum lateral tilt : 20 – 30 deg. or more (both sides)	M/s Medex India(P) Ltd. New Delhi	Maximum lateral tilt : 18 – 20 deg. or more (both sides) Reasoning: In case tilt is given more than 20 deg,the patient may tend to fall.Also most of the reputed manufacturers have 18 – 20 deg tilt.	amended as: Maximum lateral tilt : 18 – 25 deg. or more (both sides)
6	Pg 52 ,Para II(16)	Minimum working Load capacity : 270 kg or more (All positions)	M/s Medex India(P) Ltd. New Delhi	Minimum working Load capacity : 270 kg or more (All positions) and 450 Kg (Static position) Reasoning: As per European CE and USFDA the tables having patient load capacity of 270 Kgs(All position) the static patient carrying capacity should be 450Kg or more .Hence this request to make point more clear.	No Change

(Handwritten signatures and initials)


Page 1 of 3


Schedule No. 08 -OT Table (High End) (RfxEvent number 3000002608)					
Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
7			as per pre-bid discussion	Minimum working Load capacity : 250 kg or more (All positions)	
8	Pg 53 ,Para II(12)	Leg section adjustment : + 50 deg. to 90 deg.	as per pre-bid discussion	request to amend as Leg section adjustment: +10 deg. to - 90 deg.	amended as: Leg section adjustment: +10 deg. to - 90 deg.
9	Pg 53 ,Para II(15)	Longitudinal movement : ± 250 – 300 mm	as per pre-bid discussion	request to amend as Longitudinal movement: ±150 mm (both sides) from centre	amended as: Longitudinal movement: 150 mm or more from centre
10	Pg 52 ,Para III	Technical Specifications -Accessories			Amended as: Standard Accessories (required for 9 nos. OT Tables)
11	Pg 53 ,Para III(2)	Lithotomy leg holders			Amended as: Deleted
12	Pg 53 ,Para III(8)	Arm support, Perspex – 2 Pc	as per pre-bid discussion	request to amend as Lateral arm support- 1 No.	amended as: Lateral arm support- 1 No.
13		Added Para:	as per pre-bid discussion	as per NCI AllIMS Committee recommendation	Added Para: Large arm board for hand surgery- 2 nos.
14		Added Para:			Added Para: Clamp for attaching side supports: 3 nos.
15		Added Para:			Added Para: Lateral support - 1no.
16		Added Para:			Added Para: Back support: 2 nos.
17		Added Para:			Added Para: Pubic support: 1 no.
18		Added Para:			Added Para: Tunnel cushion for leg in lateral position: 1 no.
19		Added Para:			Added Para: Gel Chest Rolli: 2 nos.
20		Added Para:			Added Para: Gel Head rings Medium & Large- 1 no. each.
21	Pg 53 ,Para V.a	Gynae & Urology Accessories			Gynae and Urology accessories (required for 4 nos. OT Table)
22	Pg 53 ,Para V.a(iii)	Clamp for attaching side supports: 3 nos.			Amended as: Deleted
23	Pg 53 ,Para V.a(iv)	Lateral support: 1 no.			Amended as: Deleted
24	Pg 53 ,Para V.a(v)	Back support : 1 No.	as per pre-bid discussion	request that the specification should be deleted	amended as: Deleted.








Schedule No. 08 -OT Table (High End) (RfxEvent number 3000002608)			
Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM
		REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
25	Pg 53 ,Para V.a(vi)	Pubic Support: 1 no.	
26	Pg 53 ,Para V.a(vii)	Tunnel Cushion for leg in lateral position: 1 No.	request to clarify this specification
27	Pg 53 ,Para V.b	Neurosurgery	
28	Pg 53 ,Para V.b(i)	Connection Bracket : 1 No.	request to increase the quantity to 2 nos
29	Pg 53 ,Para V.b(iv)	Doro Skull Clamp : 1 No	request to clarify this specification
30	Pg 53 ,Para V.b(vi)	Connecting Fixture : 1 No	request to be deleted
31	Pg 53 ,Para V.b(viii)	Prone Head rest - 1no.	
32		Added Para:	
33		Added Para:	
34		Added Para:	









BOQ for OT Table:

S.No.	Name of the Item	Qty
1.	OT Table as per technical Specifications	9 nos.
2.	Standard Accessories 1. Arm Board – 2 2. Body strap – 3 3. Anaesthesia screen – 1 4. Clamp, rotary – 2 pc. 5. Clamp, circular – 2 Pc 6. Accessories stand, mobile on castors – 1 Pc 7. Lateral support Arm-1no. 8. Clamp for attaching side supports- 3 nos. 9. Lateral support- 1 no. 10. Back support- 2 nos. 11. Pubic support- 1 no. 12. Tunnel cushion for leg in lateral position: 1 no. 13. Gel pad head rings medium – 1no. 14. Gel pad head rings large – 1no. 15. Gel chest rolls – 2nos. 16. Large arm board for hand surgery- 1no.	9 sets
3.	Gynae and Urology Accessories for Lithotomy position 1. Seat Plate/Foot extension - 1 No. 2. TUR Drainage Set - 1 No. 3. Lithotomy rods with Clamp - 2 Nos. 4. Lithotomy leg holders (adult) – 1 set	4 sets
4.	Neurosurgery Accessories 1. Connection Bracket : 2 No. 2. Basic Unit for Skull Clamp : 1 No. 3. Adapter for Skull Clamp : 1 No. 4. Doro/Maryland Skull Clamp : 1 No. 5. Adult and Child Pins : 1 No. 6. Horseshoe head rest two-parts : 1 No. 7. Spinal pads : 1 No	1 set
5.	Accessories for Operating in Prone Position 1. Prone face support: 1 no. 2. Prone arm support – 2 nos. 3. Prone chest support – 1 no. 4. Prone pelvic support – 1 no.	2 sets

be mentioned in price bid

. Price for each of the accessory should

C. Thane

Pop

Subin

[Signature]

[Signature]

[Signature]

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 11 Bipolar Coagulator for Microscopic Surgery - (Rfx/Event no. 3000002611)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
				as per NCI AIIMS committee recommendation	Added Para: BOQ
				as per NCI AIIMS committee recommendation	Bipolar coagulator for Microscopic surgery - 1 No.
				as per NCI AIIMS committee recommendation	Foot switch - 1 No.
				as per NCI AIIMS committee recommendation	Consumables:
				as per NCI AIIMS committee recommendation	Double Pneumatic or electrical Pedal- 1 Nos.
				as per NCI AIIMS committee recommendation	Bipolar Forceps (Nonstick- AgCu or other alloy offering nonstick effect) Bayonet shaped - 5 Nos.
				as per NCI AIIMS committee recommendation	Suitable/compatible Bipolar Cable- 5 Nos
				as per NCI AIIMS committee recommendation	Power cables - 1 Nos

Handwritten notes and signatures:
 Swidler
 [Signature]
 [Signature]
 [Signature]

Response To Pre-Bid Queries (Pre-Bid date: 16.02.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/12/17-18 Dated: 05.02.2018

Schedule No. 12 - Hands and Foot wear Disinfection System (Rfx/Event number 3000002612)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
				as per NCI AIIMS committee recommendation	Added Para: BOQ
				as per NCI AIIMS committee recommendation	Hands and foot wear disinfection system - 1 No.
				as per NCI AIIMS committee recommendation	Compatible disinfection solution - 1 No.

Jebe

m

Sanjay

Sanjay

SP

Sanjay

Sanjay

Sanjay

Sanjay