

10-07-2017

Clarification No. 01

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-III/06/CAD/2017-18 dated 08-06-2017 read with Amendment No. 01 dated 21-06-2017 & Amendment No. 02 dated 15-07-2017

This clarification has been issued to rectify certain mismatches between the BOQ and Technical Specifications for the following schedule.

**Section VII
Technical Specification**

Sch 02 - Single plane cardiovascular Catheterisation with digital subtraction angiography lab

Event Number: 3000002005

Existing:

Essential accessories:

6. Radiation protective apron (US FDA Approved/European CE; Lead-Free, light weight) of high quality with hangers: (Total quantity 20: Front type - 10, Wrap around – 6 and Two-piece type - 4). It should be state of art light weight with a lead equivalent of 0.5mm.

Read as:

Essential accessories:

6. Radiation protective apron (US FDA Approved/European CE; Lead-Free, light weight) of high quality with hangers: (Total quantity 20: Front type - 10, Wrap around – 6, Two-piece type – 4 and **Lead Cap - 4**). It should be state of art light weight with a lead equivalent of 0.5mm.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**