

HLL/BME/ ICCONS - 05/15-16/Amt.-01

Dated: 22-12-2015

Amendment No:1

to

Ref: Tender no. HLL/BME/ ICCONS - 05/15-16 Dated: 10.12.15

For supply of medical equipment to ICCONS Hospital, Shoranur

The following amendments are incorporated in the referred tender enquiry document:

In Technical specification:

Digital Video EEG/ERP/EP System

Page No.	Para No.	Tender Specification	Amended As
45	G. 1	Adult & Paediatric EEG caps - 1 Set each as STANDARD SUPPLY and another 1 Set each should be quoted as OPTIONAL.	Adult & Paediatric EEG caps of 128 Channels - 1 Set each as STANDARD SUPPLY and another 1 Set each should be quoted as OPTIONAL. Bidder should also quote 1 set of 32 Channel-Neonatal EEG Cap as OPTIONAL.

All other contents, terms & conditions of the tender enquiry remain unchanged.