

13.03.2019

Amendment No. 08**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/10/18-19 DATED 13.12.2018.**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII
Technical Specification****Schedule No. 01
CT 128 Slice (Rfx No. 3000003617)**

Tender Page & Para	TENDER SPECIFICATION	Read as
Amendment No. 01 Para 3b, page 48	X-Ray Tube Anode Heat Storage Capacity of at least 6.0 MHU or direct cooling tube.	X-Ray Tube Anode Heat Storage Capacity of at least 7.0 MHU or direct cooling tube

All other contents of the tender enquiry including terms & conditions remain unaltered.**Note:**

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**