

Amendment No.10

Date: 27/10/2016

Sub: Amendment No.10 to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/11/16-17 dated: 22.08.2016.

The following changes are made to the Tender enquiry document.

Section – VII Technical Specifications

Item Sl. No. 5 Flexible Cysto-Nephroscope (High End)

1. Existing Para 1.1: The Flexible Cysto – Nephroscope with employs an **ultra-miniature** digital video CMOS sensor placed directly at the tip of the endoscope that captures full-motion video images in digital format. Illumination of the surgical site is provided by white light LEDs that are built into the endoscope. **This illumination technology eliminates the need for a separate high-intensity light source and related cables.**

Read as: The Flexible Cysto – Nephroscope with employs a digital video CMOS **or CCD** sensor placed directly at the tip of the endoscope that captures full-motion video images in digital format. Illumination of the surgical site is provided by white light LEDs/ **Xenon light source** that are built into the endoscope.

2. Existing Para 3.1: Field Of View: 110 degree or better Length: 37 cm (approx) Direction Of View: Straight forward (zero to six degrees). Working Channel: 6.0 Fr or better Distal tip Diameter: 14.0 Fr (approx).

Read as: Field Of View: 110 degree or better Length: 37 cm (approx) Direction Of View: Straight forward (zero to six degrees). Working Channel: 6.0 Fr or better Distal tip Diameter: **14.0 Fr to 16 Fr**

3. Added Para1: Video Processor- (Price to be quoted separately)

4. Added Para 2: 26” or more full HD Medical grade Monitor (Price to be quoted separately).

All other contents of the tender enquiry including terms & conditions remain unaltered.