

Amendment No. 1**25.10.19****Sub: Amendment to the Bidding Document****Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/43/19-20 dated 27.09.2019**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION – VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item no. 1 (Tender_ID: 2019_HLL_32911_1)****IORT Machine (Electron Based)**

Tender Page & Para	Existing Tender Specification	Amended as
Page 46, Para v & vi	<p>v) Calculate dose distribution using Pencil Beam algorithm taking into account electron density of patient's tissue and geometry of selected applicator (including bolus and shielding); and</p> <p>vi) Calculate dose distribution using electron Monte Carlo algorithm taking into account electron density of patient's tissue and geometry of selected applicator (including bolus and shielding)</p>	<p>v) Calculate dose distribution using Pencil Beam algorithm taking into account electron density of patient's tissue and geometry of selected applicator (including bolus and shielding); and/or</p> <p>vi) Calculate dose distribution using electron Monte Carlo algorithm taking into account electron density of patient's tissue and geometry of selected applicator (including bolus and shielding)</p>

Item no. 3 (Tender_ID: 2019_HLL_32911_3)**Mobile CT Scanner for IORT**

Tender Page & Para	Tender Specification	Amended as
Page 53, Para 1.7	The X-Ray detector system should have solid state detector with at least 1.25 mm to 10 mm detector width to generate multi slice CT images of soft tissue and bone.	The X-Ray detector system should have solid state detector with generated slice thickness ranging from 1.25 mm to 10 mm.

SECTION – VIII**QUALIFICATION CRITERIA****Existing:**

1. The bidders must be a manufacturer. In case the foreign manufacturer does not quote directly, they may authorise an Indian agent as per proforma of “Manufacturer Authorization Form” as given in the bidding document to quote and enter into a contractual obligation.

Read as:

1. The bidders must be a manufacturer. In case the foreign manufacturer does not quote directly, they may authorise an **agent** as per proforma of “Manufacturer Authorization Form” as given in the bidding document to quote and enter into a contractual obligation.

Important Note:

Technical Compliance Sheet ‘Ver 1.1’ for tender IDs **2019_HLL_32911_1 & 2019_HLL_32911_3** under this NIB are being uploaded under ‘Corrigendum’ on CPPP. The same need to be downloaded for submission.

All other contents of the Bidding Document including terms & conditions remain unaltered.