

11-09-2017

Amendment No. 1**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/13/ENDO/17-18 Dated:22.08.2017**

The following changes are being incorporated in the above referred Tender Enquiry Document

**Section VII
Technical Specification**

TECHNICAL AMENDMENT		
Tender ref: HITES/PCD/PMSSY-III/13/ENDO/17-18 Dated:22.08.2017		
SCH 01. Bone Mineral Densitometer (Rfx no. 3000002139)		
Para	Existing Tender Specification	Amended Tender Specification
A.2	Energy Switching technique	Energy Switching technique - Energy switching should be mentioned with supporting documents.
A.3	Internal reference system for pixel by pixel data calibration (no daily calibration required)	Internal reference system for pixel by pixel data calibration (either automated or by phantom)
B.1	Automated daily procedure	Deleted
C.2	Second spine and hip scan acquisition	Second spine and hip scan acquisition or by dual energy scanning in spine and hip
F.8	Laserjet color printer along with extra consumables. Price to be fixed for 5 years	Laserjet color printer along with consumables supplied for initial startup. All consumable prices to be quoted separately. Price to be fixed for 5 years.
	Added Para	Following accessories to be supplied as per AERB norms: 1. Lead Apron - 2 nos 2. lead protection barrier with viewing window - 1 no 3. Lead lining of door and window. Installation has to be done as per AERB norms.
SCH 02. Insulin Pump with Integrated CGMS (Rfx no. 3000002147)		
Para	Existing Tender Specification	Amended Tender Specification
3	Battery :Rechargeable battery with shelf life of 2 weeks or more with total life of battery of minimum 6 months.	Battery : Rechargeable battery with shelf life of 2 weeks or more with total life of battery of minimum 6 months. Battery to be replaced every 6 months for initial 2 years.
17	Provide extra battery for backup & data cable.	One extra data cable to be provided.

16	Accessories: a) Standard insulin reservoirs/cartridges b) Software and data cable for downloading data from pump /CGM to computer c) Tubing/Infusion sets d) Glucose sensors	Accessories: Software and data cable for downloading data from pump/ CGM to computer Consumables to be supplied initially (Price of all these items should be quoted separately & it will be valid for 5 years): a) Standard insulin reservoirs/cartridges - 30 Nos b) Tubing/Infusion sets - 30 Nos c) Glucose sensors - 15 Nos
	Added para:	Accessories will be covered under warranty however, consumables & single use items will not be covered under warranty.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids