

**Amendment No. 2**

**Date: 15/12/2014**

**Sub: Amendment to Tender Enquiry Document.**

**Ref: NIT No. HLL/PCD/GNCTD/20/GNEC/14-15 dated 05/11/2014  
Amendment No. 1 dated 08/12/2014**

The following changes have been incorporated in the referred NIT.

**Item No. 14  
Anaesthesia Workstation**

- 1. Existing Para B 8:** Should have permanent Oxygen sensor (paramagnetic type).  
  
**Read as: Should have paramagnetic/galvanic sensor and should covered under warranty & CMC.**
  
- 2. Existing Para G 2: Voltage corrector stabilizer of appropriate ratings meeting ISI specification (input 160-260 V ac and output 220-240 Vac and 50 Hz).** UPS of suitable rating shall be supplied for minimum 1 hour backup time for the entire system.  
  
**Read as:** UPS of suitable rating shall be supplied for minimum 1 hour backup time for the entire system.
  
- 3. Existing Para:** Annual supply in 2nd, 3rd, 4th and 5th year after installation:  
  
**Read as: Annual supply as per point no. 1-15 is to be supplied per year at the beginning of 2<sup>nd</sup>, 3rd, 4th and 5th year after installation**

**All other contents of the tender enquiry including terms & conditions remain unaltered.**