

MINUTES OF THE MEETING

**PRE BID MEETING OF SUPPLY, INSTALLATION, COMMISSIONING AND
VALIDATION OF ONLINE PARTICLE COUNTING SYSTEM AT PASTEUR INSTITUTE OF
INDIA, COONOR**

Document No. : NPI/110831/EQP/TD/OPC/01 dated 02.06.2017

Venue : HLL Biotech Limited, IVC, Chengalpattu

Date : 08.06.2017

Project : Pasteur Institute of India, Coonor

Attendees : Enclosed As Annexure

Issued by : Mr. Suresh S (DGM – Procurement)

Issued on : 09.06.2017

Issued from : HLL Biotech Limited, Chengalpattu

MINUTES OF THE MEETING


**PRE BID MEETING OF SUPPLY, INSTALLATION, COMMISSIONING AND
 VALIDATION OF ONLINE PARTICLE COUNTING SYSTEM AT PASTEUR INSTITUTE OF
 INDIA, COONOOR**

S.No.	Clarifications on queries
	Discussion on Tender Enquiry Document: NPI/110831/EQP/TD/OPC/01 dated 02.06.2017
1.	General Technical Specifications: General Points – Point 4. Pg 52 of 90 in the tender document: point (j) is deleted.
2.	Proforma (D: Quality) Pg: 58 of 90 : 3.6: ISO 14664 shall be read as ISO 14644.

Technical Clarifications:

S.No	Datasheet Point number and excerpt*/ description of the specification *	Point changed as / Comment
1.	2.3: Zero Count Filter: 0.2 micron absolute with tubing on all sensors.	2.3: Zero Count Filter: The filter should ensure zero count level to meet ISO: 21501-4 standards.
2.	3.4: Tube: cGMP/FDA compliant.	3.4: Tube: Vendor shall supply Bevelon or Super thin type tubes.
3.	4.5: Flow Rate Indicator: Differential Pressure Transducer with LED indicator	4.5: Flow Rate Indicator: Differential Pressure Transducer with LED indicator / Inbuilt type indicator.
4.	CMC clause in the Tender Document:	As calibration cost was not clearly mentioned in the CMC clause, vendor shall consider calibration cost for each year in 5 years in CMC cost.

For HLL Biotech Limited


 9/6/17.

Authorized Signatory