

**Response To Pre-Bid Queries (Date of Pre-Bid Meeting: 26-Nov-2020)**

| Sl. No. | Tender Clause No & Page No   | Name of Firm                 | Representation Received from the Firms   | Committee Recommendation  |
|---------|--|------------------------------|--|---|
| Ref.    | <b>Bid Number: GEM/2020/B/882203 Dated: 19-11-2020 - Deep Freezer For Universal Immunization Program</b><br><b>Bid Number: GEM/2020/B/881345 Dated: 19-11-2020 - Solar direct drive combined refrigerator/freezer(small)</b><br><b>Bid Number: GEM/2020/B/883362 Dated: 19-11-2020 - Walk in Freezer (20 cum.)</b><br><b>Bid Number: GEM/2020/B/884231 Dated: 20-11-2020 - Walk in Cooler (16.5 cum.)</b><br><b>Bid Number: GEM/2020/B/884873 Dated: 20-11-2020 - Walk in Cooler (32 cum.)</b><br><b>Bid Number: GEM/2020/B/886407 Dated: 21-11-2020 - Walk in Cooler (40 cum.)</b>  |                              |  |   |
| 1       | Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion. | Remi Elektrotechnik Ltd.     | <b>Past Performance:</b> 50%<br><br><b>Query &amp; Suggestion-</b> Past Performance to be deleted<br><br><b>Justification :</b> We understand that under "Atmanirbhar Bharat" Scheme of Govt. of India, past performance is not mandatory in tenders for evaluation Purpose.   | No Change   |
| 2       | If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed  |                              | <b>Query &amp; Suggestion-</b> MSME including Medium Enterprises having UAM needs to be considered for EMD, Past performance and Purchase preferences.<br><br><b>Justification :</b> This will encourage Medium Enterprises also to take part and offer competitive pricing.   | Exemption shall be provided as per MSE policy.  |
| 3       | Consignee list   |                              | <b>Query &amp; Suggestion-</b> Bidder Should be allowed to quote on Zone wise basis for any or all zones. Each Zone should be treated as separate Tender for Evaluation Process.<br><br><b>Justification :</b> We understand that under "Atmanirbhar Bharat" Scheme of Govt. of India, past performance is not mandatory in tenders. | No Change   |
| 4       | As per the bid documents delivery period is 90 days  | Godrej & Boyce Mfg. Co. Ltd. | it should be at least 180 day as we have to deliver the product till end users.  | No Change   |
| 5       | As per tender condition 75% payment can be claimed after CRC submission and rest 25 % after FAC submission   |                              | As we have to deliver the product till end location hence 100 % payment should be processed after CRC submission.  | No Change   |
| 6       |  | B Medical System             | Can you also kindly confirm that there is NIL EMD and tender fee for this tender?  | Yes, There is no EMD and Tender Fee for this tender. However bidders has to provide a "Bid Security Declaration". |
| 7       | Under Bid Document- E-PBG detail   | Rattan Sales Corporation     | 3. The Performance Bank Guarantee should be for 60 months instead of 62 months. As our and few other banks were not issuing the PBG more than 60 months as per their protocols. Or We will extend the PBG for another 2 months before expiry of 60 months later on. For which we will provide declaration/undertaking for the same.  | No Change   |
| 8       | We request that delivery and installation site should be ready before the delivery of WIC & WIF. As this is a very major problem for these types of equipment, which is to be assemble and install at the installation site.   |                              |  |   |

**Response To Pre-Bid Queries (Date of Pre-Bid Meeting: 26-Nov-2020)**

| Sl. No. | Tender Clause No & Page No   | Name of Firm                           | Representation Received from the Firms | Committee Recommendation              |
|---------|--|--|--|---------------------------------------|
| Ref.    | <p><b>Bid Number: GEM/2020/B/882203 Dated: 19-11-2020 - Deep Freezer For Universal Immunization Program</b><br/> <b>Bid Number: GEM/2020/B/881345 Dated: 19-11-2020 - Solar direct drive combined refrigerator/freezer(small)</b><br/> <b>Bid Number: GEM/2020/B/883362 Dated: 19-11-2020 - Walk in Freezer (20 cum.)</b><br/> <b>Bid Number: GEM/2020/B/884231 Dated: 20-11-2020 - Walk in Cooler (16.5 cum.)</b><br/> <b>Bid Number: GEM/2020/B/884873 Dated: 20-11-2020 - Walk in Cooler (32 cum.)</b><br/> <b>Bid Number: GEM/2020/B/886407 Dated: 21-11-2020 - Walk in Cooler (40 cum.)</b></p>   |  |  |                                       |
| 9       | <p>We request purchasing authority, that after submission of all documentation part, for claiming payment as per NOA, the payment should be released within 7-10 days' time. As it is very much required for proper flow of payment during this COVID19 time.</p>  |  |  | <p align="center">No Change</p>       |
| 10      | <p>As there is only one lab in India which is WHO certified/NABL for carrying out the required test WHO/PQS/E03/FZ03.1 (or its updated version as per WHO PQS from time to time) for Deep freezers. The lead time for required test is minimum 4 to 6 weeks and due to COVID-19 Pandemic situation, there is difficulties to schedule the test immediately. Tender has been floated on date 19th Nov 2020 and submission date 19th Dec 2020 for the tender. Span is only of 30 days. It is very difficult to carry out such crucial test in such a shot span. So we hereby request kindly allow us period of 4 to 6 weeks to submit the Test Report as per standard WHO/PQS/E03/FZ03.1 from WHO certified lab/ NABL/ STQC lab.</p> | <p align="center">Mark Enterprises</p> |  | <p align="center">Refer Amendment</p> |

3. **Walk-in-Freezer (WIF):**The TSDG recommendations for WIF were discussed during TSC meeting.

- i. The specification Minimum thickness, Storage temperature, Gross internal volume, "Calibration of Thermostat" and Environmental factors had some typographical error. So all may be amended as per existing MoHFW approved specification
- ii. The recommendations with TSC approvals are as follows:

| Tender Page & Para | Specification Name  | Name Of The Firm  | Representation Received From The Firms            | TSC Recommendations | Justifications  |
|--------------------|---|-------------------|---|---------------------|---|
| Page No. 3         | Density (min)(kg/m3)  | Hitch Associates  | ±2 kg/m3  | Change              | Tolerance Suggested is acceptable subject to the submission of lab certification ±2 kg/m3   |
| Page No. 3         | Outer and inner panels  | Hitch Associates  | 0.5mm   | Change              | No change in specifications. However, if sheet thickness is mentioned in mm it must be close to 22 SWG and Bidder has to provide DATA sheet accordingly and approved lab certification.   |
| Page No. 3         |   | Crystal Logistics | Need to know which material is required for panel |                     |   |
| Page No. 3         | Roof and panel skin material  | Hitch Associates  | Inner side 0.5mm                                  | No Change           | No change, refer MoHFW approved specifications  |
| Page No. 3         |   | Crystal Logistics | Need to know which material is required for panel |                     |   |
| Page No... 3-4     | Construction  | Hitch Associates  | Solid concrete base (whom scope)                  | Change              | <b>May be amended as:</b><br>"Concrete base of 200 mm is under the scope of purchaser, however providing the insulated base for WIC", either in the form of prefabricated floor panel or layered structured" will be sole responsibility & under the scope of bidder. |
| Page No. 4         | Internal temperature  | Hitch Associates  | • -20 deg C                                       | Change              | May be amended as per MoHFW approved specifications:<br>-25Degree C to -15Degree C adjustable (i.e. during +43Degree C continuous ambient , ii. 32Degree C continuous ambient, iii. 43/05Degree C day / night cycling temperatures.)                                  |
| Page No. 4         | Depending upon the internal room location refrigeration units may be one of the following types | Hitch Associates  | <b>Condenser Wall/floor mounted</b>               | No Change           | No change, refer MoHFW approved specifications  |

| <b>Tender Page &amp; Para</b> | <b>Specification Name</b>                       | <b>Name Of The Firm</b> | <b>Representation Received From The Firms</b>   | <b>TSC Recommendations</b> | <b>Justifications</b>   |
|-------------------------------|---|-------------------------|---|----------------------------|---|
| Page No. 4                    | Storage condition                               | Hitch Associates        | -20 deg C   | Change                     | In case of WIF Storage conditions to be maintained at -20Degree C $\pm$ 2Degree C continuously, control by thermostat on each refrigeration unit.   |
| Page No. 4                    | Base 1st layer                                  | Crystal Logistics       | We suggest 60 mm Cement & 15mm Plywood for checker Plate  | No Change                  | No change, refer MoHFW approved specifications  |
| Page No. 4                    | 3rd layer of specified                          | Hitch Associates        | <ul style="list-style-type: none"> <li>• 2MM</li> <li>• Not advisable to put <b>Aluminum</b> chequered on concrete base.</li> <li>• Ready made floor with top of <b>Aluminum</b> chequered plate</li> </ul> | No Change                  | No change, refer MoHFW approved specifications  |
| Page No. 5                    | Condensing unit (s) to comprise compressor with |                         | To be updated as per approved specification   | Change                     | <b>In case of WIF it May be amended as:</b><br>Forced air condenser, Oil level glass, Oil separator, Liquid receiver to carry full charge, Filter/dryer with flare connections, Isolating stop valves, Fixed high and low pressure dial gauges, Fitted with high and low pressure cut-outs, Time-operated electric defrost control. It should also have run hour meter. |
| Page No. 5                    | Refrigeration system feature                    | Crystal Logistics       | Define refrigerant name   | No Change                  | CFC free refrigerant already mentioned  |
| Page No. 6                    | Temperature Monitoring recording unit feature   | Crystal Logistics       | If not Thermal paper than printer to be attached and data to be extracted using USB   | No Change                  | Wall mounted seven days graphic temperature recorder with pen recording mechanism.  |
| Page No. 6                    | Temperature Control, monitoring & Recording     | Hitch Associates        | *-20  | Change                     | <b>May be amended as:</b><br>All parts of the room designated for vaccine storage must remain between -25Degree C to -15Degree C when measured under any loading condition between empty and full and over the full ambient temperature range of the required temperature zone  |
| Page No. 7                    | Installation                                    |                         | To be updated as per approved specification   | Change                     | <b>May be amended as:</b><br>Complete Installation, testing and commissioning is to be done by supplier inclusive of - Installation of stabilizer, Drain system, Assembly of the panels, Refrigerator units, Data logger, Adequate smoke evacuation system, Generator as per <b>CPCB guidelines</b> . All other related work required for                               |

| <b>Tender Page &amp; Para</b> | <b>Specification Name</b>  | <b>Name Of The Firm</b> | <b>Representation Received From The Firms</b> | <b>TSC Recommendations</b> | <b>Justifications</b>  |
|-------------------------------|----------------------------|-------------------------|---|----------------------------|--|
|                               |                            |                         |   |                            | installation as per WHO PQS and guidelines   |
| Page No. 8                    | Alternator poer supply     |                         | To be updated as per approved specification   | Change                     | <b>May be amended as:</b><br>Alternator <b>power</b> supply : 240/415V, Single/Three phase, 50 Hz  |
| Page No. 10                   | 5 year Warranty            | Hitch Associates        | 1 year warranty 5 years CMC                   | No change                  | No change in warranty period   |
| Page No. 10                   | Servo voltage Power Supply | Hitch Associates        | CE certified                                  | No change                  | No change, refer MoHFW approved specifications   |
| Page No. 11                   | 5 year Warranty            | Hitch Associates        | 1 year warranty, 5 years CMC                  | No change                  | No change in warranty period   |
| Page No. 12                   | Certification and manual   | Hitch Associates        | NABL  | Change                     | <b><u>NABL is already there.However for more clarity to be amended as below:</u></b><br>At the time of prototype inspection "Test certificate" as per WHO quality assurance protocol WHO/PQS/E001/CR-FR01-VP1.2 is mandatory from an independent laboratory approved/ recognize byWHO/NABL/STQC laboratory Separate certificate of inspection for tendered items from an independent laboratory approved / recognize by WHO/NABL//STQC Labs or 3rd party inspection agency as mentioned in the NOA is essential and It is required to be submitted at the time of delivery the list of important spare parts and accessories with their part number and costing. |