

02-08-2019

Amendment No. 5

Ref.: Tender Enquiry: HITES/PCD/MP/CLINICAL/RC-04/19-20 dated 01.06.2019 and subsequent Amendments

Sub: Amendment No. 5 to the referred tender enquiry

The following changes are being incorporated in the above referred Tender Enquiry Document

SECTION I**NOTICE INVITING TENDER (NIT)**

The **revised** time line for of the above referred TED is as below:

S.No.	Rfx. No.	Name of the Items	Closing date for submission of online bids Time : 13:00 hrs.	Closing date for submission of tender processing fee and EMD in physical form* Time : 14:00 hrs.	Date of opening of online bids Time : 14:30 hrs.
17	3000004099	Echocardiography machine	14-08-2019	14-08-2019	14-08-2019

SECTION - VII**TECHNICAL SPECIFICATIONS**

Sl. No. 17 – Echocardiography Machine (Rfx. No. 3000004099)		
	Existing Specification	Clarification
page 74 para 16	Peadiatric and Adult Transthoracic 3 D Probe-01 Nos. each	1. Adult Transthoracic 3 D Probe-01 Nos. 2.Peadiatric Transthoracic 2 D Probe-01 Nos.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.

- i. **If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of ender opening.**