

**AMENDMENT NO: 04 Dated 01.11.2022**

**AMENDMENT TITLE:**

Change in Technical Specification.

**REF NO:**HLL/SD/HCS/2022-23/Tender 17 dated 28<sup>th</sup> September 2022

E-TENDER FOR SUPPLY OF FULLY AUTOMATIC BIOCHEMISTRY ANALYSER – HIGH THROUGHPUT (WET CHEMISTRY) FOR HLL LIFECARE DIAGNOSTIC LABORATORIES LOCATED AT VARIOUS LOCATION PAN INDIA ON REAGENT RENTAL BASIS

The following amendment has been incorporated to the bid document for the above tender;

In ANNEXURE 04, TECHNICAL SPECIFICATION AND PARAMETERS

<b>SN</b>	<b>Existing Technical Specifications as per tender</b>	<b>Amendment</b>
16	Photometer should consist of Minimum 13 wavelengths in the range 340 to 800 nm. It should be capable of doing Monochromatic and Bichromatic measurements.	Photometer should consist of Minimum 10 or higher wavelengths in the range 340 to 700 nm or greater. It should be capable of doing Monochromatic and Bichromatic measurements.

All relevant clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

VICE PRESIDENT (SOURCING DIVISION)  
HLL LIFECARE LTD  
HLL BHAVAN, POOJAPURA  
THIRUVANANTHAPURAM