

Amendment no. 01

Date: 21.04.2015

Subject: Amendment no. 1 to the Tender Enquiry Document

Ref: Tender Ref: HLL/PCD/PMSSY-II/NAGPUR/05/14-15 dated 26.03.2015

The due date for submission of bids is revised as follows:

Section I

Notice Inviting Tenders (NIT)

For:-

Sl.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	22-April-2015,1700 hrs (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	23-April-2015, 1230 hrs IST
e	Time and date of opening of online bids	24-April-2015, 1230 hrs IST

Read as:-

Sl.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	19-May-2015,1700 hrs (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	20-May-2015, 1700 hrs IST
e	Time and date of opening of online bids	21-May-2015, 1230 hrs IST

Note: If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e, upto 02.11.2015.

For:-

IMPORTANT NOTE :- Tender fee (Rs.5,000/-) and EMD(As applicable) should be deposited in the Tender Box located at HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh on or before 24-April-2015, 1700 hrs (IST) . Submission beyond stipulated date & time would result in REJECTION of BID.

Read as:-

IMPORTANT NOTE :- Tender fee (Rs.5,000/-) and EMD(As applicable) should be deposited in the Tender Box located at HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh on or before **19-May-2015**, 1700 hrs (IST) . Submission beyond stipulated date & time would result in REJECTION of BID.

Technical Specifications

Item Sl. No. 7: Upper GI End View Gastroscope

Existing Para: Scope should be capable to produce chrome endoscopy images like NBI/SPIES/FICE etc

Read as: Scope should be capable to produce chrome endoscopy images like NBI/SPIES/FICE /I-Scan etc.

Item Sl. No. 8: HEAD-LIGHT WITH XENON COLD-LIGHT SOURCE

Existing Para 2: Adjustable spot light 15mm to 100mm

Read as: Adjustable spot light 15mm-30mm to 80-100mm

Item Sl. No. 24: Neuro Endoscope system of Neurosurgery

Existing Para 3.15 : High Definition camera with resolution 1920x1080p integrated image processing module, color systed: pal, power supply: 100-240 VAC, 50/60 Hz, including: 3 chip camera head with optical zoom lens, camera control unit with integrated image processing module, connecting cable length 180 cm, connecting cable set length 180 cm, key board or touch screen controlled camera console, 2 connecting cables, for connecting video printers or recorders-1 nos.

Read as : High Definition camera with resolution 1920x1080p integrated image processing module, color systed: pal, power supply: 100-240 VAC, 50/60 Hz, including: **3 chip CCD HD camera head** with optical zoom lens, camera control unit with integrated image processing module, connecting cable length 180 cm, connecting cable set length 180 cm, key board or touch screen controlled camera console, 2 connecting cables, for connecting video printers or recorders-1 nos.

The subsequent amendment related to other schedules will be uploaded shortly.

All other terms and conditions of the tender enquiry remain unaltered.