

Amendment No. 4

Date: 21.02.2017

Subject: Amendment no. 4 to the Tender Enquiry Document

Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY-II/NAGPUR/18/16-17 dated 21.12.2016

Section I

Notice Inviting Tenders(NIT)

Existing para 12. h.

8 x medical gas terminal & Hose assembly

(2 x Oxy, 2 x Vac, 1 x Air (4bar), 1x N2O, 1 Air (7bar), 1 CO2). Provision for AGSS outlet & AGSS piping should be available in the pendant also. Inside OT medical gas pipeline should be done by the bidder for all the gases including AGSS and should be connected (or should be drawn till outside of OT, incase existing medical gas pipeline for any gas is not available) with existing medical gas pipeline available outside OT.

Read as:

10 x medical gas terminal & Hose assembly

2 x Oxy, 2 x Vac, 1 x Air (4bar), 1x N2O, 1 Air (7bar), 1 CO2, **1 AGSS, 1 extra (type of outlet shall be as per the requirement of the client at the time of Installation & Commissioning)**. Inside OT medical gas pipeline should be done by the bidder for all the gases including AGSS and should be connected (or should be drawn till outside of OT, incase existing medical gas pipeline for any gas is not available) with existing medical gas pipeline available outside OT.

All other terms and conditions of the tender enquiry remain unaltered.