

Date: 24-04-2019

Amendment No. 5**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/38/Mix/18-19 dated 01/03/2019 read with Amendment No.01, 02, 03, 04 Dated- 06.03.2019, 07.03.2019, 28.03.2019, 15.04.2019 respectively.**

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION – VII
TECHNICAL SPECIFICATIONS

HITES/PCD/PMSSY-III/38/Mix/18-19 Dated: 01.03.2019		
Item No.29 Video Polysomnography with 1 Camera		
Para	TENDER SPECIFICATION	READ AS
Page 94 Para 3	Amplifier must also have built-in Oximetry (SPO2), patient event button and photic connectivity with the amplifier box/ PC	Amplifier must also have built-in Oximetry (SPO2), patient event button and photic connectivity within the amplifier box
Pg 94 Para 6	Amplifier must offer Ethernet/USB/ Wireless connectivity to acquisition PC with hi speed data transfer	Amplifier must offer Ethernet/USB connectivity to acquisition PC with hi speed data transfer.

All other contents of the tender enquiry including terms & conditions remain unaltered.**Note:**

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**