

## **Amendment No. 3**

**Date: 26/06/2014**

**Subject: Amendment to the tender Enquiry Document**

**Ref: Tender Enquiry No.: HLL/PCD/HP/03/14-15 dated 19/05/2014 read with Amendment No.2 dated 23/06/2014**

In pursuant to the prebid discussions held on 27.05.2014, the following amendments have been authorised and incorporated in the referred tender enquiry.

### **Section – VII**

#### **Technical Specifications**

##### **Item No. 1**

#### **FOR REMOTE ECG ON MOBILE SYSTEM**

**1. Existing -**

Para: Inbuilt GSM/GPRS modem for ECG transmission

Para: Software support for all types of smart phones (Android, Black Berry, Windows, I Phone)

**Read As**

GSM/GPRS modem for ECG transmission (if it is through mobile, both transmitter and receiver mobile should be supplied. Make and model of the mobiles to be mentioned)

**2. Existing-**

Para: Two way transmission should be possible i.e. From Technician to Doctor and from Doctor to Technician

Para: Support web portal for anywhere, anytime access of ECG

Para: Support Auto-interpretation & Manual reporting from web portal

**Read As**

Para: Two way transmission should be possible i.e From Technician to Doctor and from Doctor to Technician

Para: Support web portal for anywhere, anytime access of ECG

Para: Support Auto-interpretation & Manual reporting from web portal

**The server and its maintenance is responsibility of the bidder and maintenance cost of the server (if any) should be quoted separately on yearly basis (after warranty period) and this cost shall be loaded for 3 years (say) for evaluation and ranking purpose**

**Section IX**  
**Qualification Criteria**

**3. Existing:**

Para. 2a: The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, at least 33% of the quoted quantity of the similar equipment meeting major parameters of technical specification which is functioning satisfactorily. (For equipment which are consumable in nature, as identified in the list of requirement, proof of delivery/acceptance by consignee/purchaser shall also be considered acceptable)

**Read as:**

The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, at least 33% of the quoted quantity of the similar equipment meeting major parameters of technical specification which is functioning satisfactorily. **Proof of acceptance and/or satisfactory installations issued by consignee/purchaser shall be accepted for this qualification criterion.**

**4. Existing:**

Para. 2b: The Tenderers quoting as authorized representative of the manufacturer **meeting the above criteria 2 (a) should have executed at least one contract** in the last five years from the date of tender opening of similar equipment meeting major parameters of technical specification which is functioning satisfactorily, anywhere in India of the same manufacturer

**Read as**

The Tenderers quoting as authorized representative of the manufacturer **meeting the above criteria 2 (a) should have executed at least one contract** in the last five years from the date of tender opening of similar equipment meeting major parameters of technical specification which is functioning satisfactorily, **anywhere in India of any manufacturer. Proof of acceptance and/or satisfactory installations issued by consignee/purchaser shall be accepted for this qualification criterion.**

**All other contents of the tender enquiry including terms & conditions remain unaltered.**