

NOTICE INVITING TENDER

TENDER FOR THE SUPPLY, INSTALLATION, TESTING & COMMISSIONING AND ONSITE SUPPORT FOR MEDICAL EQUIPMENTS AND MATERIAL TO COTE D'IVOIRE (IVORY COST)

Tender No. HLL/SD/RBD/2018-19/TENDER/10 DT 24.10.2018



HLL Lifecare Limited
(A Govt. of India Enterprise)
HLL Bhavan, Poojappura,
Thiruvananthapuram -695012
Kerala, India
Tel: +91 471 2354949, 2350959
Website – www.lifecarehll.com

HLL LIFECARE LIMITED
(A Govt. of India Enterprise)
HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012, Kerala, India
Tel: +91 471 2354949, 2350959, 2350961, 2356352.
Website – www.lifecarehll.com

TENDER FOR THE SUPPLY, INSTALLATION, TESTING & COMMISSIONING AND ONSITE SUPPORT FOR MEDICAL EQUIPMENTS AND MATERIAL TO COTE D'IVOIRE (IVORY COST)

TENDER NO.	:	HLL/SD/RBD/2018-19/TENDER/10 DT 24.10.2018
DATE OF COMMENCEMENT OF BIDDING	:	24.10.2018, 02:00pm
LAST DATE AND TIME FOR SUBMISSION OF BIDS	:	13.11.2018, 03:00pm
TIME AND DATE OF OPENING OF TECHNICAL BIDS	:	13.11.2018, 04:00pm
PLACE OF SUBMISSION AND OPENING OF BIDS	:	HLL Lifecare Limited Corporate & Registered Office HLL Bhavan, Poojappura, Thiruvananthapuram -695012 Kerala, India
ADDRESS FOR COMMUNICATION	:	HLL Lifecare Limited Corporate & Registered Office HLL Bhavan, Poojappura, Thiruvananthapuram -695012 Kerala, India Email – sdrbdsouth@lifecarehll.com Ph: 0471 2354949

TABLE OF CONTENTS

SECTION I: INVITATION FOR BIDS (IFB)	4-10
SECTION II: INSTRUCTIONS TO BIDDERS (ITB)	11-17
SECTION III: SPECIAL CONDITIONS OF CONTRACT (SCC)	18-20
SECTION IV - TECHNICAL SPECIFICATION	21-34
Annexure - I	35
Annexure - II	36-37
Annexure - III	38
Annexure - IV	39
Annexure - V	40-42
Annexure - VI	43-45
Annexure – VII	46
Annexure – VIII	47-48
Annexure - IX a,b,c	49-51
Annexure - X	52
Annexure – XI	53
Annexure – XII	54-60

SECTION I

INVITATION FOR BIDS (IFB)

HLL LIFECARE LIMITED (HLL), a 'Mini Ratna' Public Sector Enterprise under the Ministry of Health & Family Welfare, Govt. of India, on behalf of Ministry of External Affairs (MEA) invites Bids (Tender) on two bid system for the Supply, Installation, Testing & Commissioning and Onsite Support Of Medical Equipments and material as mentioned in Schedule A for project at Cote D'Ivoire (Ivory Cost).

Schedule A - Cote D'Ivoire (Ivory Cost)

Serial No.	Description	Qty (nos)
1	Mucosity vacuum cleaner / Suction cup	72
2	Autoclave Cocotte/ Table top Autoclave 20-25 litre capacity	31
3	Basin shape plastic slipper (plastic basin), Plastic bowl, Urinal plastic women	9
4	Suture box (With Instruments & SS box):	2
5	Complete Delivery box (With Instruments & SS Box)	147
6	Episiotomy box (With instruments & SS Box)	85
7	Care trolley (baby carrying bassinet)	1
8	Step ladder (2 steps) / Foot Step	21
9	Insufflator (resuscitator) manual baby	1
10	Inpatient bed + accessories / Bed observation + mattress	4
11	Stainless steel kidney tray / Stainless steel trays with lids / sterilization drums	6
12	Baby scale with direct display	18
13	Weigh Machine(150 kg)/ Bathroom scales	20
14	Serum hanger (IV Stand) / Stem on wheel	12
15	Waste Bin Pedal Operated(20L)-SS	1
16	Double head stethoscope/Standard bi- auricular stethoscope	7
17	Fetal Doppler Echo	15
18	Delivery Bed (Childbirth bed/Table)	40
19	BP Machine (Adult)	1
20	BP machine (Pediatric)	1
21	Infrared Thermometer	1
22	Tensioning device (Salters scale and complete tripod stand Hanging)	31
23	Neonatal vacuum cleaner	14
24	Basin	76
25	BedPan	8
26	Box of dressing	1
27	Stainless steel bin/ Average stainless steel bin INOX	8
28	Heating Lamp	66

29	Gynecological / Examination Light Stands	42
30	Delivery Kit	65
31	Pediatric Examination Table (Pediatric Bed + Mattress)	15
32	Mask and new-born balloon:	18
33	Mattress (adult)	19
34	New-born mattress	1
35	Endo-uterine vacuum cleaner	2
36	Neonatal Resuscitation Unit	4
37	Monitors (Patient Monitors)	1
38	Oxygen (SpO2 monitor)/Pulse-oxymeter	29
39	Bed Side Screen/ Three part Screen:	35
40	Nursing tray with lid	4
41	Medium care tray with lid	2
42	Stainless Pedal Wastebin	142
43	Poupinel Sterilizer / Dry Heat Sterilizer	10
44	Sterile speculum	27
45	Obstetrical stethoscope	15
46	Gynaecological examination table	7
47	Heating table (Warmer)	9
48	Neonatal Open Care System	7
49	Sphygmomanometer	22
50	Spengler blood pressure monitor	6
51	Stadiometer(Height Measurement)	5
52	Infantometer (Baby Height Measurement)	19

1. Contact Information

Senior Manager (SD-RBD)
HLL Lifecare Ltd.
(A Govt. of India Enterprise)
Corporate & Registered Office
HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012,
Kerala, India Tel: +91 471 2354949 (extn: 242)
Email –sdrbdsouth@lifecarehll.com
Website – www.lifecarehll.com

2. Critical Date Sheet

Tender documents may be downloaded from www.lifecarehll.com as per the critical date sheet below:

Published Date	24.10.2018
Commencement of Bid Document	24.10.2018, 02:00 pm

Bid Submission Start Date	24.10.2018, 02:30 pm
Closing date & time for submission of bids	13.11.2018, 03:00pm
Technical Bids Opening Date	13.11.2018, 04:00pm
Place of opening of Technical bids	HLL Lifecare Limited Corporate & Regd Office HLL Bhavan, Poojappura, Thiruvananthapuram - 695012 Kerala, India

3. Submission of Bids - Two Bid System

Part 1: "Technical Bid"

- Tender fee of Rs 3,000/-** (Rupees Three Thousand only) in the form Demand Draft drawn in favour of HLL Lifecare Limited, payable at Thiruvananthapuram drawn from any nationalized bank/scheduled bank.
- Earnest Money Deposit of Rs.1,00,000 (Rupees One Lakh Only)** in the form of Demand Draft/ Bank Guarantee towards Earnest Money Deposit as mentioned at Para 6 (Section II) from any Nationalized/ Scheduled Bank, in favor of 'HLL Lifecare Ltd, payable at Thiruvananthapuram . In case of non-submission of original payment instrument for EMD, the bid shall be summarily rejected.
- Central Public Sector Enterprises/SSI Units registered with NSIC (certified copy required) shall be exempted from payment towards Earnest Money Deposit & security Deposits as per rules (upto their monetary limit). Those bidders should submit attested hard copy of their **UDYOG AADHAR MEMORANDUM** along with registration certificate. In case of non-submission of relevant document for exempting EMD and security deposit, the bid shall be summarily rejected.
- Duly filled Technical Bid with proper seal and signature of authorized person on each page of the bid submitted. The person signing the bid should be the duly authorized representative of the Firm/ Bidder whose signature should be verified and certificate of authority should be submitted as per format given at **Annexure I**. The power or authorization or any other document consisting of adequate proof of the ability of the signatory to bind the firm/ Bidder should be annexed to the bid.
- Bid Form as per format given at **Annexure II** must be submitted by the bidder along with Technical Bid.
- Declaration as per the format given at **Annexure III** must be submitted by the bidder along with Technical Bid for undertaking that the Machines that shall be supplied for Cote D'Ivoire (Ivory Cost) will not be a refurbished / duplicate equipment or acquired illegally in India / Third Country.
- The installation and warranty services are required for the quoted equipments listed in schedule A at Cote D'Ivoire (Ivory Cost). The bidder should submit the signed and scanned copy of the detail plan for providing warranty services at sites. The bidders should have its own Branch Office/ Service Centre in Cote D'Ivoire (Ivory Cost) for the quoted equipments listed in schedule A or should have arrangement to provide services through local dealer/ service provider or need to provide some mechanism to provide the warranty services. In all these cases, a signed and scanned undertaking to this effect should be submitted along with the technical bid as per format given at **Annexure IV**. The details viz. name, address, contact person, telephone / fax, email etc. should be provided along with an undertaking from this local dealer/ service provider within 60 (sixty) days of the receipt of Supply Order from HLL Lifecare Ltd or in case of some other mechanism need to submit self- attested document on Bidder letter head by describing mechanism to provide warranty services.

8. Signed and scanned copy of the Checklist as per format given at **Annexure VII** must be submitted along with Technical Bid.
9. Technical Compliance Document as per **Annexure VIII** along with detailed technical specification, make, model and leaflets/brochures should be submitted in the technical bid, failing may lead to the rejection of the bid.
10. Bidder needs to submit an undertaking as **Annexure X** and declare its consortium partners, if any, while bidding and after that bidder is not allowed to change their consortium partners.
11. Bidder and consortium partner need to submit an undertaking (attached as **Annexure XI**) that its consortium partners in this Tender will not be participating as a bidder separately. If found so, bidder and its consortium partner's bid will be disqualified
12. Signed Pre-Contract Integrity Pact - **Annexure XII**
13. Copy of Supply Orders/ Contracts/ Agreements for similar work in any foreign country, successfully executed by the bidders (and its consortium partners, if any) in last five years ending March 31, 2018.
14. **Bidder is also responsible to furnish the following relevant documents of consortium partners :**
 - GST registration certificate
 - CA audited balance sheet for last three years
 - Work Orders / Letter of Award / Completion certificate of executed projects pertaining to required work experience
 - Certificate of incorporation of consortium partner, confirming consortium partner(s)'s existence for more than 3 years
 - Copy of PAN/ Registration certificate, as applicable shall be submitted.
 - Copy of a certificate by the auditor/ CA/ CS indicating the turnover of the firm shall be submitted
 - Copy of Non Conviction Notarized self attested certificate to be submitted along with technical bid
15. The bidder should have total average turnover of Rs.10.00 Crores in last 3 (three) financial years ending 31st March 2018 (if audited) else year ending 31st March 2017. In case, bidder doesn't have required turnover or other eligibility criteria, bidder can choose to form a consortium with a partner having experience in similar areas of business as listed in the requirements. The turnover and other criteria of the consortium partner can be counted towards ascertaining the eligibility criteria.
16. Copy of the GST registration Certificate and PAN card of the bidder shall be submitted.
17. Copy of documents related with Firm Registration/ Partnership Deed/ Articles of Memorandum of Association or Proprietorship Deed, Certificate of Incorporation whichever is applicable shall be submitted.
18. Signed and scanned copy of other related documents, mentioned in the Tender document but not listed here.
19. Notarized self attested Non Conviction certificate to be submitted along with technical bid

Part II - Price Bid:

Prices are to be quoted in the prescribed Price Bid format provided along with the Tender (Annexure IXa, IXb, IXc) enquiry in the Tender document. .

The price quoted shall be considered firm and no price escalation will be permitted.

The prices quoted should be inclusive of all applicable costs like equipment cost, profit margin, freight, insurance, packing, port charges, documentation charges, demurrage charges, training cost, installation, testing & commissioning charges including manpower charges, warranty charges, loading & unloading charges, applicable statutory charges (if any) at the destination country etc till final destination as per the supply order. Applicable statutory government taxes in India (if any) like GST need to be indicated separately in the prescribed price bid format. However MEA would reimburse taxes as per the applicable rates as on the date of invoice. The packing shall be transport worthy so as to prevent their damage or deterioration to goods during transit to their final destination as indicated in this document. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit and open storage. Packing case size and weights shall take into consideration, where appropriate, and the remoteness of the Goods final destination and the absence of heavy handling facilities at all point in transit. However risk in good shall continue with supplier till goods are delivered in good condition and installed at end user's site duly certified by HLL Lifecare Ltd, end- user & Indian Mission in Cote D'Ivoire (Ivory Cost)

The price should be quoted for the accounting unit indicated in the Tender document **against each equipment and the L1 will be finalized equipment wise.**

Note:

It is the responsibility of bidder to go through the Tender document to ensure furnishing of all required documents.

4. Amendment of Tender documents

At any time prior to the deadline for submission of bids, the Purchaser may, for any reason, modify the Tender Documents and the amendments will be published in HLL website/ CPP Portal only.

5. Language of Bid

All correspondence and documents related to the bid shall be in English.

6. Submission of documents

6.1 Sealing and Marking of Bids

a) The Bidder shall prepare and seal in **separate packets** the following

- **Technical Bid**
- **Financial Bid**

Bid shall contain the Technical Bid and Financial Bid in **separate envelopes**, after duly properly marking the envelopes as **TECHNICAL BID** and **FINANCIAL BID**. All the above packets shall then be sealed in an outer envelope, duly marking the envelope as BIDS FOR IFB No. HLL/SD/RBD/2018-19/TENDER/10 DT. 24.10.2018

i) The inner and outer envelopes shall be addressed to HLL at the address given below;

Senior Manager
Sourcing Division - RBD
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.
Ph.no: 0471 2353932.
sdrbdsouth@lifecarehll.com

ii) **Bear the Invitation for Bids title and number, and the statement DO NOT OPEN**

BEFORE. (Mention the date of opening of the bid as given in the tender documents).

- iii) The inner envelopes shall also indicate the name and address of the Bidder so that the bid can be returned unopened in case it is declared late.
- iv) If the outer envelope is not sealed and marked as required, HLL will assume no responsibility for the bids misplacement or premature opening.

7. Opening and Evaluation of Technical bids:

Opening of Bids by HLL

Bids received before the dead line of the submission of the bid will be opened on the date and time of opening mentioned in this tender enquiry.

Bidders wishing to be present at the time of such opening may send their duly authorized representative. Only Technical Bids will be opened on that day.

Evaluation of Bids

The purchaser will scrutinize the technical bid for compliance to the specifications and documentation requirement as per the bid document. HLL reserves the right at its sole discretion to visit the manufacturing facility of the bidder to ascertain its capability and for verification of relevant documents as a part of technical evaluation. The bidders will be short-listed on the basis of responsiveness of technical bid and the visit report if any. The price bid of the bidders who are disqualified at the technical evaluation will be returned unopened. The short listed bidders will be informed about the time, date and venue of the price bid opening.

For opening of Financial Bid, only those Bidders qualified in the Technical bid will be considered.

HLL will examine the bids to determine whether they are complete, whether any computational errors have been made, whether required securities have been furnished, whether the documents have been properly signed, and whether the bids are generally in order. HLL will ensure that each bid is from an eligible Bidder.

Arithmetical errors will be received on the following basis. If there is a discrepancy between the unit price and the total price, which is obtained by multiplying the unit price and quantity, or between subtotals and the total price, the unit or subtotal price shall prevail, and the total price shall be corrected. If there is a discrepancy between words and figures, the amount in words will prevail. If a Bidder does not accept the correction of errors, its bid will be rejected.

HLL may waive any minor nonconformity, or irregularity in a bid that does not constitute a material deviation, provided such waiver does not prejudice or affect the relative ranking of any Bidder.

8. Opening of Commercial Bids

The commercial bids of the bidders who are short listed after Technical Evaluation only will be opened in the presence of the bidders or their authorized representative (maximum two per bidder), who choose to attend, at the time place and date to be informed later.

The authorized representative of bidders, present at the time of opening of the bids shall be required to sign an attendance sheet as a proof of having attended the commercial bid opening

The bidder's name, bid prices and such other details considered as appropriate by HLL Lifecare Ltd will be announced at the time of the opening of the bids.

Lowest bidder of each equipment listed in Schedule A would be finalized independently from all

technically qualified bidders, subject to approval of MEA.

Note: Technically accepted competitive bids ONLY will be considered for the opening of Commercial Bids.

9. Modification and Withdrawal of Bids

Bids once submitted should not be modified. However in exceptional cases where modification is inevitable, the following procedure for the same should be adopted.

Modification will be permitted only if a written notice of the same is received by HLL prior to the deadline prescribed for bid submission.

1. A Bidder wishing to withdraw its bid shall notify HLL in writing prior to the deadline prescribed for bid submission. The withdrawal notice shall:
 - a) Be addressed to HLL at the address named in the bid data sheet and bear the Contract name, and the words BID WITHDRAWAL NOTICE. Bid withdrawal notices received after the bid submission deadline will be ignored and the submitted bid will be deemed to be a validly submitted bid.
2. No bid may be withdrawn in the interval between the bid submission deadline and the expiration of the bid validity period specified.
3. In the event of the date specified for bid receipt and opening being declared as a closed holiday for purchaser's office, the due date for submission of bids and opening of bids will be the following working day at the appointed times.
4. The purchaser may, at its discretion, extend this deadline for submission of bids by amending the Tender Documents or any other reasons, in which case all rights and obligations of the Purchaser and Bidder previously subject to the deadline will thereafter be subject to the deadline as extended, in HLL website.
5. Purchaser will not be held responsible for the postal delay, if any, in the delivery of the tender document or the non-receipt of the same. Bids sent by Telex/Fax/Telegraph will not be accepted.
6. HLL reserves the right to club or split the items of works, change the qualifying criteria at their discretion and to reject the bid or cancel the tender without assigning any reason thereof.

END OF SECTION I

SECTION II

INSTRUCTIONS TO BIDDERS (ITB)

1. Scope of Work / Project Timelines

The bidder is required to Supply, Install, Test & Commission and has to provide on-site support of equipment/s awarded to the bidder in Cote D'Ivoire (Ivory Cost). The delivery and installation at site must be completed within **150 days** from the date of placement of supply order by HLL Lifecare Ltd. The shipment from India/ third country is required to be done by sea/ air/ road. The delivery at site must be completed within **120 days** including the custom clearance & local transportation to the site by the bidder. It is mandatory for the bidders who respond to this bid to meet these expectations as time is the essence of this contract and is tightly linked to HLL Lifecare Ltd plans of completing the project within the available time frame. **Bidder is also requested to provide the comprehensive warranty for 1 year after commencement of the project.** A detailed SLA will be signed between HLL and the response bidder highlighting all the terms and conditions.

Responsibilities of the Implementing Agency (HLL Lifecare Ltd):

- To identify supplier/suppliers to supply medical equipments and other required materials as specified in this Tender up to site of installation i.e: in Cote D'Ivoire (Ivory Cost).
- To monitor the installation & commissioning of medical equipments at site by selected vendor.
- To facilitate selected vendor for warranty support services for the equipment for any manufacturing defect for a period of 12 months from the date of commissioning and 15 months from the date of shipment of the equipments to project site, whichever is earlier.

Responsibilities of Recipient Country /Counterpart Organization:

- To nominate a National Project Coordinator (Focal point) to closely co-ordinate and jointly work with team of experts of selected vendor and HLL.
- Appropriate selection mechanism and mobilization of citizens of Cote D'Ivoire (Ivory Cost) for training
- Making necessary arrangements required for training
- Working space including computer with Internet connection, printer, scanner, telephone, fax, photocopier, stationary (paper, CD, etc.) and other necessary items.
- To facilitate for short / long term visa / work permit exemption certificate for the experts.
- To provide power supply of three phase/single phase at 380/220 volts, 50 cycle/sec power. The site should be adequately furnished with work benches/office space and furniture viz. tables, chairs, etc.
- To assist in providing skilled, semi-skilled and unskilled workers for installation, commissioning and training. Host shall also provide interpreter at their cost, if required.
- To provide furniture & fixtures and office equipment for the centre.
- Physical space for center along with Infrastructure and logistics support such as Civil and electrical works, provision of data communication / telecom links.
- Administration and operating expenses of the center including maintenance and upgrading of the premises, furniture, basic equipment and consumables. Etc

Responsibilities of Bidder

- Supply, Installation, Commissioning, Testing & maintenance of equipment in Cote D'Ivoire

(Ivory Cost).

- The bidder shall be solely responsible right from shipping of the equipemnt from the bidders warehouse upto the destination site and subsequent installation, testing and commissioning as per the terms and conditions of this tender. The bidder should quote accordingly in the price bid covering various cost aspects like packing of the equipment, inland transportation of the equipment to the port of loading, applicable insurance coverages, documentation requirements, demurrage etc
- Bidder is also responsible to provide necessary services during warranty period to maintain service level agreements as defined in warranty clause by providing necessary spares and visits of engineers.
- Bidders should also ensure that equipment quoted should be compatible with power requirements and should meet all equipments related safety and security standards/regulations as applicable in Cote D'Ivoire (Ivory Cost).
- Any civil alterations, cabling, flooring, safety protective furnishings, signage, interiors not made available by host country; would be responsibility of the vendor, if required.
- Bidder has to position at least 2 trained staff for each set of equipments for a month, who will also be responsible to provide trainings to master trainer
- Necessary Comprehensive Insurance may be taken for the workmen involved in all the related activities with regard to supply, installation, testing, commissioning, training etc.
- Bidder needs to provide comprehensive 1 year warranty and has to arrange necessary spare parts along with delivery of equipment for proper maintenance (quantum of spare parts will be at discretion of bidder to ensure full compliance to the to Service Level requirements envisaged in the warranty clause.
- After successful installation and training, bidder has to provide necessary manuals (booklets, hand leaves, audio and visual clips) for the reference for the Government with complete right to share with any one.
- Installation, Commissioning and acceptance certificate from the competent authorities from the host country is to be obtained in a format acceptable to HLL, host country and Indian Mission at Cote D'Ivoire (Ivory Cost) and same is to be provided to HLL with complete right to share with anyone.

2. Locations for the Supply, Installation & Warranty Services

The items as detailed in this document are required to be supplied and installed at Cote D'Ivoire (Ivory Cost). The detail of consignee & firm address of the site(s) for supply, installation, Testing & commissioning and on-site warranty support will be communicated in the Supply Order.

3. Order Placements and Release of Payment

The supply order and payment shall be released by:

HLL Lifecare Limited,
HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012, Kerala,
India Tel: +0471 2354949, 2350959, 2350961, 2356352

4. Eligible Bidders

- a. The bidder and its consortium partner(s) (if any), should be incorporated in India and registered with Government under appropriate law or Act, and should be in the business of the similar activities.**

- b. Bidder & its consortium partner(if any) should provide supporting document for incorporation, turn over, balance sheet, successful completion of work order / purchase order / Letter of award, etc as detailed elsewhere in the Tender document.**
- c. The equipments quoted shall be of Indian brand or Indian OEM (Original Equipment Manufacturer). The bidder shall quote other brands only if Indian brands are not available. The Supporting document shall be provided along with the technical document. Acceptance of the equipment other than Indian make is subject to the approval from MEA.**
- d. Bidder must have an average annual turnover of INR 10 Crores in last three years (either 2014-2015; 2015-2016; 2016-2017 or 2015-2016; 2016-2017; 2017-2018 audited).**
- e. Bidder must have experience of execution of similar projects in last 5 years of either 1 project of minimum INR 5 Crores or 2 projects of minimum INR 3 Crores each or 3 projects of minimum INR 2 Crores each or total cumulative executed projects in last 5 years (ending 31st March 2018) of minimum value INR 30 Crores; The similar work means Supply, Installation, Testing & Commissioning and Onsite support of Medical Equipments in any foreign country.**

In case, bidder doesn't have required turnover or other eligibility criteria, bidder can choose to form a consortium with a partner having experience in similar areas of business as listed in the requirements. The turnover and other criteria of the consortium partner can be counted towards ascertaining the eligibility criteria.

5. Amendment of Tender Document

- At any time prior to the deadline for submission of bids, HLL Lifecare Ltd may, for any reason, whether on its own initiative or in response to the clarification request by a prospective bidder, modify the bid document.
- HLL Lifecare Ltd at its discretion may extend the deadline for the submission of bids if the bid document undergoes changes during the bidding period, in order to give prospective bidders time to take into the consideration the amendments while preparing their bids.
- HLL Lifecare Limited reserves right to add or delete any of the equipments in Schedule A, changes in quantity of equipments and its subsequent clauses at any time prior to the bid submission deadline.
- Bidders should keep viewing HLL website (www.lifecarehll.com) for any corrigendum/ change. If any bidder misses the information published on the website and their bid is rejected, no complaint would be entertained.

6. EARNEST MONEY DEPOSIT (EMD)

6.1. The Tender documents must be accompanied by Earnest Money Deposit:

Earnest Money in the form of Demand Draft/ Bank Guarantee towards Earnest Money Deposit from any Nationalized/ Scheduled Bank, in favour of 'HLL Lifecare Ltd' payable at Thiruvananthapuram. The original instrument in respect of Earnest Money Deposit must be submitted to HLL Lifecare Ltd along with the bid document as mentioned in critical date sheet. In case of non- submission of original payment instrument for EMD, the bid shall be rejected.

EMD will not be accepted in the form of cash/ cheque / FDR or any other form except DD or BG. No interest shall be payable on EMD.

6.2. The EMD will be returned to the bidder(s) whose offer is not accepted by HLL Lifecare Ltd within 120 days from the date of opening of commercial bids. In case of the bidder(s) whose offer is accepted the EMD will be returned on submission of Performance Bank Guarantee (Refer Clause 7 of Section II). However if the return of EMD is delayed for any reason, no

interest/ penalty shall be payable to the bidder.

6.3. The successful bidder, on award of contract/ order, must send the contract/ order acceptance in writing, within 7 days of award of contract/ order, failing which the EMD will be forfeited.

6.4. The EMD shall be forfeited:

- If the bidder withdraws the bid during the period of bid validity specified in the Tender.
- In case a successful bidder, fails to furnish the Performance Bank Guarantee (Clause 7 of Section II) in lieu of performance warranty.
- If the bidder fails to furnish the acceptance in writing, within 7 days of award of contract/ order.
- If the bidder fails to supply the material.

7. PERFORMANCE BANK GUARANTEE (PBG):

The successful bidder within 30 days must submit a PBG equivalent to 10% of the order value on receipt of supply order from HLL Lifecare Ltd as per the format provided in Annexure VI of this Tender document. This Bank Guarantee should remain valid 24 (Twenty Four) months from the date of its submission to HLL Lifecare Ltd.

8. PERIOD OF VALIDITY OF BIDS

Bids shall be valid for minimum 180 days from the date of opening of price bid. Bid valid for a shorter period shall stand rejected.

HLL Lifecare Ltd may ask for the bidder's consent to extend the period of validity. Such request and the response shall be made in writing only. The bidder is free not to accept such request without forfeiting the EMD. A bidder agreeing to the request for extension will not be permitted to modify his bid.

9. INTEGRITY PACT

The Integrity pact annexed shall be part and parcel of this document, and has to be signed by bidder(s), as a pre bid obligation and should be submitted along with the technical bid. All the bidders are bound to comply the Integrity Pact clauses. Bids submitted without signing Integrity Pact will be ab initio rejected without assigning any reason. The integrity pact must be submitted in their firm's letterhead.

10. LIQUIDATED DAMAGES

If the Supplier fails to deliver any or all of the Goods or perform of services within the time period(s) specified in the Contract, the Purchaser shall without prejudice to its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to 0.5 percent of the delivered price of the delayed Goods or unperformed Services for each week of delay or part thereof until actual delivery or performance, up to a maximum deduction of 5 percent of the delayed Goods or Services contract price. Once the maximum is reached, the Purchaser may consider termination of the Contract. If the Supplier fail to comply with specific packing descriptions or instructions, the loss incurred by the purchaser on this account shall be indemnified by the supplier.

If L1 or any other parties defaults (fails to deliver goods on time) then the purchaser reserves the right to purchase the goods from L2 or higher bidder or from market at the risk and cost of supplier and if the purchase happens at a price higher than the ordered rates, the purchaser shall have the right to claim the difference upon whom order was originally placed and supplier will be under obligation to pay the same. The purchaser has the right to forfeit the performance security in the event of default. In addition the purchaser is entitled to recover the business loss suffered by the purchaser consequent to default for supplying the product.

11. PURCHASER'S RIGHT TO VARY QUANTITIES AT TIME OF AWARD

The Purchaser reserves the right at the time of award of contract to increase or decrease the quantity of goods and services originally specified in the bid document without any change in unit price or other terms and conditions

12. PURCHASER'S RIGHT TO ACCEPT ANY BID AND TO REJECT ANY OR ALL BIDS

The Purchaser reserves the right to accept or reject any bid, and to annul the bidding process and reject all bids at any time prior to award Contract award, without thereby incurring any liability to the affected bidder or bidders.

13. SUBMISSION OF BIDS

The bid shall be neat, plain and intelligible. Each page of the bid should be signed by the authorized person. They should not contain any terms and conditions, printed or otherwise, which are not applicable to the Bid. The conditional bid will be summarily rejected. Insertions, postscripts, additions and alterations shall not be recognized, unless confirmed by bidder's signature.

14. DEADLINE FOR SUBMISSION OF BIDS

Bids must be submitted to the undersigned office before the due date and time as specified in the critical date sheet.

HLL Lifecare Ltd may extend this deadline for submission of bids by amending the bid documents and the same shall be suitably notified on the websites only.

15. LATE BIDS

The bid submission would not be possible after the deadline for submission of bids.

16. BID OPENING AND EVALUATION OF BIDS

The technical bids of only the eligible bidders shall be considered for further processing (technical evaluation).

Purchaser may depute its competent officers to the premises of the bidder qualified on the basis of technical scrutiny, for on-site evaluation of the claims made in the technical bid, if deemed appropriate on purchaser's sole discretion. The bidders will be short-listed on the basis of responsiveness of technical bid as well as report of on-site technical evaluation, if any.

HLL Lifecare Ltd will open commercial bids of only the technically qualified and short listed bids at the time and date to be informed later.

17. COMPARISON OF TENDERS

The comparison of respective of Tender shall be carried out equipment wise on Delivery Duty Paid (DDP) consignee site basis.

18. INDEMNITY:

The Bidder shall indemnify, defend and hold harmless Government of India and HLL, its Affiliates, officers, directors, employees, agents, and their respective successors and assigns, from and against any and all loss, damage, claim, injury, cost or expenses (including without limitation reasonable attorney's fees), incurred in connection with third Party claims of any kind that arise out of or are attributable to (i) Manufacturer's/Bidders breach of any of its warranties, representations, covenants or obligations set forth herein or (ii) the negligent act or omission of the Manufacturer /Bidders.(iii) any product liability claim arising from the gross negligence or bad faith of, or intentional misconduct or intentional breach of this Contract by bidder or its affiliate.

19. AWARD OF CONTRACT

Evaluation of Proposals & Award Criteria

The bidder can go ahead with any one or more equipments listed in Schedule A. The lowest price criteria shall be applied equipment wise on the total composite amount (inclusive of all charges except customs duty at destination country) for the equipment quoted **[Annexure IX(a)]**. The price quoted for CMC [Annexure IX(b)], Consumables and spares [Annexure IX(c)] (if any) will not be taken for arriving the L1 price. However these rates will be finalized for future requirements of supplies and the requirement will be intimated to the party either by HLL or MEA or by the host in the destination country and this price will be valid for period of four years after the warranty period.

Preliminary scrutiny of the proposal will be made to determine whether they are complete, required bid security (EMD) have been furnished, whether the documents have been properly signed and whether the bids are generally in order. Proposals not conforming to such preliminary requirements will be prima facie rejected.

Bids complying with all the eligibility requirements mentioned under Section III Clause 4 of this Tender document and fulfilling the specifications and schedule of requirements mentioned in (please refer annexure VIII for sheet) Section V shall be treated as substantially responsive bids. Responsiveness of the bids shall be determined on the basis of the contents of the bid itself and shall not be determined by extrinsic evidences.

HLL Lifecare Ltd, if required, may ask bidders for presentation on the solution offered. Failure on part of bidder to arrange the presentation on the date & place fixed by HLL shall result in the rejection of technical bids and financial bids of these bidders shall not be opened. Also, if it is found after presentation that the solution offered is not meeting the specifications prescribed by, such bidders shall be treated as substantially non- responsive. HLL Lifecare Ltd decision shall be final in this regard. The place for presentation shall be conveyed to the bidders at an appropriate date.

HLL reserves the right at its sole discretion to visit the manufacturing facility of the bidder to ascertain its capability and for verification of relevant documents as a part of technical evaluation.

Commercial bids of only those bidders will be opened who are found to be substantially responsive and the work shall be awarded to the commercially lowest bidder.

The bidder should carefully cross check the prices entered in figures with corresponding figures converted in words. In case of discrepancy between words and figures, the rates quoted in words shall be treated as final. The correct amount will be calculated by multiplying unit price with quantity and in case of any discrepancy, the corrected amount shall be considered and total of all corrected amount shall be bidder's total quoted amount.

In the copies of supply order/ contract/ agreement/ experience certificate submitted by the bidder, if the currency is other than Indian Rupees, the value of work in Indian Rupees shall be determined by using the exchange rate declared by Reserve Bank of India as on the last date of submission of technical/ commercial bids and the eligibility of the bidder shall be determined accordingly.

If more than one bidder happens to quote the same lowest price, HLL Lifecare Ltd reserves the right to split the order and award the contract to more than one bidder.

20. PURCHASER'S RIGHT TO AMEND SCOPE OF WORK

If, for any unforeseen reasons, HLL Lifecare Ltd is required to change the Scope of Supply, this change shall be acceptable to the bidder without change in the unit price quoted.

HLL Lifecare Ltd reserves the right to reject one/ all the bids or cancel the Tender without assigning any reasons there for.

HLL Lifecare Ltd reserves the right to accord relaxation uniformly to all the bidders in case the bid submitted by all the bidders are found to have minor deviation.

21. CORRUPT OR FRAUDULENT PRACTICES

It is expected that the bidders who wish to bid for this project have highest standards of ethics.

HLL Lifecare Ltd will reject bid if it determines that the bidder recommended for award has engaged in corrupt or fraudulent practices while competing for this contract;

HLL Lifecare Ltd may declare a vendor ineligible, either indefinitely or for a stated duration, to be awarded a contract if it at any time determines that the vendor has engaged in corrupt and fraudulent practices during the execution of contract.

22. INTERPRETATION OF THE CLAUSES IN THE TENDER DOCUMENT / CONTRACT DOCUMENT

In case of any ambiguity/ dispute in the interpretation of any of the clauses in this Tender Document; Director (Marketing), HLL Lifecare Ltd interpretation of the clauses shall be final and binding on all parties.

END OF SECTION II

SECTION III

SPECIAL CONDITIONS OF CONTRACT (SCC)

1. Prices:

The price quoted shall be considered firm and no price escalation will be permitted.

The prices quoted should be inclusive of all applicable costs like equipment cost, profit margin, freight, insurance, packing, port charges, documentation charges, demurrage charges, training cost, installation, testing & commissioning charges including manpower charges, warranty charges, loading & unloading charges, applicable statutory charges (if any) at the destination country etc till final destination as per the supply order. Applicable statutory government taxes in India (if any) like GST need to be indicated separately in the prescribed price bid format. However MEA would reimburse taxes as per the applicable rates as on the date of invoice. The packing shall be transport worthy so as to prevent their damage or deterioration to goods during transit to their final destination as indicated in this document. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit and open storage. Packing case size and weights shall take into consideration, where appropriate, and the remoteness of the Goods final destination and the absence of heavy handling facilities at all point in transit. However risk in good shall continue with supplier till goods are delivered in good condition and installed at end user's site duly certified by HLL Lifecare Ltd, end- user & Indian Mission in Cote D'Ivoire (Ivory Cost).

2. Taxes and Duties:

Applicable statutory government taxes in India (if any) like GST etc would be reimbursed by MEA as per the applicable rates as on the date of invoice. However bidder should indicate the necessary details in the prescribed price format. The items being imported in Cote D'Ivoire (Ivory Cost) as per schedule A from India/ third country will be exempted from payment of Custom Duty at Cote D'Ivoire (Ivory Cost). The Government of Cote D'Ivoire (Ivory Cost) end user on the request of supplier shall provide necessary Custom Duty Exemption Certificate but the custom clearance will be the responsibility of the successful bidder. No Concession Tax Form (C/D) will be given by HLL Lifecare Ltd.

Bidder shall arrange to clear the consignment after following customs formalities at Cote D'Ivoire (Ivory Cost) and shall arrange to deliver the consignment to the end user individual site(s). The cost and risk of the consignment rests with the bidder till it is delivered to the end user individual site(s) and till the completion of Installation, testing, Commissioning and acceptance as per the relevant claused mentioned in the e-tender.

3. Software Licenses:

Standard software with basic package should be available with equipment wherever applicable and it should be updated time to time without any additional cost.

4. Chartered Engineer Certificate:

The successful bidder will be required to furnish the certificate from a Registered Chartered Engineer certifying that the items supplied and their specifications are in compliance with the requirements of the supply order issued by HLL Lifecare Ltd.

5. Completeness Responsibility:

Notwithstanding the scope of work, engineering, supply and services stated in bid document, any equipment or material, engineering or technical services which might not be even specifically mentioned under the scope of supply of the vendor and which are not expressly excluded there from but which are necessary for the performance of equipments in accordance with the

specifications and executing the contract to establish achievement of performance guarantee parameters, are to be provided for and rendered by the bidder without any extra charge so that the said project is completed in all respect. Even bidder has done consortium; in case also bidder is responsible.

6. Warranty and Support (also Service Level Agreement):

All the items covered in the schedule of requirements, shall carry minimum 12 months (twelve) on site comprehensive warranty from the date of its installation & commissioning or minimum 15 (fifteen) months from the date of delivery at site whichever is earlier. The bidder must undertake to provide the on-site support during the warranty in Cote D'Ivoire (Ivory Cost). The repairing/ rectification/ replacement/ configuration required, if any, must be done at site only. During the warranty, all complaints should be rectified within 14 (fourteen) working days from the date of intimation. In case the rectification of fault involves replacement of some hardware/software the same should be carried out within 21 (Twenty One) working days from the date of intimation.

Failure to do so would result in the levy of penalties. The PBG will be released by HLL Lifecare Ltd only after the submission of satisfactory performance certificate issued by end- user & verified by Indian Mission in Cote D'Ivoire (Ivory Cost) after the completion of warranty period.

7. Payments:

Payment will be done for each Schedule separately as per milestones below:

- a. HLL Lifecare Ltd shall release 70% of the total cost to the supplier on dispatch of all items as per supply order from India/ third country and after receipt of funds from MEA, on due verification of documents pertaining to proof of dispatch and receipt of Performance Bank Guarantee.
- b. HLL Lifecare Ltd shall release 20% total cost to the supplier after delivery of all item(s) as per supply order at all site(s) and after receipt of funds from MEA, on verification by HLL Lifecare Ltd , end user and Indian Mission in Cote D'Ivoire (Ivory Cost).
- c. Balance 10% of the total cost to the supplier shall be released by HLL Lifecare Ltd after successful installation & commissioning of all item(s) at all site(s) and after receipt of funds from MEA on physical verification at site and certification by a Project Monitoring Committee (PMC) duly constituted by MEA and represented by IFD, MEA.

8. Performance Bank Guarantee (PBG):

The successful bidder within 30 days must submit a PBG equivalent to 10% of the order value on receipt of supply order from HLL Lifecare Ltd as per the format provided in Annexure VI of this Tender document. This Bank Guarantee should remain valid 24 (Twenty Four) months from the date of its submission to HLL Lifecare Ltd.

9. Transport/ Shipping Documents:

After the consignment is ready for dispatch, the successful bidder shall be required to furnish the following documents:

- Chartered Engineer's Certificate
- Bill of Lading/ Air Way Bill (booking details)
- Packing List
- Insurance Policy
- Invoice & other relevant document(s)

Actual shipment should be done only after receipt of concurrence from HLL Lifecare Ltd based upon the above-mentioned documents.

10. Penalty for delayed Services:

HLL Lifecare Ltd reserves the right to levy penalty as per the relevant clauses of this Tender. The penalties, if any shall be recovered from the Performance Bank Guarantee (PBG) submitted by the successful bidder or from the Balance payment reserved with HLL Lifecare Ltd. HLL Lifecare Ltd may consider relaxing the penalty and delivery requirements, as specified in this document, if and to the extent that, the delay in performance or other failure to perform its obligations under the contract is the result of circumstances not attributed to the bidder. For any such relaxation, the bidder should sought prior written approval from HLL Lifecare Ltd by submitting proper justification with documentary evidences.

11. Jurisdiction:

The disputes, legal matters, court matters, if any shall be subject to Thiruvananthapuram jurisdiction only.

12. Force Majeure:

HLL Lifecare Ltd may consider relaxing the penalty and delivery requirements, as specified in this document, if and to the extent that, the delay in performance or other failure to perform its obligations under the contract is the result of a Force Majeure. Force Majeure is defined as an event of effect that cannot reasonably be anticipated such as acts of God (like earthquakes, floods, storms etc.), acts of states, the direct and indirect consequences of wars (declared or undeclared), hostilities, national emergencies, civil commotion and strikes at successful bidder's OEM premises or at the destination site.

13. Arbitration:

All disputes of any kind arising out of supply, commissioning, acceptance, warranty maintenance etc., shall be referred by either party (HLL Lifecare Ltd or the bidder) after issuance of 30 (thirty) days' notice in writing to the other party clearly mentioning the nature of dispute to a single arbitrator acceptable to both the parties. The venue for arbitration shall be specified in the purchase agreement. The jurisdiction of the courts shall be specified in the purchase agreement. The Arbitration and Conciliation Act 1996 and the rules made there under with all/ any modifications or amendments thereof for the time being in force shall apply to the arbitration proceedings.

END OF SECTION III

SECTION IV

TECHNICAL SPECIFICATION

Schedule B - Cote D'Ivoire (Ivory Cost)

1. Mucosity vacuum cleaner / Suction cup:

Nasal aspirator with high quality medical grade rubber

2. Autoclave Cocotte/ Table top Autoclave 20-25 litre capacity

- Sterilizer Type: Table Top Sterilizer.
- Capacity: 20-25 litres
- Chamber Size: The sterilizer should have Rectangular/Cylindrical chamber with suiting the volume for maximum processing capacity per charge.
- Quality System Compliance: Sterilizer should comply the quality systems as per ISO 9001:2000, EN ISO 13485:2003, and ISO 14001:2004. Sterilizer should be CE or FDA or BIS Certified
- Types of Cycles Process: Table Top Sterilizers should be equipped with temperature cycles.
- Chamber: Should be made of S.S.316 & should comply the Pressure Equipment Directive (PED) & EN 13445 norms or equivalent, Chamber should confirm 44-50,000 process minimum life. Chamber should have working pressure 2.2 bar & design pressure up to 3.8 bar. Chamber should have Stress & Fatigue analysis reports for material & construction of the pressure vessel. Chamber should be equipped with electrically heated jacket for preheating on standby mode. Door should have horizontal sliding/Hinged door with silicon elastomer rubber gasket to withstand temperature up to 140°C. 5 spare gaskets need to be provided along with sterilizer.
- Air Filter: A disposable air filter should be provided for filtering the atmospheric air before entering inside the chamber and the filter separation efficiency should be higher than 99.998% for particle size less than 0.3µm. 50 nos of filters need to be provided along with the sterilizer.
- Water Storage Tank: Sterilizer should have inbuilt water reservoir with storage capacity up to 5 Litres. The water reservoir should have easy access for cleaning & to avoid bio film.
- Steam Generator: Sterilizer should have inbuilt steam generator with integrated energy storing system for building up power for sterilization loads in short time. 5 spare heating elements need to be supplied with the sterilizer.
- Control Panel: The control system should have cycle START AND STOP switches. Emergency breaker switch need to be provided. Should have analog temperature and pressure measuring devices.
- Accessories: The sterilizer unit should include rack with 3 levels & suitable size instrument trays should be the part of the supply. The Sterilizer should have water circulation system.
- The complete instrument and accessories must be supplied with one set of spares apart from the spares mentioned above.

3. Basin shape plastic slipper (plastic basin), Plastic bowl, Urinal plastic women:

High grade plastic

4. Suture box (With Instruments & SS box):

All Necessary instruments and box are made up of medical grade 304 steel

5. Complete Delivery box (With Instruments & SS Box):

All Necessary instruments and box are made up of medical grade 304 steel

6. Episiotomy box (With instruments & SS Box)

All Necessary instruments and box are made up of medical grade 304 steel

7. Care trolley (baby carrying bassinet):

- a. Single moulded polycarbonate tub
- b. Easily Removable and washable foam bed and infant tub
- c. Over all Dimension: Length~765mm, Width~400mm, Height~1055mm
- d. Infant bed (Tub) dimension: Length~ 765mm, Width~400, Height~250mm

- e. Weight: Around 20 Kg
- f. Maneuverability : 4" castors with brake
- g. Mattress: High Quality PU foam
- h. Coating: Epoxy/Powder coated for scratch resistance and rust protection
- i. Environmental Specification:
 - 1. Operating Temperature range: 10 C to 60 C
 - 2. Transport or Storage Temperature : -10C to 60 C
 - 3. Humidity range: 5% to 100% RH
 - 4. Atmospheric Pressure: 500 to 1060 HPA

8. Step ladder (2 steps) / Foot Step:

- Polished SS tubular legs fitted with high quality nylon gromets.
- SS made top having high grade 3mm nylon matting and aluminum beading
- Overall size: - 460(L)*500(W)*460(H) in mm (for Double Foot Step Stool)
- All the Stainless Steel should be seamless conforming to 304 grade/ 16 gauge and polished finished.

9. Insufflator (resuscitator)manual baby:

- Should be Medical grade material.

10. Inpatient bed + accessories / Bed observation + mattress:

- Inpatient bed should be of following minimum dimension: 2080 mm L x 920 mm W x 560 mm H (without mattress)
- Should have at least four sections. Top flat platform made of detachable CRCA perforated sheet.
- Bed frame must be sturdy and stable to support weight of at least 200 kg. The frame structure should be made up of at least 16 G CRC, rectangular / circular pipe of 60 mm x 30 mm.
- Bed frame mounted on round / rectangular tube of minimum 30 mm diameter and 16 G thicknesses, fitted with 125 mm castors, 2 with breaks and having rustproof body.
- All adjustments for fowler position must be obtained from crank shaft, manually operated with stainless steel foldable handle on both the shaft.
- The finished bed must be rust proof, pre-treated and treated with epoxy polyester antimicrobial powder coated to increase the bacteriostatic property.
- The bed should have a pair of swing down type full length side rails, MS powder coated with self-locking.
- Should have easily removable head and foot panels made up of SS with four corner buffers.
- Mattress area of approximate Length 2000 to 2010 mm X Width 900 to 910 mm, Height: - 500 to 550mm(without mattress).
- Bed should have the side detachable protection rail.
- There should be suitable buffer mechanism to avoid hitting of the bed to the wall.
- Should have provision of fixing suitable rod for hanging intravenous / irrigation fluid bottle on both sides at head end and foot end. Each bed should be supplied with 2 nos. good quality telescopic I.V. rods.
- Should have hooks on bed frame on both side for holding urine / drainage bag (at least 4 Nos. in each bed).
- Each bed must be quoted with one no., four section mattress of dimension (2000 mm X 900 mm) with washable cover of lasting quality. The mattress should be made of high density PU foam of 100 mm thickness.
- Back rest and leg rest both shall have three mattress guards.
- All SS parts should be matt finish and made of 304 grades / 16 gauge.

11. Stainless steel kidney tray / Stainless steel trays with lids / sterilization drums:

- Made up of 304 SS

12. Baby scale with direct display:

- Weight range 0-20 kg (minimum weight to be weigh 20 gm).
- Accuracy +/- 5gms, resolution 5 gms.

- Unit should have facility to accurately weighs the hectic / active baby and retain the digital display for 30 sec. even baby is removed from the scale.
- Zeroing facility.
- Durable HIP molded baby tray, it should be detachable, to weigh standing babies.
- Baby Tray construction should not allow baby to be injured or slip from the scale.
- Large bright display for strain free reading
- Display in kg and lbs., easy switch between kg and lbs.
- Reading time max 5 seconds

13. Weigh Machine(150 kg)/ Bathroom scales:

- Should have an accuracy of 500 gms.
- Should be dial type having a magnifying lens to see the measurement.
- Should measure a maximum weight of 150kgs.
- Should have zero adjustment.
- Should bear certification for each machine by department of legal metrology
- Should be Round shape of diameter 300mm (minor variations will be accepted)
- Shall be made of Metal, epoxy powder coated with rust proof parts.

14. Serum hanger (IV Stand) / Stem on wheel:

- Overall approx. size: height 150cm-230 cm (with telescopic adjustable height)
- Main Frame: Strong & Sturdy stainless steel tubular construction mounted on four pronged tubular/rectangular base fitted with four swivel rust proof castors 50mm diameter .
- Stainless steel rod with double hooks
- All the Stainless Steel should be 304 grade
- Should be pre-treated and epoxy coated finish

15. Waste Bin Pedal Operated(20L)-SS

- Should be made up of high quality stainless steel.
- Should have minimum capacity of 20 liters.
- The covering lid should be open able by pressing the plate attached to the bottom.

16. Double head stethoscope/Standard bi- auricular stethoscope:

- Should have double tube from diaphragm
- Should be four eartips with single diaphragms
- Should be able listen the lung and heart sounds Chest piece - Ø45mm(adult) , - Ø35mm (paediatric)
- Should be light weight
- Length should be 27" to 29"
- Should have high quality sound transmission.
- Preferable colour -black Should have tight and soft sealing ear tips
- Configuration Accessories, spares and consumables: • Rims• Binaurals • Eartips • Diaphragms •
- Documentation: System Manufacturer should have ISO certification for quality standards

17. Fetal Doppler Echo:

- Should be high sensitivity hand-held fetal doppler to detect fetal heart rate.
- Should be light weight
- Should have LCD display to show the detected fetal heart rate (FHR).
- Heart rate rage should be 60-210 born
- Should be battery operated.
- Battery should be rechargeable type.
- Should have auto power down facility to conserve battery life.
- Battery life should be > 6 Hrs.
- Should have low battery indicator.
- Should have following probes.
 - 2MHz probe suitable for general obstetric use
 - 3MHz probe suitable for detecting heartbeat in very early gestation

- System Configuration Accessories, spares and consumables
 - System as specified - 01 no
 - 2 Hz probe - 01 no
 - 3MHz probe - 01 no
 - Extra Battery 01 no
 - Carrying case bag - 01 no
- Standards, Safety and Training
 - Manufacturer should have ISO certification and copy of the same should be submitted along with the technical bid.
 - The quoted model should be FDA/CE approved and copy of the same should be submitted along with the technical bid.
 - Documentation
 - User/Technical manual in English should be supplied.
 - Warranty and CMC as per tender terms.
 - Must submit atleast 3 user list and performance report within last 5 years from major hospitals.
 - Copy of the same should be submitted along with the technical bid.

18. Delivery Bed (Childbirth bed/Table):

- Delivery bed should be supplied with all accessories
- Overall size 2100 – 2400 mm L when extended , 900mm– 1100 mm W (with and without side rails up), 440 mm to 900 mm H (without mattress).
- Bed should have three section high pressure laminated top. It should have detachable laminated top for middle section and leg section for ease of cleaning. Leg section should be telescoped under backrest for lithotomy position.
- It should have control devise for making height and back adjustments.[Manual as well as remote control].
- Frame should be made of mild steel which should be pre treated and powder coated.
- Bed shall have swing down type side railings; railings shall be made from non-rusting polymer moulded material.
- It should have three sectional mattress and seat section should have large perineal cut.
- Mattress should be made of high resilience high quality P.U foam with stain resistant cover and zip.
- Should have a pair of bearing down patient hand grips with PVC covers and with location adjustability for patient convenience.
- All positions should be operated by electromechanical adjustment through a hand set; battery backup with inbuilt battery charger should be provided. The hand set should have indications for power on and the battery charge.
- It should have polymer moulded easily detachable head and foot side panels. It should have four corner rubber buffers.
- Bed should be mounted on 12 cm – 15 cm non rusting castor wheels with locking mechanism. Castor should be made from high grade non floor staining synthetic material. Wheel center should have precision ball bearing to run smoothly.
- Should have infusion rods which have adjustable heights, quick release and attaches to all corners of bed.
- Should have a pair of upholstery aluminium crutches mounted on SS tubes.
- Should have sliding waste collection bowl at perineal part of table.
- It should be able to give trendelenburg, reverse trendelburg and 60 - 70 degree sitting position both mechanically and electronically.
- It should have adjustable foot supports for nursing staff.
- All consumables required for installation and standardization of system to be given free of cost. Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. Or should comply with 89/366/EEC; EMC-directive.
- The unit shall be capable of operating continuously in ambient temperature of 10-40 deg C and relative humidity of 15-90%.
- The unit shall be capable of being stored continuously in ambient temperature of 0-50deg C and relative humidity of 15-90%.
- Power input to be 220-240VAC, 50Hz fitted with Indian plug.

- UPS of suitable rating with voltage regulation and spike protection for 60 minutes back up.
- Should be FDA, CE, UL or BIS approved product
- Manufacturer should have ISO certification for quality standards.
- Comprehensive training for lab staff and support services till familiarity with the system.
- Comprehensive warranty for 2 years and 5 years CMC after warranty including UPS.
- Electrical safety conforms to standards for Electrical Safety IEC 60601-2-38 Particular safety requirements for Electrically operated hospital beds.
- User/Technical/Maintenance manuals to be supplied in English.
- List of Equipments available for providing calibration and routine Preventive Maintenance Support as per manufacturer documentation in service/technical manual.
- Certificate of calibration and inspection.
- Compliance Report to be submitted in a tabulated and point wise manner clearly mentioning the page/para number of original catalogue/data sheet. Any point, if not substantiated with authenticated catalogue/manual, will not be considered.
- List of important spare parts and accessories with their part number and costing.
- Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.

19. BP Machine (Adult)

- Display : LCD Digital Display
- Measurement Method : Oscillometric
- Measurement Range : Pressure: 0 to 299 mmHg
- Pulse: 40 to 180 beats/ min.
- Accuracy : Pressure: ± 3 mmHg
- Pulse: $\pm 5\%$ of display reading
- Inflation : Fuzzy-logic controlled by electric pump
- Deflation : Automatic pressure release valve
- Memory : At least Last Measurement
- Power Source: 4 "AA" batteries 1.5V or AC adapter (optional, INPUT AC100- 240V50/60Hz 0.12A)
- Cuff/ Tube Material: Nylon, polyester, polyvinyl chloride
- Different Cuff size: Adult, Obese

20. BP machine (Pediatric)

- Display : LCD Digital Display
- Measurement Method : Oscillometric
- Measurement Range : Pressure: 0 to 299 mmHg
- Pulse: 40 to 180 beats/ min.
- Accuracy : Pressure: ± 3 mmHg
- Pulse: $\pm 5\%$ of display reading
- Inflation : Fuzzy-logic controlled by electric pump
- Deflation : Automatic pressure release valve
- Memory : At least Last Measurement
- Power Source: 4 "AA" batteries 1.5V or AC adapter (optional, INPUT AC100- 240V50/60Hz 0.12A)
- Cuff/ Tube Material: Nylon, polyester, polyvinyl chloride
- Different Cuff size: Pediatric

21. Infrared Thermometer

- Has dual mode (object and human) so measures temperature of milk, water, etc for child
- Operating environment is - 0 degree Celsius to 60 degree Celsius
- Has audio beeper for indicating measurement
- Degree Celsius to Fahrenheit convertible
- Gives measurement in just 2 seconds
- For use of forehead
- CE or ISO or FDA approved
- Has an memory button which stores last 10 readings
- Sensor Type - Infrared Thermometer

- Batteries Required - Yes
- Battery – 1 or 2 AAA Alkaline batteries
- Material - BPA Free Plastic

22. Tensioning device (Salters scale and complete tripod stand Hanging):

- Portable, spring, Salter, hanging type,
- capacity 25kg, markings 500g with 100g graduations
- Supplied with trousers and slings Spare: slings and trousers,
- Tripod Stand for hanging Spring Salter of SS

23. Neonatal vacuum cleaner:

- Nasal aspirator with high quality medical grade rubber

24. Basin:

- Stainless steel with 4 wheel stand with breaks

25. BedPan:

- Medical grade 304 stainless steel

26. Box of dressing:

- With all necessary instruments and its box of 304 SS

27. Stainless steel bin/ Average stainless steel bin INOX:

- Should be constructed from stainless steel with lid
- Should have removable inner container of S.S. material.
- Capacity 10 litre or more
- Size: 28 x 36cm (approximate)

28. Heating Lamp

- Radiant Heat & Light
- special single bulb that furnishes both radiant heat and clear white light
- Adjustable height & light hood
- Easy to set at the correct distance and angle
- Control panel set at just the right height for use
- Should have 15 minute on/off safety timer and an adjustable intensity control
- Warm-Lamps have a roll-around base with five swivelling casters with break for ease of movement, and are counterbalanced for stability & safety.
- Should have adjustable height – from 58 to 74 in. (147.3 to 187.9 cm) – so the lamp can be placed at just the right distance from the baby
- Incandescent light. 250W
- Should be provided 2 spare bulbs with each lamp and should be user replaceable
- Should be ergonomically made for the use of Neonatal/paediatric care.

29. Gynecological / Examination Light Stands:

- Should be movable on wheel 5 wheels
- Should be able to adjust position.
- Should use a LED light source.
- It should have minimum light intensity of 39,000 lux
- It should have 4300 chromaticity
- CRI should be minimum 90
- It should have minimum 6 LED
- It should have on/OFF switch on light head handle
- Color temperature should be 4000K
- Stand height should be minimum 38" from the floor and arm length 36"
- Arm should can adjust Horizontally and vertically

- It should can fit provision for either ICU cot / labor cot / table
- Powers supply should be SMPS
- It should be CE/FDA certified

30. Delivery Kit:

- Gown(Non woven fabric material 5'X5' : 2
- Shoe Cover (Plastic): 2 pair
- Cap and Mask: 2 pair
- Gloves(Latex): 2 pair
- Perineal sheet Non woven fabric material 4'X6'): 1
- Couch Sheet(Non woven fabric material 4'X6'): 1
- MOPS(Egyptian Cotton 15"X15"): 3
- Table drape(Plastic): 2
- Umbilical cord clamp(plastic): 1

31. Pediatric Examination Table (Pediatric Bed + Mattress)

- Built-In Scale (Height & Weight- digital)
- Ranges in one ounce increments (measurements given in both imperial and metric)
- Paper Roll Holder
- Storage: Large drawer with Door/drawer locks
- Patient Weight Capacity: 4.5 lb (2 kg) - 40 lb (18 kg)
- Adjustable stirrups
- Electrical: AC adaptor 100 to 250 VAC, 50-60 Hz
- Dimensions: Minimum 35.8" (90.8 cm) H x 42" (106.7 cm) L x 20.75" (52.7 cm) W

32. Mask and new-born balloon:

- Infant mask with ambu bag

33. Mattress (adult):

- High Quality PU foam with vinyl cover
- Mattress area of approximate Length 2000 to 2010 mm X Width 900 to 910 mm, 100 mm thickness

34. New-born mattress:

- High Quality PU foam with vinyl cover with 100 mm thickness

35. Endo-uterine vacuum cleaner:

- Vacuum aspiration kit

36. Neonatal Resuscitation Unit

- Should have microprocessor based heater control and manual modes of operation.
- Should have both skin and air modes of operation.
- Should have user friendly touch sensitive control panel with large easy to read LED displays for air and skin temperatures.
- Should have Quartz Infrared Heater with parabolic reflector for uniform heat radiation
- The unit should contain an in-built double surface phototherapy unit with acceptable distance for effective treatment.
- The heater unit should be protected by a suitable grill
- The heater unit should be swiveling type and should be swivelled effortlessly.
- The probes should be detachable type and should be interchangeable.
- Should have memory back up to retrieve set data against power failure
- Should have calibration free temperature sensors.
- The heater should automatically cut off at 38 degree Celsius irrespective of the set parameters.
- Should be mounted on four smooth running swivelling casters with integrated brakes.
- Should have a monitor stand and IV drip pole.
- Should have alarms with visual indicator for the following
 - a. Temp high
 - b. Temp low

- c. Probe failure
- d. Power failure
- e. Heater failure
- Should have an examination light with ON/OFF switch.
- Should be provided with integrated baby bed system with cassette tray compatible for taking X-ray.
- Should be provided with withdraw able bed with head raising facility on both end.
- Should be supported with easily removable side flaps.
- Should have an in built suction unit with pressure control.
- Should have an in built humidified oxygen outlet with flow meter control.
- Should be supplied with suitable filled oxygen cylinder with cylinder holding facility at the rear side.
- Should have manual resuscitation unit with PEEP and airway pressure control facility.
- The unit should be made of mild steel tubular structure pretreated and powdercoated.
- Should operate on mains supply 200 to 240V ac, 50 Hz
- Should have safety certificate from a competent authority CE / FDA (US) /STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL./test report from ETDC. Copy of the certificate /test report shall be produced along with the technical bid.

37. Monitors (Patient Monitors):

- For adult, pediatric and infant use
- 8.4 inch to 12.5 inch size, LED display with minimum 5 and up to 8 waveforms
- User configurable touch screen
- Quick keys to rapidly access frequently used functions such as trend review, alarm setting and useful display modes including large fonts and mini trends
- No-fan design for a quiet care environment
- Parameters including ECG, RESP, SpO₂, NIBP, PR, TEMP
- Upgradable to IBP and EtCo₂
- ECG:
 - 3 lead and 5 lead ECG
 - ST analysis
 - Arrhythmia detection
 - Defibrillator protection: withstand 360 J of defibrillation
 - Pacemaker detection
- Heart rate:
 - 15-300 bpm
 - Accuracy ± 1 bpm
- Respiration
 - 0-120 rpm
 - Accuracy 7-120 rpm ± 2 rpm
 - Lead I or II
- SpO₂ (technology: Nelcore or Massimo)
 - 0-100%
 - Accuracy $\pm 2\%$ (70-100%, Non motion)
 - Accuracy $\pm 3\%$ (70-100%, motion)
- Pulse rate
 - 20-250 bpm
- NIBP
 - Method - Automatic oscillometric
 - Manual, Auto and stat operation modes
 - Accuracy: maximum mean error ± 5 mm Hg
- Temperature
 - Range: 0-50 degree C
 - Accuracy: ± 0.1 degree C

- Side stream CO2
 - Range: 0-99 mm Hg
 - Accuracy 0-40 mm Hg \pm 2 mm Hg, 41- 76 mm Hg \pm 5% of reading, 77- 99 mm Hg \pm 10% of reading
 - Sample flowrate: 70, 100 ml/min
- Required software for integration with central monitoring system
- Portable, weight less than 4 kgs
- Up to 7 hours battery backup for continuous monitoring with large capacity Li-ion battery
- 360-degree visible alarm indicator
- 120 hours trend and 48 hours waveforms reviewing
- Standard and optional parameter configuration
- With LAN and WiFi capability: Integration with central monitoring system
- Quick access ergonomic buttons
- provision for IBP dual channel
- Must be supplied with the following accessories for each monitor
- Reusable NIBP Cuff Adult: 5nos
- Reusable NIBP Cuff Pediatric: 5nos
- Reusable NIBP Cuff Neonatal: 5nos
- Reusable SpO2 Probe Adult: 5nos
- Reusable SpO2 Probe Pediatric: 5nos
- Reusable 5 Lead ECG cable set: 5nos
- Reusable Temperature probe: 5nos

38. Oxygen (SpO2 monitor)/Pulse-oxymeter :

- Compact portable bedside pulse oxymeter with LCD display.
- Continuous monitoring of SpO2 (arterial blood oxygen saturation), pulse rate and signal strength
- Measuring range:
 1. SpO2: 30 to 100 %, minimal graduation 1%.
 2. Pulse rate: 20 to 250 bpm, minimal graduation 1 bpm.
 3. Accuracy SpO2: 50 to 69% (\pm 3%), 70 to 100% (\pm 2%).
 4. Display shows SpO2(%), HR(bpm) and signal strength bar
- Large display readable from distance, display cover durable plastic
- User preset of high/low alarms on SpO2 and pulse rate monitoring
- Audio visual alarm for SpO2 and pulse rate in case measurements are outside preset range
- Silencing feature for audio alarm
- Display reports system errors, probe failure and built-in battery status
- Automatic switch from mains to batteries in case of power failure
- Power requirements: 220 V/50 Hz and internal re-chargeable battery (autonomy approx 6 hrs, automatic recharge).
- Power consumption: 50 W.
- Device is produced by ISO 9001 certified manufacturer.
- **Supplied with:**
 1. 5 x reusable SpO2 sensors pediatric, clip-on type (including connection cable).
 2. 5 x reusable SpO2 sensors adult, clip-on type (including connection cable)
 3. 5 x reusable SpO2 sensors neonate, wrap around type (including connection cable).
 4. 2 x spare set of fuses.
 5. 5 extension cables
 6. User manual with trouble shooting guidance, in English.

- Technical manual with maintenance and first line technical intervention instructions, in English

39. Bed Side Screen/ Three part Screen:

- SS screen stand
- Overall size: 1680 mm L x 2450 mm W approximately
- Three folding partitions
- Tubular framework made from 19 mm dia x 1.2 mm ERW tube
- Base tubular made from 25.4 mm dia x 1.2 mm tube, fitted with 50 mm dia non rusting castor

- Middle frame 1220 mm and side frame 610 mm W and each frame attached with four hooks
- All MS parts are pre-treated and powder coated
- All SS parts to be 304 grade/ 16 gauze

40. Nursing tray with lid:

- All Necessary instruments and tray with lid are made up of medical grade 304 steel.

41. Medium care tray with lid:

- All Necessary instruments and tray with lid are made up of medical grade 304 steel.

42. Stainless Pedal Wastebin:

- Should be constructed from stainless steel.
- Should have removable inner container of S.S. material.
- Capacity 10 litre or more
- Should have foot operated lid opening.
- Size: 28 x 36cm (approximate)

43. Poupinel Sterilizer / Dry Heat Sterilizer

- Should be suitable for glassware and metal as well as liquids with low moisture content.
- Microprocessor based system with temperature controller with integrated auto diagnostic system with fault indicator.
- Thermostatically controlled system.
- External : Stainless steel casing : 380(W) x 365 (D)x 390mm (H)(all dimensions are approximate).
- Insulated stainless steel door with locking
- Capacity of dry heat sterilizer: 50 litres
- Interior: 380mm x 365mm x 390 mm, 50 litres approx easy to clean interior, made of SS 304 , with supports on the three sides for two adjustable perforated stainless steel shelves.
- Forced air circulation by quiet air fan to ensure uniform temperature.
- Fitted with load indicator and safety thermostat take over indicator lamp.
- LCD/LED display
- Temperature variation $\pm 1^\circ$ C.
- Temperature range - ambient to 300 degree Celsius.

44. Sterile speculum:

- ETO sterilized high grade Plastic speculum

45. Obstetrical stethoscope:

- Should have double tube from diaphragm
- Should be four eartips with single diaphragms
- Should be able listen the fetal heart sounds with metallic pinard horn of around 8 inch long
- Should be light weight
- Length should be 27" to 29"
- Should have high quality sound transmission.

- Preferable colour -black Should have tight and soft sealing ear tips
- Documentation: System Manufacturer should have ISO certification for quality standards

46. Gynaecological examination table:

- Epoxy powder coated steel frame
- Two part padded, shaped mattress base with removable extension for leg section
- Easy-to-clean leather upholstery
- Gas spring assisted adjustment of the back section and hydraulic height adjustment
- Castors with central brake
- Dimensions
 - Minimum length 1500 mm without extension and 1900mm with extension
 - Minimum width 700mm
- Should be provided with following accessories
 - Paper roll holder
 - Bowl and holder
 - Accessory rails for the back section, pair
 - Accessory rails, pair
 - Knee supports, pair
 - Heel supports, pair
 - Clamp for fixing knee supports
- Standard
- Should be European CE and ISO approved.

47. Heating table (Warmer):

- Should have microprocessor based heater control and manual modes of operation.
- Should have user friendly touch sensitive control panel with large easy to read LED displays for actual and set temperatures.
- Should have Quartz Infrared Heater/ Calrod Heater with parabolic reflector / J shaped reflector for uniform heat radiation
- The heater unit should be protected by a suitable grill.
- The heater unit should be swiveling type/ recessed heater type and should be able to position effortlessly for performing various procedures including X rays etc .
- The probes should be detachable type and should be supplied as 2nos for each machines. The probes should be covered under warranty and CAMC.
- Should have memory back up to retrieve set data against power failure
- Should have calibration free temperature sensors.
- The heater should automatically cut off at 38 degree Celsius irrespective of the set parameters.
- Should be mounted on four smooth running swiveling casters with integrated brakes.
- Should have a monitor stand and IV drip pole.
- Should have alarms with visual indicators for the following
 - i. Temp high
 - ii. Temp low
 - iii. Probe failure
 - iv. Power failure
 - v. Heater failure
- Should have an examination light with ON/OFF switch.
- Should be provided with integrated baby bed system with cassette tray compatible for taking X-ray.
- Should be provided with withdraw able bed with head raising facility on both end.
- Should be supported with easily removable side flaps.
- The unit should be made of mild steel tubular structure pretreated and powder coated.
- Should work with input 200 to 240Vac 50 Hz supply.
- Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL./test report from ETDC. Copy of the certificate / test report shall be produced along with the technical bid.

48. Neonatal Open Care System

HLL/SD/RBD/2018-19/TENDER/10 DT. 24.10.2018

- Large clear display and control panel
- Easy access to baby
- Side panels
- Highly sensitive thermistor probe with no field calibration required
- Splash-proof heater unit – manual and servo modes
- Provision for X-ray cassette
- Bed tilt able for Fowler positioning
- Warmer unit swivels to accommodate the X-ray units
- Clear visual and acoustic alarms
- Smooth running swivel castors with brakes
- IV Stand
- Examination lamp

49. Sphygmomanometer:

- Should be Portable mercurial type.
- Should have ISI mark.
- Should have ON and OFF provision for mercury reservoir.
- Should have a measuring range from 0 to 300 mmHg.
- Should be provided with adult arm cuff of size 225 x 125 mm and pediatric cuffs of size 180 x 100 mm and 120 x 60 mm.
- The control valve should have a knurled thumb control device. The leak rate should not exceed 10 mm of mercury per minute.
- The manometer scale markings and graduations should be engraved or etched and filled with pigments and should be permanent and it should meet the requirements of boil test.
- All metal parts should be powder coated and non corrosive.
- Glass parts should be made of clear glass and should be subjected to boil test.
- The internal diameter of the manometer glass tube should be 4.1 ± 0.1 mm and the thickness not less than 2 mm.
- All plastic parts used should not crack, flake, peel or disintegrate in normal use.
- The inflating rubber bag should be capable of withstanding an internal pressure of 450 mmHg without leaking.
- The inflating bulb should be soft and should not have any joints or ridges.
- The mercury used should be clean, double distilled and of 99.9% purity.
- The fastening arrangements of the cuff should be of hook and loop type (Velcro). It should have a minimum width of 4 inches and length of 6 inches.
- The threading and fastening arrangement of the cuff should show no sign of slip or failure when subjected to the maximum test conditions.
- The rubber tubes used should have an internal diameter of 3 ± 0.5 mm and the external diameter should not be less than 8mm.
- The tubes should be fitted with male and female leur connectors.
- The housing case should be of robust design. It should have press to release lock. It should have metal hinges. The tube should be secured with metal screws and clamps. It should have mechanism to hold the lid in right angles and should prevent accidental dropping. All parts should be replaceable incase of breakage.
- A cleaning brush to clean the manometer tube and a set of spare washers may be provided with each unit.

50. Spengler blood pressure monitor

- Display : LCD Digital Display
- Measurement Method : Oscillometric
- Measurement Range : Pressure: 0 to 299 mmHg
- Pulse: 40 to 180 beats/ min.
- Accuracy : Pressure: ± 3 mmHg
- Pulse: $\pm 5\%$ of display reading
- Inflation : Fuzzy-logic controlled by electric pump
- Deflation : Automatic pressure release valve

- Memory : Minimum 50 Last Measurements
- Continuous measurement and auto measurement with time setting
- Power Source: 4 “AA” batteries 1.5V or AC adapter (optional, INPUT AC100- 240V50/60Hz 0.12A)
Cuff/ Tube Material: Nylon, polyester, polyvinyl chloride
- Different Cuff size: Pediatric, Adult, obese

51. Stadiometer(Height Measurement):

- Designed to measure height of adult and children aged 24 months and above in vertical positions.
- Firm flat vertical surface.
- Fold up mechanism and low weight making it compact and easy to transport.
- No part should be loose or shaking during transportation.
- Simple and fast set up, no tool (screw driver etc.)should be required.
- Stable connection, assuring precise and accurate measurement.
- No need for calibration as all parts have prefixed position.
- Double sided graduation for easy measurements.
- Orientation of numerals on the measurement scale: parallel to the board.
- Unit of measure: centimeters.
- Smallest graduation: 0.1 cm.
- Measurement range: Should be able to measure upto 200cm at least;
- Accuracy (+) or (-)0.2 cm.
- Width/depth of the board: approx 5.5 cm.
- Large footplate, providing extra stable base and smoothly gliding measuring slide/wedge.
- Adjustable feet for stability on uneven or soft ground.
- Spirit level indicator to check flate surface.
- Friction feature between board and measuring slide/wedge for preventing the latter from drooping when release.
- No sharp edge or corner.
- Designed for heavy duty use in demanding circumstances.
- Durable, resistant to effect of excessive humidity and high temperature, water resistant and shock resistant.
- Board should be made of Mild Steel (Rust proof)/Aluminium.23.Precision: (+)or (-)0.2 cm; Supplied with:
- Carry bag or carry case(i)made of long lasting fabric, (ii)zipper closer (iii)removable, adjustable shoulder/back strap(s) and (iv) small side pocket for storage of operating instructions, loose parts and/or stationery material.
- Instructions for use, training, maintenance and trouble shooting in English, Hindi and any other regional language as per requirement, appropriately illustrated with pictograms.
- Packaging: Packed in a carton box which should be of sturdy quality and provides adequate protection of the goods while transportation

52. Infantometer (Baby Height Measurement):

- Portable board for measuring length of baby.
- Designed to measure length of baby in recumbent position.
- The board shall be firm, flat horizontal surface.
- Fold-up mechanism and low weight, compact and easy to transport.
- No part should be loose or shaking during transportation.
- Simple and fast set up, no tools (screwdriver, etc.) required.
- Stable connections, assuring precise and accurate measurement.
- No need for calibration, all parts to be in fixed position.
- Long lasting, easy to read in low light conditions.
- Double sided graduation for easy measurements.
- Orientation of numerals on the measurement scale: parallel to the board.
- Facilitating quick and easy reading of results.
- Unit of measure: centimeters.
- Smallest graduation: 0.1 cm.

- Precision:(+) or (-)0.2 cm.
- Width on the board 30 cm approx
- Fixed head board and smoothly moving foot board.
- Angled edges(curved)to prevent baby from rolling sidelong
- No sharp edges or corners.
- Design for heavy duty use in demanding circumstances.
- Durable, resistant to effects of excessive humidity and high temperature, water resistant and shock resistant.
- Board should be made of 6mm (Six millimetre)ABS Plastic.
- Measurement range: Should be able to measure upto 100 cm at least;
- Accuracy; (+) or (-) 0.2 cm.

Note:

- It is core responsibility of bidder to understand the site before bidding and make sure arrangement being provided by Government of Cote D'Ivoire (Ivory Cost) as per schedule A.
- Any, remaining parts such as: Base for Machines, Sheds, Electrical Wire, Water Pipeline or any kind of additions / changes / modifications required would be responsibility of bidder only
- It is the responsibility of bidder to arrange all parts for installation which is not covered by local government to be locally done within given timeline.
- Bidder has to make arrangement to provide services / spare parts for all quoted machines in short time. It can be done by visit of engineer or setting up a local officer or by availing spare parts in advance.

END OF SECTION IV

Annexure - I

(On Rs 200/- stamp paper duly notarized)

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

Corporate & Registered Office

HLL Bhavan, Poojappura,

Thiruvananthapuram - 695012,

Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

Subject: Authorization to sign the bid document

This has a reference to your Tender no..... dated for Supply, Installation, Testing & Commissioning and Onsite Support of Medical Equipments and Materials to be Supplied to Cote D'Ivoire (Ivory Cost)

It is hereby confirmed that, Sh./working as is entitled to act on behalf Smt.....of our corporation/Bidder/ firm/organization and empowered to sign this document as well as such other documents, which may be required in this connection.

The specimen signature of Sh./ Smt.....is as given below.

(Signature 1)

(Signature 2)

I, certify that I am <designation> of <Bidder>, and that Sh./ Smt..... whose specimen signatures are given above is authorized to bind the corporation by authority of its governing body.

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name

Designation

Email

Mobile No.

Annexure - II

(On Bidder Letter Head)

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)
Corporate & Registered Office
HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012,
Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

Subject: Bid Form

This has a reference to your Tender no.....dated for Supply, Installation, Testing & Commissioning and Onsite Support of Medical Equipments and Materials to be Supplied to Cote D'Ivoire (Ivory Cost) .

Having examined the 'Invitation for Bids', 'Instructions to Bidders' 'Special Conditions of Contract', 'Technical Compliance Document', 'Schedule of Requirements' and the Annexure for the above referred Tender, we the undersigned offer to supply, install, test & commission and guarantee the whole of the said 'Scope of Work' in conformity with the said Special Conditions of Contract and Technical Specifications for the sum mentioned in Commercial Bid submitted separately, or such other sum as may be ascertained in accordance with the conditions. The details of the items/ services for which we have submitted our Technical Bid and for which we have quoted the rates in our Financial Bid are given in required formats.

We have independently considered each and every clause of this Tender document and given our informed consent to them.

We undertake, if our Bid is accepted, we shall commence the supply of items/ services/ manpower as per delivery schedule offered by us so as to fulfil our obligations as per this Tender Document to full satisfaction of purchaser.

If our Bid is accepted we will furnish a Bank Guarantee of 10% order value with a validity of 24months to cover delivery, installation, Testing & warranty period for the equipments.

We have independently considered the amount shown in 'Special Conditions of Contract' as per-estimated liquidated damages and agree that they represent a fair estimate of the damages likely to be suffered by you in the event of the work not being completed in time.

We agree to abide by this Bid for a minimum period of 180 days from the date of opening and it shall be remain binding upon us and may be accepted at any time before the expiration of that period or any extended period mutually agreed to.

This bid, together with any further clarification/ confirmation given by us and your written acceptance thereof, shall constitute a binding contract between us.

We understand that you are not bound to accept the lowest or any offer you may receive against this bid.

If our Bid is accepted, we understand that we are to be held solely responsible for the due performance of the Contract.

All enclosures and relevant documents forming a part of the proposal are complete and attached herewith duly verified by officials authorized to do the same. All documents submitted are genuine and if HLL Lifecare Ltd finds that we have submitted any forged document, HLL Lifecare Ltd may reject our bid; forfeit the EMD/ BG and take stringent action against us as per Government of India guidelines.

The Bidder shall indemnify, defend and hold harmless Government of India and HLL, its Affiliates, officers, directors, employees, agents, and their respective successors and assigns, from and against any and all loss, damage, claim, injury, cost or expenses (including without limitation reasonable attorney's fees), incurred in connection with third Party claims of any kind that arise out of or are attributable to (i) Manufacturer's/Bidders breach of any of its warranties, representations, covenants or obligations set forth herein or (ii) the negligent act or omission of the Manufacturer /Bidders.(iii) any product liability claim arising from the gross negligence or bad faith of, or intentional misconduct or intentional breach of this Contract by bidder or its affiliate.

We will accept HLL Lifecare Ltd's internal technical and financial evaluation procedure and will not interfere in the process after submission of the bid. We shall not deviate from the bid process and not try to stall the process; if do so, HLL Lifecare Ltd may take stringent action against us.

We undertake that our Bidder is not under the declaration of ineligibility for corrupt & fraudulent practices. We also undertake that our Bidder has not been blacklisted by any Government entity in India/ abroad till today. We will follow all guideline mentioned in the Tender documents.

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name
Designation
Email
Mobile No.

Annexure - III

(On Bidder Letter Head)

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012,
Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

Subject: Declaration Letter

This has a reference to your Tender no.....dated..... for Supply, Installation, Commissioning and Onsite Support of Medical Imaging Equipments and Material to be supplied to Cote D'Ivoire (Ivory Cost) and as per schedule A respectively.

We confirm that the equipment supplied against Tender no:.....dated..... will not be refurbished/ duplicate or acquired illegally in India/ third country. We undertake, if our bid is accepted, we will submit a letter from OEM endorsing that the equipment are brand-new at the time of supply.

The details of all the OEMs, make & model with detail technical specifications is offered in our technical bid for items. If HLL Lifecare Ltd find that any information/ details submitted by us is incorrect, HLL Lifecare Ltd may black list us and forfeit the EMD/ take stringent action against us as per Government of India guidelines.

By signing this declaration letter we accept all terms and conditions specified in the Tender document and agree that HLL Lifecare Ltd's decision will be final and binding on us.

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name
Designation
Email
Mobile No.

Annexure - IV

(On Bidder Letter Head)

Date:.....

Tender No.:.....

To

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

HLL Bhavan, Poojappura,

Thiruvananthapuram - 695012,

Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

Sub: Undertaking to the effect that onsite support during the warranty in Cote D'Ivoire (Ivory Cost) will be provided by <Bidder Name> through its own Branch Office/ Service Centre/local partner

This has a reference to your Tender no.....dated for Supply, Installation, Commissioning and Onsite Support of Medical Imaging Equipments to be Supplied to Cote D'Ivoire (Ivory Cost) as per schedule A.

We undertake to provide the onsite support in Cote D'Ivoire (Ivory Cost) as per schedule A during the period of warranty through our own Branch Office/ Service Centre/ local partner or In case <Bidder Name> has any other mechanism to provide support and services for warranty period, please specify the mechanism

<Bidder Name> has to provide details viz. Name, Address, Contact Person, Telephone/ Fax, mail etc. of our own Branch Office/ Service Centre/ local partner along with an undertaking from the local dealer/ service provider (if applicable) will be submitted by <Bidder Name> before the placement of supply order by HLL Lifecare Ltd.,

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name

Designation

Email

Mobile No.

Annexure - V

PERFORMA FOR BANK GUARANTEE TOWARDS EMD

(on non-judicial paper of appropriate value)

Bank Guarantee No. Dated: To

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

HLL Bhavan, Poojappura,

Thiruvananthapuram - 695012,

Kerala, India

Dear Sir(s),

Whereas the HLL Lifecare Ltd., HLL Bhavan, Poojappura, Thiruvananthapuram - 695012, Kerala, India (hereinafter called the HLL) which expression shall, unless repugnant to the context or the HLL thereof, include all its successors, administrators, executors and assignees has invited Tender No. _____ and M/s _____ having Registered /head office at _____ (hereinafter called the "Bidder" which expression shall, unless repugnant to the context or the HLL thereof, HLL and include all its successors, administrators executors and assignees) have submitted a quotation Reference No. _____ and Bidder having agree to furnish as a conditions precedent for participation in Tender as unconditional and irrevocable bank guarantee of Rs _____ (Rupees _____ only) for the due performance of _____

Bidder's obligations as contained in the terms of the Notice inviting Tender and other terms and conditions contained in the Tender Documents supplied by the HLL specially the conditions that (a) bidder shall keep his bid open for a period of day i.e. from , _____ to _____ or any extension thereof, and shall not withdraw or modify it in a manner not acceptable to the HLL (b) the Bidder will execute the contract, if awarded, and shall furnish performance guarantee in the format prescribed by the HLL within the required time. The Bidder has absolutely and unconditionally accepted these conditions. The HLL and the Bidder have agreed that Tender document is an offer made on the condition that the bids, if submitted would be kept open in its original form without variation or modification in a manner not acceptable to the HLL for a period of _____ days i.e. from _____ to _____ or any, extension thereof and that submission of the bid itself shall be regarded as an unconditional and absolute acceptance of the conditions, contained in the Tender documents. They have further agreed that the contract consisting of Tender documents as the OFFER and submission of the bids as the ACCEPTANCE shall be a separate contract distinct from the contract, which will come into existence when the bid is finally accepted by the HLL. The consideration for this separate initial contract preceding the main contract is that the HLL is not agreeable to sell the Tender documents to the Bidder and to consider the Tender to be made except on the condition that the bids shall be kept open for the period indicated above and the Bidder desires to submit bid on this condition after entering into this separate initial contract with the HLL promises to consider the Tender on this condition and Bidder agrees to keep this bid open for the required period. These reciprocal promises form the CONSIDERATION for this separate initial contract between the parties.

Therefore, we _____ registered (indicate the name of Bank) under the laws of _____ having head /registered office at (hereinafter referred to as the "Bank") which expression shall, unless repugnant to the context or HLL thereof, include all its successors, administrators and executors hereby issue irrevocable and unconditional bank

guarantee and undertake to pay immediately on first demand in writing Rupees all money to the extent of Rs _____ (Rupees _____ only) at any time immediately on such demand without any demur, reservations, recourse, contest or protest and/ or without any reference to the Bidder and any such demand made by the HLL on the bank shall be conclusive and binding notwithstanding any difference between the HLL and the Bidder or any dispute pending before any court/arbitrator or any other matter whatsoever. We also agree to give that Guarantee herein the HLL in writing. This guarantee shall not be determined / discharged / affected by the liquidation, winding up, dissolution or insolvency of the Bidder and will remain valid, binding and operative against the bank.

The bank also undertakes that the HLL at the option shall be entitled to enforce this guarantee, against the Bank as a principal debtor, in the first instance, without proceeding against the Bidder.

The bank further agree that as between the bank and the HLL, purpose of the guarantee, any notice of the breach of the terms and conditions contained in the bid Documents as referred above given to the bank by the HLL shall be conclusive and binding on Bank, without any proof, notwithstanding any other matter or difference or dispute whatsoever. We further agree that this guarantee shall not be affected by any change in our constitution, in the constitution of the HLL or that of the Bidder. We also undertake not to revoke, in any case, this Guarantee during its currency.

The bank agree with the HLL that the HLL shall have the fullest liberty without our consent and without affecting in any manner our obligations hereunder to vary any of the terms of the Tender or get extension of the validity period from time to time. We shall not be relieved from our liability by reason of any such variation or extension of the validity period or for any forbearance, act of omission and commission on the part of the HLL or any indulgence shown by the HLL to the said Bidder or by any such matter or thing whatsoever which under the law relating to sureties, would, but for this provision, have the effect of so relieving us.

Notwithstanding anything contained here in above our liability under his Guarantee is limited to Rs. _____ (Rupees _____ only) in aggregate and it shall remain in full force up to _____ (24 months from the date of bid opening) unless extended further from time to time, for such period as may be instructed in writing by M/s _____ on whose behalf this guarantee has been given, in which case, it shall remain in full force up to the expiry of extended period. Any claim under this guarantee must be received by us before _____ (date of expiry of validity period) or before the expiry of extended period, if any. If no such claim is received by us within the said date/extended date, the rights of the HLL under this guarantee will cease. However, if such a claim has been received by us within and upto the said date/extended date, all right of the HLL under this guarantee shall be valid and shall not cease until we have satisfied that claim.

In case contract is awarded to the Bidder here in after referred to as "Contractor" the validity of this Bank Guarantee will stand automatically extended until the Bidder furnished to the HLL a bank guarantee for requisite amount towards performance guarantee for satisfactory performance of the contract. In case of failure to furnish performance bank Guarantee in the format prescribed by the HLL by the required date the claim must be submitted to us within validity period or extended period, if any. If no such claim has been received by us within the said date /extended date, rights, of the Ministry under this guarantee will cease. However if such a claim has been received by us within the said date/extended date all rights of the HLL under this guarantee shall be valid and shall not cease until we have satisfied that claim,

In witness where of the Bank, through its authorised officer, has sent its hand & stamp on this _____ day of _____ of _____ at _____ of _____ (month & year).

Signature
(Full name in capital letters)

Designation with bank stamp
(Full name and address in capital letters)

Attorney as per power of attorney

No

Date.....

PROFORMA FOR BANK GUARANTEE TOWARDS PERFORMANCE GUARANTEE

(on non-judicial paper of appropriate value)

Dated:

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012,
Kerala, India

BANKS GUARANTEE NO:

Dear Sir(s),

This has reference to the Supply Order No . _____ dated _____ placed by HLL Lifecare Ltd, Thiruvananthapuram on M/s , _____ for Supply, Installation, Commissioning and Onsite Support of Medical Imaging Equipments to be Supplied to Cote D'Ivoire (Ivory Cost) as per schedule A

The conditions of this order provide that the vendor shall,

- Arrange to deliver the items listed in the said order to the consignee, as per details given in said order, and
- Arrange to install and commission the items listed in said order at client's site, to the entire satisfaction of HLL Lifecare Ltd and
- Arrange for the comprehensive warranty service support towards the items supplied by vendor on site in Cote D'Ivoire (Ivory Cost) as per schedule A as per the warranty clause in said purchase order.

M/s _____ has accepted the said purchase order with the terms and conditions stipulated therein and have agreed to issue the performance bank guarantee on their part, towards promises and assurance of their contractual obligations vide the Supply Order No. _____ on M/s. _____ holds an account with us and has approached us and at their request and in consideration of the promises, we hereby furnish such guarantees as mentioned hereinafter.

HLL shall be at liberty without reference to the Bank and without affecting the full liability of the Bank hereunder to take any other undertaking of security in respect of the suppliers obligations and / or liabilities under or in connection with the said contract or to vary the terms vis-a-vis the supplier or the said contract or to grant time and or indulgence to the supplier or to reduce or to increase or otherwise vary the prices or the total contract value or to forebear from enforcement of all or any of the obligations of the supplier under the said contract and/or the remedies of HLL under any security (ies) now, or hereafter held by HLL and no such dealing(s) with the supplier or release or for bearance whatsoever shall have the effect of releasing the bank from its full liability of HLL hereunder or of prejudicing right of HLL against the bank.

This undertaking guarantee shall be a continuing undertaking guarantee and shall remain valid and irrevocable for all claims of HLL and liabilities of the supplier arising up to and until _____.

This undertaking guarantee shall be in addition to any other undertaking or guarantee or security whatsoever the that HLL may now or at any time have in relation to its claims or the supplier's obligations/liabilities under and/ or in connection with the said contract and HLL shall have the full authority to take recourse to or enforce this undertaking guarantee in preference to the other undertaking or security (ies) at its sole discretion and no failure on the part of HLL in enforcing or requiring enforcement of any other undertaking or security shall have the effect of releasing the bank from its full liability hereunder.

We _____ hereby agree and irrevocably undertake and promise that if in your (HLL's) opinion any default is made by M/s _____ in performing any of the terms and/or conditions of the agreement or if in your opinion they commit any breach of the contract or there is any demand by you against M/s _____, then on notice to us by you, we shall on demand and without demur and without reference to M/s _____, pay you, in any manner in which you may direct, the amount of Rs. _____ /- (Rupees _____ Only) or such portion thereof as may be demanded by you not exceeding the said sum and as you may from time to time require. Our liability to pay is not dependent or conditional on your proceeding against M/s _____ and we shall be liable & obligated to pay the aforesaid amount as and when demanded by you merely on an intimation being given by you and even before any legal proceedings, if any, are taken against M/s _____. The Bank hereby waives all rights at any time inconsistent with the terms of this undertaking guarantee and the obligations of the bank in terms hereof shall not be anywise affected or suspended by reason of any dispute or disputes having been raised by the supplier (whether or not pending before any arbitrator, Tribunal or Court) or any denial of liability by the supplier or any order or any order or communication whatsoever by the supplier stopping or preventing or purporting to stop or prevent payment by the Bank to HLL hereunder.

The amount stated in any notice of demand addressed by HLL to the Bank as claimed by HLL from the supplier or as suffered or incurred by HLL on the account of any losses or damages or costs, charges and/or expenses shall as between the Bank and HLL be conclusive of the amount so claimed or liable to be paid to HLL or suffered or incurred by HLL, as the case may be and payable by the Bank to HLL in terms hereof.

You (HLL's) shall full liberty without reference to us and without affecting this guarantee, postpone for any time or from time to time the exercise of any of the powers and rights conferred on you under the contract with the said M/s _____ and to enforce or to forbear from endorsing any power or rights or by reason of time being given to the said M/s _____ which under law relating to the sureties would but for the provisions have the effect of releasing us.

You will have full liberty without reference to us and without affecting this guarantee, postpone for any time or from time to time the exercise of any of the powers and rights conferred on you under the contract with the said M/s _____ and to enforce or to forbear from endorsing any power or rights or by reason of time being given to the said M/s _____ which under law relating to the sureties would but for the provisions have the effect of releasing us.

Your right to recover the said sum of Rs. _____ (Rupees , _____ only) from us in manner aforesaid will not be affected/or suspended by reason of the fact that any dispute or disputes have been raised the said M/s _____ and/ or that any dispute or disputes are pending before any officer, tribunal or court or Arbitrator.

The guarantee herein contained shall not be determined or affected by the liquidation or winding up, dissolution or change of constitution or insolvency of the said M/s. _____ but shall in all respects and for all purposes be binding and operative until payment of all dues to HLL in respect of such liability or liabilities.

Our liability is restricted to under this guarantee Rs. _____ (Rupees _____ Only).
Our guarantee shall remain in force until unless a suit action to enforce a claim under guarantee is filed against us within three months from (which is date of expiry of guarantee) all your rights under the said guarantee shall be forfeited and we shall be relieved and discharged from all liabilities there under.

We have power to issue this guarantee in your favour under Memorandum and Articles of Association of our Bank and the undersigned has full power to do under the power of Attorney dated.

Notwithstanding anything contained herein:

- Our liability under this guarantee shall not exceed Rs _____ (in words)
- This bank guarantee shall be valid up to& unless a suit for action to enforce a claim under guarantee is filed against us within three months from the date of expiry of guarantee. All your rights under the said guarantee shall be forfeited and we shall be relieved and discharged from all liabilities there after i.e. after three months from the date of expiry of this Bank guarantee
- We are liable to pay the guaranteed amount or any parts thereof under this bank guarantee only and only if you serve upon us a written claim or demand on or before _____
- The Bank guarantee will expire on _____ granted by the Bank

Yours faithfully,

For (Name of Bank)

SEAL OF THE BANK

Signatory

Annexure - VII

CHECK LIST (On Bidder Letter Head)

SNo	Description	Details	Attached(Yes/No)
1	Bidder need to submit a covering page mentioning in which all Schedules bidder is participating and sequencing of documents for each and every page with page number.		
2	Name and Address of the bidder & its consortium partners(clearly marked with separators) and its incorporation details (e.g. copy of MOA and AOA)		
4	Copy of Earnest Money Deposit of Rs.1,00,000.00 (please refer to section II Point 6. for complete details and Tender Fee of Rs.3,000/-		
5	Duly filled Technical Bid with proper seal and signature of authorized person on each page of the bid submitted. Certificate of authority should be scanned and submitted as perform at given at Annexure I		
6	Both bidder and its consortium partners need to provide copy of Supply Orders for/Contracts/ Agreements for similar work, executed by the bidders in last 5 (five) years ending with December 31, 2016 along the work orders / letter of award / completion certificate duly issued by the end user. Copy of the purchase order and work completion report duly notarized along with contact details of the end user. For details of work orders and turnover and other necessary documentation please refer to section II; point 3.c for complete details and amount. Make sure to categorize describe in technical bid for how many schedules bidder is participating with sequencing of page numbers		
7	Both Bidder and its consortium partner need to provide GST number and PAN number		
8	Last three financial years annual turnover. For turnover details Bidder & consortium partner need to refer to point 3.c for complete details.		
9	Bid Form by the bidder as per format given at Annexure II		
10	Declaration by the bidder as perform at given at Annexure III		
11	Undertaking to the effect that onsite support during the warranty in Cote D'Ivoire (Ivory Cost) will be provided by onsite support through local dealer / service provider as per format given at Annexure IV		
12	Signed and Scanned Copy of Check List - Annexure VII		
13	Technical Compliance Document (Annexure VIII) with detailed technical specification, make, model and leaflets/brochures in compliance with the data sheets or OEM Compliance (please refer to Annexure VIII)		
14	Annexure X		
15	Annexure XI		
16	Annexure XII		
17	Copy of Non Conviction certificate		
18	Copy of Supply orders/Contract/Agreements for similar work executed by bidders		
19	Copy of all other relevant Documents, if any apart from mentioned above as mentioned in Section II - Invitation of Bids, Clause 4 - Part I		

Thanking you,

Sincerely yours,

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name

Designation

Email

Mobile No.

Annexure - VIII

(On Bidder Letter Head)

TECHNICAL COMPLIANCE SHEET

Tender Title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

(Bidder need to put "Yes" in compliance column only when 100% compliance as per specification mentioned Section - V; else need to mention deviation specifically in remarks column. Bidder need to provide sufficient data sheets of products also to justify compliance)

Technical Compliance Sheet

Sl · N o.	Description	Quoted or Not	100% Compliance as per specification mentioned Section V	If No, Please mention the details of Deviation
1	Mucosity vacuum cleaner / Suction cup			
2	Autoclave Cocotte/ Table top Autoclave 20-25 litre capacity			
3	Basin shape plastic slipper (plastic basin), Plastic bowl, Urinal plastic women			
4	Suture box (With Instruments & SS box):			
5	Complete Delivery box (With Instruments & SS Box)			
6	Episiotomy box (With instruments & SS Box)			
7	Care trolley (baby carrying bassinet)			
8	Step ladder (2 steps) / Foot Step			
9	Insufflator (resuscitator) manual baby			
10	Inpatient bed + accessories / Bed observation + mattress			
11	Stainless steel kidney tray / Stainless steel trays with lids / sterilization drums			
12	Baby scale with direct display			
13	Weigh Machine(150 kg)/ Bathroom scales			
14	Serum hanger (IV Stand) / Stem on wheel			
15	Waste Bin Pedal Operated(20L)-SS			
16	Double head stethoscope/Standard bi-auricular stethoscope			
17	Fetal Doppler Echo			
18	Delivery Bed (Childbirth bed/Table)			

19	BP Machine (Adult)			
20	BP machine (Pediatric)			
21	Infrared Thermometer			
22	Tensioning device (Salters scale and complete tripod stand Hanging)			
23	Neonatal vacuum cleaner			
24	Basin			
25	BedPan			
26	Box of dressing			
27	Stainless steel bin/ Average stainless steel bin INOX			
28	Heating Lamp			
29	Gynecological / Examination Light Stands			
30	Delivery Kit			
31	Pediatric Examination Table (Pediatric Bed + Mattress)			
32	Mask and new-born balloon:			
33	Mattress (adult)			
34	New-born mattress			
35	Endo-uterine vacuum cleaner			
36	Neonatal Resuscitation Unit			
37	Monitors (Patient Monitors)			
38	Oxygen (SpO2 monitor)/Pulse-oxymeter			
39	Bed Side Screen/ Three part Screen:			
40	Nursing tray with lid			
41	Medium care tray with lid			
42	Stainless Pedal Wastebin			
43	Poupinel Sterilizer / Dry Heat Sterilizer			
44	Sterile speculum			
45	Obstetrical stethoscope			
46	Gynaecological examination table			
47	Heating table (Warmer)			
48	Neonatal Open Care System			
49	Sphygmomanometer			
50	Spengler blood pressure monitor			
51	Stadiometer(Height Measurement)			
52	Infantometer (Baby Height Measurement)			

Thanking you,

Sincerely yours,

For M /s _____ (Name of the bidder)

Signature & Bidder seal

Name

Designation

Annexure - IX

(On Bidder Letter Head)

(Bidder need to sign and submit blank copy of BOQ along with other essential documents asked above in check list)

Pricing Sheet (Inclusive of all Taxes, Transportation and all other kind of charges In INR)

Cote D'Ivoire (Ivory Cost) (Schedule A)

Annexure IX (a)

Name of the bidder:

Ref: HLL/SD/RBD/2018-19/TENDER/10 DT. 24.10.2018

Price Format (offered in INR)											
1	2	3	4	5	6	7	8	9	10	11	12
Sl. No.	Description of goods as per specification	Make/Model	Country of Origin	HS/HSN Code	Unit of measurement	Quantity (as per Schedule A)	Unit Price at Consignee Site (excluding GST)	Applicable GST (%)	Applicable GST value/ unit (8 x 9)	Unit Price at Consignee Site (8+10)	Total Price at Consignee Site (7 x 11)

Note: 1. All the information must be entered in the relevant columns of the price format.

2. Any Incidental Services (including Installation, Testing & Commissioning, Supervision, Demonstration and Training) at the Consignee site would be the responsibility of bidder only

4. It is core responsibility of bidder to understand the site before bidding and make sure arrangement being provided by Government of Cote D'Ivoire (Ivory Cost) as per schedule A.

5. Any, remaining parts such as: Base for Machines, Sheds, Electrical Wire, Water Pipeline or any kind of additions / changes / modifications required would be the responsibility of bidder only

6. It is the responsibility of bidder to arrange all parts for installation which is not covered by local government to be locally done within given timeline.

7. Bidder has to make arrangement to provide services / spare parts for all quoted machines in short time. It can be done by visit of engineer or setting up a local officer or by availing spare parts in advance.

Annexure IX (c)

Consumables and Spares Quotation							
S.No	Description of goods as per specification	Category Spare/ Consumable	Spare or Consumable Description	Ordering Code	Unit of Measurement	Packing Size	Unit Price Incl. of all taxes
1							
2							
3							
4							

Note: The price must be fixed and valid up to a period of 5 years from the date of Installation and commissioning at site

Annexure - X

(Undertaking On Bidder Letter Head)

Date:.....
Tender No.:.....
To

Senior Manager (SD-RBD)
HLL Lifecare Ltd.
(A Govt. of India Enterprise)
HLL Bhavan, Poojappura,
Thiruvananthapuram - 695012,
Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments and Material To Cote D'Ivoire (Ivory Cost)

Sub: Confirmation of consortium partner and participation

This has a reference to your Tender no.....dated for Supply, Installation, Commissioning and Onsite Support of Medical Imaging Equipments and material to be Supplied to Cote D'Ivoire (Ivory Cost) as per schedule A.

<Bidder Name>; We as an bidder confirm that we have a consortium partner <name of consortium partner> for<Schedule X>. Also we confirm that we will not change the consortium partner during or after the bid during complete project period. Change in any partner will lead us to disqualify from project immediately

Thanking you,

For M /s _____
(Name of the bidder)

Thanking you,

For M /s _____
(Name of the Consortium Partner)

Signature & Bidder seal of Bidder

Signature & Bidder seal of Consortium partner

Bidder :
Name :
Designation :
Email :
Mobile No. :

Consortium Partner :
Name :
Designation :
Email :
Mobile No. :

Annexure - XI

(Undertaking On Bidder Letter Head)

Date:

Tender No.:

Senior Manager (SD-RBD)

HLL Lifecare Ltd.

(A Govt. of India Enterprise)

HLL Bhavan, Poojappura,

Thiruvananthapuram - 695012,

Kerala, India

Dear Sir,

Tender title: Tender For The Supply, Installation, Testing & Commissioning And Onsite Support For Medical Equipments And Material To Cote D'Ivoire (Ivory Cost)

Sub: Confirmation of non participation of Consortium Partner as a Bidder

This has a reference to your Tender no.....dated Supply, Installation, Commissioning and Onsite Support of Medical Imaging Equipments and material to be supplied to Cote D'Ivoire (Ivory Cost) as per schedule A.

<Bidder Name>; we as an bidder confirm that our consortium partner <name of consortium partner> for<Schedule X> is not bidding himself as a bidder in this project.

<Consortium partner name>; We as a consortium partner confirm that we will not participate as a bidder in this project in any respect.

We both (Bidder and consortium partner(s)) abide to and confirm that if consortium partner is found to be bidding himself as a bidder in this project, both of us will be disqualified from project immediately.

Thanking you,

Thanking you,

For M /s _____
(Name of the bidder)

For M /s _____
(Name of the Consortium Partner)

Signature & Bidder seal of Bidder

Signature & Bidder seal of Consortium Partner

Bidder :
Name :

Consortium Partner :
Name :

Designation :
Email :
Mobile No. :

Designation :
Email :
Mobile No. :

Annexure - XII

PRE-CONTRACT INTEGRITY PACT

This Pre-Contract Integrity Pact (herein after called the Integrity Pact) is made on -----^t day of the month of -----,

Between

HLL Life Care Limited, a Government of India Enterprise with registered office at HLL Bhavan, Poojappura, Thiruvananthapuram 695 012, Kerala, India. (Hereinafter called “HLL”, which expression shall mean and include, unless the context otherwise requires, his successors in office and assigns) of the First Party.

And

----- India represented by Shri -----
----- (hereinafter called the “BIDDER / Seller” / Contractor which expression shall mean and include, unless the context otherwise requires, his successors and permitted assigns) of the Second Party.

Preamble

[Both HLL and BIDDER referred above are jointly referred to as the Parties]

HLL intends to award, under laid down organizational procedures, Purchase orders / contract/s against Tender /Work Order. HLL desires full compliance with all relevant laws and regulations, and the principles of economic use of resources, and of fairness and transparency in its relations with its Bidder/s and Contractor/s.

NOW, THEREFORE,

To avoid all forms of corruption by following a system that is fair, transparent and free from any influence /prejudiced dealings prior to, during and subsequent to the currency of the contract to be entered into with a view to:-

1. Enable HLL to obtain the desired materials/ stores/equipment/ work/ project done at a competitive price in conformity with the defined specifications by avoiding the high cost and the distortionary impact of corruption on public procurement; and
2. Enable the BIDDER to abstain from bribing or indulging in any corrupt practice in order to secure the contract by providing assurance to them that their competitors will also abstain from bribing and other corrupt practices and HLL will commit to prevent corruption, in any form, by its officials by following transparent procedures.

The parties hereto hereby agree to enter into this Integrity Pact and agree as follows:

Clause.1. Commitments of HLL

- 1.1 HLL undertakes that HLL and /or its Associates (i.e. employees, agents, consultants, advisors, etc.) will not demand, take a promise for or accept, directly or through intermediaries, any bribe, consideration, gift, reward, favour or any material or immaterial benefit or any other advantage from the BIDDER, either for themselves or for any person, organization or third party related to the contract in exchange for an advantage in the bidding process, bid evaluation, contracting or implementation process related to the contract.

- 1.2 HLL will, during the Tender process / pre-contract stage, treat all BIDDERS with equity and reason, and will provide to all BIDDERS the same information and will not provide any such information or additional information, which is confidential in any manner, to any particular BIDDER which could afford an advantage to that particular BIDDER in comparison to other BIDDERS in relation to Tendering process or during the contract execution.
- 1.3 All the officials of HLL will report to Chief Vigilance Officer of HLL (CVO), any attempted or completed breaches of the above commitments as well as any substantial suspicion of such a breach.
- 1.4 HLL will exclude from the process all known prejudiced persons and persons who would be known to have a connection or nexus with the prospective bidder.
- 1.5 If the BIDDER reports to HLL with full and verifiable facts any misconduct on the part of HLL's Associates (i.e. employees, agents, consultants, advisors, etc.) and the same is prima facie found to be correct by HLL, necessary disciplinary proceedings, or any other action as deemed fit, including criminal proceedings may be initiated by HLL. Further, such an Associate may be debarred from further dealings related to the contract process. In such a case, while an enquiry is being conducted by HLL the proceedings under the contract would not be stalled.

Clause 2. Commitments of BIDDERS/ CONTRACTORS

2. The BIDDER commits itself to take all measures necessary to prevent corrupt practices, unfair means and illegal activities during any stage of its bid or during any pre-contract or post-contract stage in order to secure the contract or in furtherance to secure it and in particular commit itself to the following:-
 - 2.1 The BIDDER will not offer, directly or indirectly (i.e. employees, agents, consultants, advisors, etc.) any bribe, gift, consideration, reward, favour, any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of HLL, connected directly or indirectly with the bidding process, or to any person, organization or third party related to the contract in exchange for any advantage in the bidding, evaluation, contracting and implementation of the contract.
 - 2.2 The BIDDER further undertakes that it has not given, offered or promised to give, directly or indirectly any bribe, gift, consideration, reward, favour, any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of HLL or otherwise in procuring the contract or forbearing to do or having done any act in relation to obtaining or execution of the contract or any other contract with the Government for showing or forbearing to show favour or disfavor to any person in relation to the contract or any other contract with the Government.
 - 2.3 The BIDDER will not engage in collusion, price fixing, cartelization, etc. with other counterparty(s).
 - 2.4 The counterparty will not pass to any third party any confidential information entrusted to it, unless duly authorized by HLL.
 - 2.5 The counterparty will promote and observe ethical practices within its Organization and its affiliates.

- 2.6 BIDDER shall disclose the name and address of agents and representatives and Indian BIDDERS shall disclose their foreign principals or associates.
- 2.7 The counterparty will not make any false or misleading allegations against HLL or its Associates.
- 2.8 BIDDERS shall disclose the payments to be made by them to agents / brokers or any other intermediary, in connection with this bid/contract.
- 2.9 The BIDDER further confirms and declares to HLL that the BIDDER is the original integrator / manufacture /authorized government sponsored export entity of the defense stores and has not engaged any individual or firm or company whether Indian or foreign to intercede, facilitate or in any way to recommend to HLL or any of its functionaries, whether officially or unofficially to award the contract to the BIDDER, nor has any amount been paid, promised or intended to be paid to any such individual, firm or company in respect of any such intercession, facilitation or recommendation.
- 2.10 The BIDDER while presenting the bid or during pre-contract negotiations or before signing the contract, shall disclose any payments he has made, is committed to or intends to make to officials of HLL or their family members, agents, brokers or any other intermediaries in connection with the contract and the details of services agreed upon for such payments.
- 2.11 The BIDDER will not accept any advantage in exchange for any corrupt practice, unfair means and illegal activities.
- 2.12 The BIDDER commits to refrain from giving any complaint directly or through any other manner without supporting it with full and verifiable facts.
- 2.13 If the BIDDER or any employee of the BIDDER or any person acting on behalf of the BIDDER, either directly or indirectly, is a relative of any of the officers of HLL, or alternatively, if any relative of an officer of HLL has financial interest /stake in the BIDDER's firm, the same shall be disclosed by the BIDDER at the time of filing of Tender.

The term 'relative' for this purpose would be as defined in Section 6 of the Companies Act 1956.
- 2.14 The BIDDER shall not lend to or borrow any money from or enter into any monetary dealings or transactions, directly or indirectly, with any employee of HLL.
- 2.15 The BIDDER will not collude with other parties interested in the contract to impair the transparency, fairness and progress of the bidding process, bid evaluation, contracting and implementation of the contract, and will not enter into any undisclosed agreement or understanding with other Bidders, whether formal or informal. This applies in particular to prices, specifications, certifications, subsidiary contracts, submission or non-submission of bids or any other actions to restrict competitiveness or to introduce cartelization in the bidding process.
- 2.16 The BIDDER will not commit any offence under the relevant Indian Penal Code, 1860 or Prevention of Corruption Act, 1988; further the Bidder(s)/ Contractor(s) will not use improperly, for purposes of competition or personal gain, or pass on to others, any information or document provided by the HLL as part of the business relationship, regarding plans, technical proposals and business details, including information contained or transmitted electronically. The BIDDER also undertakes to exercise due and adequate care lest any such information is divulged.

2.17 The BIDDER will not instigate third persons to commit offences outlined above or be an accessory to such offences.

2.18 The Bidder(s)/Contractors(s) of foreign origin shall disclose the name and address of the Agents /representatives in India, if any. Similarly the Bidder(s) /Contractors(s) of Indian Nationality shall furnish the name and address of the foreign Principal(s), if any.

Clause.3. Previous contravention and Disqualification from Tender process and exclusion from future contracts

3.1 The BIDDER declares that no previous contravention occurred in the last three years immediately before signing of this Integrity Pact, with any other company in any country in respect of any corrupt practices envisaged hereunder or with any Public Sector Enterprise in India or any Government Department in India that could justify BIDDER's exclusion from the Tender process

3.2 The BIDDER agrees that if it makes incorrect statement on this subject, BIDDER can be disqualified from the Tender process or the contract, if already awarded, can be terminated for such reason.

If BIDDER before award or during execution has committed a contravention through a violation of Clause 2, above or in any other form such as to put his reliability or credibility in question, HLL is entitled to disqualify the BIDDER from the Tender process.

Clause .4. Equal treatment of all Bidders / Contractors /

Subcontractors

4.1 The Bidder(s) / Contractor(s) undertake(s) to demand from his Subcontractors a commitment in conformity with this Integrity Pact.

4.2 HLL will enter into agreements with identical conditions as this one with all Bidders and Contractors.

4.3 HLL will disqualify from the Tender process all bidders who do not sign this Pact or violate its provisions.

Clause .5. Consequences of Violation / Breach

5.1 Any breach of the aforesaid provision by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER) shall entitle HLL to take all or any one of the following action, wherever required:-

i. To immediately call off the pre-contract negotiations without assigning any reason or giving any compensation to the BIDDER. However, the proceedings with the other

ii. If BIDDER commits violation of Integrity Pact Policy during bidding process, he shall be liable to compensate HLL by way of liquidated damages amounting to a sum equivalent to 5% to the value of the offer or the amount equivalent to Earnest Money Deposit /Bid Security, whichever is higher.

iii. In case of violation of the Integrity Pact after award of the contract, HLL will be entitled to terminate the contract. HLL shall also be entitled to recover from the contractor liquidated damages equivalent to 10% of the contract value or the amount equivalent to security deposit/ performance guarantee, whichever is higher.

- iv. To immediately cancel the contract, if already signed, without giving any compensation to the BIDDER.
 - v. To recover all sums already paid by HLL, and in case of an Indian BIDDER with interest thereon at 2% higher than the prevailing Prime Lending Rate of State Bank of India, while in case of a BIDDER from a country other than India with interest thereon at 2% higher than the LIBOR. If any outstanding payment is due to the BIDDER from HLL in connection with any other contract for any other stores, such outstanding payment could also be utilized to recover the aforesaid amount.
 - vi. To encash the advance bank guarantee and performance guarantee / warranty bond, if furnished by the BIDDER, in order to recover the payments already made by HLL, along with interest.
 - vii. To cancel all or any other contract with the BIDDER. The BIDDER shall be liable to pay compensation for any loss or damage to HLL resulting from such cancellation/recession and HLL shall be entitled to deduct the amount so payable from the money(s) due to the BIDDER.
 - viii. To debar the BIDDER from participating in future bidding processes of HLL for a minimum period of five (5) years, which may be further extended at the discretion of HLL or until Independent External Monitors is satisfied that the Counterparty will not commit any future violation.
 - ix. To recover all sums paid in violation of this Pact by BIDDER(s) to any middleman or agent or broker with a view to securing the contract.
 - x. In cases where irrevocable Letters of credit have been received in respect of any contract signed by HLL with the BIDDER, the same shall not be opened.
 - xi. Forfeiture of performance guarantee in case of a decision by HLL to forfeit the same without assigning any reason for imposing sanction for violation of the pact.
- 5.2 HLL will be entitled to all or any of the actions mentioned in para 5.1(i) to (x) of this pact also on the commission by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER), of an offence as defined in Chapter IX of the Indian Penal Code, 1860 or Prevention of Corruption Act, 1988 or any other statute enacted for prevention of corruption.
- 5.3 The decision of HLL to the effect that a breach of the provisions of this Pact has been committed by the BIDDER shall be final and conclusive on the BIDDER. However, the BIDDER can approach the Independent External Monitor(s) appointed for the purposes of this Pact.

Clause.6. Fall Clause

The BIDDER undertakes that it has not supplied/is not supplying similar product/systems or subsystems OR providing similar services at a price / charge lower than that offered in the present bid in respect of any other Ministry/Department of the Government of India or PSU and if it is found any stage that similar product/systems or sub systems was supplied by the BIDDER to any to the Ministry/Department of the Government of India or a PSU at a lower price, then that very price, with due allowance for elapsed time will be applicable to the present case and the difference in the cost would be refunded by the BIDDER to HLL, if the contract has already been concluded.

Clause .7. Independent External Monitor(s)

- 7.1 HLL has appointed Independent External Monitor(s) (hereinafter referred to as Monitor(s)) for this Pact in consultation with the Central Vigilance Commission (Name and addresses of the Monitor(s) to be given).
- 7.2 The responsibility of the Monitor(s) shall be to review independently and objectively, whether and to what extent the parties comply with the obligations under this Pact.
- 7.3 The Monitor(s) shall not be subject to instructions by the representatives of the parties and perform their functions neutrally and independently.
- 7.4 Both the parties accept that the Monitor(s) have the right to access all the documents relating to the project/ procurement, including minutes of meetings.
- 7.5 As soon as the Monitor(s) notices, or has reason to believe, a violation of this pact, he will so inform the CVO.
- 7.6 The BIDDER(S) accepts that the Monitor(s) have the right to access without restriction to all project documentation of HLL including that provided by the BIDDER. The BIDDER will also grant the Monitor(s), upon his request and demonstration of a valid interest, unrestricted and unconditional access to his project documentation. The same is applicable to subcontractors engaged by the BIDDER. The Monitor(s) shall be under contractual obligation to treat the information and documents of the BIDDER/ Subcontractor(s) with confidentiality.
- 7.7 HLL will provide to the Monitor(s) sufficient information about all meetings among the parties related to the Project provided such meeting could have an impact on the contractual relation between the parties. The parties will offer to the Monitor(s) option to participate in such meetings.
- 7.8 The Monitor(s) will submit a written report to the CVO of HLL within 8 to 10 weeks from the date of reference or intimation to him by HLL/BIDDER and, should consent arise, submit proposals for correcting problematic situations.

Clause.8.Criminal charges against violating Bidder(s)/

Contractor(s)/ Subcontractor(s)

If HLL obtains knowledge of conduct of a Bidder, Contractor or Subcontractor, or of an employee or a representative or an associate of a Bidder, Contractor or Subcontractor which constitutes corruption, or if HLL has substantive suspicion in this regard, HLL will inform the same to the Chief Vigilance Officer.

Clause.9. Facilitation of Investigation

In case of any allegation of violation of any provisions of this Pact or payment of commission, HLL or its agencies shall be entitled to examine all the documents, including the Books of Accounts of the BIDDER and the BIDDER shall provide necessary information and documents in English and shall extend all possible help for the purpose of such examination.

Clause.10. Law and Place of Jurisdiction

Both the Parties agree that this Pact is subject to Indian Law. The place of performance and hence this Pact shall be subject to Thiruvananthapuram Jurisdiction.

Clause.11. Other legal Actions

The actions stipulated in the Integrity Pact are without prejudice to any other legal action that may follow in accordance with the provisions of the extant law in force relating to any civil or criminal proceedings.

Clause.12. Validity and Duration of the Agreement

This Pact begins when both parties have legally signed it. It expires for the Contractor/Successful bidder 12 months after the last payment under the contract or the complete execution of the contract to the satisfaction of the both HLL and the BIDDER /Seller, including warranty period, whichever is later, and for all other Bidders/unsuccessful bidders 6 months after the contract has been awarded.

If any claim is made / lodged during this time, the same shall be binding and continue to be valid despite the lapse of this pact as specified above, unless it is discharged / determined by Chairman and Managing Director of HLL.

Clause. 13. Other provisions

- 13.1 Changes and supplements as well as termination notices need to be made in writing. Both the Parties declare that no side agreements have been made to this Integrity Pact.
- 13.1 If the Contractor is a partnership or a consortium, this agreement must be signed by all partners or consortium members.
- 13.1 Should one or several provisions of this agreement turn out to be invalid, the remainder of this agreement remains valid. In this case, the parties will strive to come to an agreement to their original intentions

INWITNESS THEREOF the parties have signed and executed this pact at the place and date first above mentioned in the presents of following witnesses:

HLL

BIDDER

Witness

Witness

1.....

1.....

2.....

2.....

* Provisions of these clauses would be amended /deleted in line with the policy of the HLL in regard to involvement of Indian agents of foreign suppliers.