

Corrigendum No: 1 Issued on dated 27/02/2014

Ref: Tender Enquiry No: HLL/PCD/ICCONS-03/13-14 and Pre-Bid meeting held on Pre-Bid meeting held on 17/02/2014 at HLL Lifecare Ltd. Annex. Building, Tenra 22, Palathinkara,TC 24/606, Thycaud, Thiruvananthapuram - 695 014

The following amendments are issued with respect to the above tender Enquiry Document.

Section I, Notice Inviting Tenders (NIT)

For:

2. Tender No.: HLL/PCD/ICCONS-03/13-14

Sl. No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	05.02.2014 to 03.03.2014, in all working days, during 10:00 am to 4:00 pm
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Limited Procurement & Consultancy Division B-14A, Sector-62, Noida -201 307
iii.	Cost of the Tender Enquiry Document	Rs. 3,000.00
iv.	Time and date of Pre-bid meeting	17.02.2014 at 11:00 am (IST)
v.	Closing date & time for receipt of Tender	04.03.2014 at 2:00 pm. (IST)
vi.	Time and date of opening of Techno-Commercial tenders	04.03.2014 at 2:30 pm. (IST)
vii.	Venue for Pre-bid Meeting, Closing and opening of Tender	HLL Lifecare Ltd. Annex. Building, Tenra 22, Palathinkara, TC 24/606, Thycaud, Thiruvananthapuram - 695 014, Ph.0471-2330447

Read As:

Sl. No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	05.02.2014 to 13.03.2014 , in all working days, during 10:00 am to 4:00 pm
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Limited Procurement & Consultancy Division B-14A, Sector-62, Noida -201 307
iii.	Cost of the Tender Enquiry Document	Rs. 3,000.00
iv.	Time and date of Pre-bid meeting	17.02.2014 at 11:00 am (IST)
v.	Closing date & time for receipt of Tender	14.03.2014 at 2:00 pm. (IST)
vi.	Time and date of opening of Techno-Commercial tenders	14.03.2014 at 2:30 pm. (IST)
vii.	Venue for Pre-bid Meeting, Closing and opening of Tender	HLL Lifecare Ltd. Annex. Building, Tenra 22, Palathinkara, TC 24/606, Thycaud, Thiruvananthapuram - 695 014, Ph.0471-2330447

I. Item No.1 ICU VENTILATOR

(1) Existing Specification:

Point 2.3 (c) : 3 loops- P-V, F-V, P-F with facility of saving of 3 Loops for reference.

Read As:

2.3 (c) : At least 2 Loops should be provided in, P-V, F-V, P-F loops. Specify.

(2) Existing Specification:

Point 2.4 : Trending facility for 72 hours with minimum 5 minutes resolution for recent 24 hours.

Read As:

2.4 : Should have trending facility for atleast 24 hours.

(3) Existing Specification:

Point 2.6 (a) : Tidal Volume :50 to 3000 ml

Read As:

2.6 (a) : Tidal Volume : 50 to 2000 ml

(4) Added specification:

Point 2.8 (i) : **APRV**

(5) Existing Specification:

Point 2.15 (a) : Stand-alone/integrated Medical Air compressor

Read As:

2.15 (a) : Medical Air Compressor should be inbuilt compressor or mountable on the trolley. It should meet the required flow.

(6) Added specification:

Point 2.16 : Should provide ET-tube leak compensation .

(7) Existing Specification:

Technical Specifications for reusable face mask & nasal mask.

Read As:

3. Technical Specifications for reusable face mask

(8) Existing Specification:

Point 3.1: Reusable face & nasal mask with textured dual flap silicon cushion flap for easy fit.

Read As:

3.1: Reusable face with textured dual flap silicon cushion flap for easy fit.

(9) Existing Specification:

Point 4.3 (a) : Reusable Masks (Small, Medium, Large) with each machine. - 02 sets each

Read As:

4.3 (a) Reusable Masks (Small, Medium, Large) with each machine. - 05 sets each

(10) Existing Specification:

Point 4.5: Filter paper for humidifier for 100 uses - 01

Read As:

4.5 Deleted

(11) Added specification:

Point 2.17 : Trolley for mounting ventilator should be original from the same manufacturer

II. Item No.21 . 32-CHANNEL EEG SYSTEM

(1) Existing Specification:

Point 1.3.3: Noise < .2uV peak to peak.

Read As:

1.3.3 Noise < 1 .5uV peak to peak

(2) Existing Specification:

Point 1.3.9 : Acquisition Sensitivity from 1 microvolt per mm to 2000 microvolt per mm.

Read As:

1.3.9 Acquisition Sensitivity from 1 microvolt per mm to 200 uV /mm.

(3) Existing Specification:

Point 1.6.2 : Should have an option for upgrading the digital EEG to Video EEG with day/night camera using MPEG-2, 3rd generation technology

Read As:

1.6.2 : Should have an option for upgrading the digital EEG to Video EEG with day/night camera using MPEG-2, 3rd generation technology . Should quote video camera and software as OPTIONAL.

III. Item No.18 . BERA/ASSR/OAE/VEMP

(1) Existing Specification:

Point 4.1 : Should do DP and/or TE OAEs having following specifications.

Read As:

4.1 Should do DPOAE having following specifications.

(2) Existing Specification:

Point 4.1 (b) : TEOAE

Read As:

Deleted

IV. Item No.19 . OAE SCREENER

(1) Existing Specification:

Point 1.3 : System should be integrated with evoked otoacoustic emissions (OAE) with built-in printer.

Read As:

1.3 System should be integrated with evoked otoacoustic emissions (OAE) with printer.

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.

.....**The End**.....