

AMENDMENT NO: 04 Dated 01.11.2022

AMENDMENT TITLE:

Change in technical specification.

REF NO:HLL/SD/HCS/2022-23/Tender 15 dated 28th September 2022

E-TENDER FOR SUPPLY OF FULLY AUTOMATIC BIOCHEMISTRY ANALYSER - LOW THROUGHPUT (WET CHEMISTRY) FOR HLL LIFECARE DIAGNOSTIC LABORATORIES LOCATED AT VARIOUS LOCATION PAN INDIA ON REAGENT RENTAL BASIS

The following amendment has been incorporated to the bid document for the above tender;

In ANNEXURE 04, TECHNICAL SPECIFICATION AND PARAMETERS

Sr. No.	Existing Technical Specifications as per tender	Amendment
2	Should have a throughput up to or above 200 photometric tests per hour	Should have a throughput 180 and above photometric tests per hour
14	Should have photometric wavelength of 300-800 nm	Should have photometric wavelength of 340-700 nm or greater.

All relevant clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

VICE PRESIDENT (SOURCING DIVISION)
HLL LIFECARE LTD
HLL BHAVAN, POOJAPURA
THIRUVANANTHAPURAM