

**Techno Commercial Responsiveness & Non Responsiveness of Projects Concluded by PCD from July 2013 to October 2013**

Tender Ref No	Item No	Item Description	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
HLL/PCD/IMMU-03/11-12	1	Hub Cutters	Mediwaves	Responsive		Non-Responsive	Test certificate : the model passes for 5000 cuts	Non-Responsive
			S.S Medical System Pvt. Ltd.	Non-Responsive	End user certificate not submitted	Responsive		Non-Responsive
			Ratan Sales Corporation	Non-Responsive	End user certificate not submitted	Non-Responsive	Minimum number of cut not evident from the test certificate	Non-Responsive
			AOV International	Non-Responsive	End user certificate not submitted. Short EMD	Responsive		Non-Responsive
			GLO Lifecare equipment pvt. Ltd.	Non-Responsive	Performance statement not furnished.	Non- Responsive	Test certificate not submitted. Average force to penetrate not specified	Non-Responsive
	2	Ice Lined Refrigerator (small)	Voltas Ltd.	Responsive		Non- Responsive	Temperature going below 2 deg. C during continues energy supply. Certificate CE/FDA/BIS not submitted	Non-Responsive
			Haier Electrical Appliances	Responsive		Responsive		Responsive
	3	Ice Lined Refrigerator (Large)	Voltas Ltd.	Responsive		Non- Responsive	Temperature going below 2 deg. C during continues energy supply. Certificate CE/FDA/BIS not submitted	Non-Responsive
			Haier Electrical Appliances	Responsive		Responsive		Responsive
			Ratan sales Coop.	Non- Responsive	Not meeting the qualification requirement	Non- Responsive	Test certificate not submitted	Non-Responsive
	4	Vaccine Carrier	Blow King	Responsive		Responsive		Responsive ( Responsive for all schedules 4a,4d & 4e)
			Apex International	Non-Responsive	Not meeting the qualification requirement. End user certificate not submitted.	Responsive		Non-Responsive
			AOV International	Responsive		Responsive		Responsive ( Responsive for 4a,4d & 4e)
			Nil Kamal	Non- Responsive	BG for short period	Responsive		Non-Responsive

	5	Vaccine Van (2WD)	Tata Motors	Responsive		Non- Responsive	Responsive: Weight carrying capacity 495 kg. Against the tender requirement of 1000kg excluding the weight of cargobox.	Non-Responsive
			Force Motors	Responsive		Responsive		Responsive
<b>Tender Ref No</b>	<b>Item No</b>	<b>Name of Equipment</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason For Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
HLL/PCD/IMMU-04/11-12	1	Deep Freezer (Large)	Voltas Ltd.	Responsive		Non- Responsive	No CE /FDA/BIS certification.	Non- Responsive
			Ratan Sales Corporation	Non Responsive	Not meeting the qualification requirement. End user certificate not submitted	Non- Responsive	No test certification according to the required protocol.	Non- Responsive
			Pawar Electro System Pvt. Ltd.	Non Responsive	Not meeting the qualification requirement. End user certificate not submitted	Non- Responsive	No test certification according to the required protocol.	Non- Responsive
			Prama Instrument Pvt. Ltd.	Non- Responsive	Not meeting the qualification requirement of authorized agent.	Non- Responsive	Test certificated for hold over time as asked in the specifications not submitted.	Non- Responsive
			Khera Instrument Pvt. Ltd.	Non- Responsive	End user certificate not submitted	Non- Responsive	No test certification according to the required protocol.	Non- Responsive
	2	Stem Thermometer	No Bid Received					
	3	Ice Packs	Apex (Guj.) Plastochem Pvt. Ltd.	Responsive		Non- Responsive	Compliance statement not given. Not clarified.	Non- Responsive
			Xinxiang Dengke Cold Chain Equipment Ltd.	Non- Responsive	Clarification asked for certificates from end users- Not submitted by the supplier till 15.02.2013. Mail received from bidder that it may take another 10 days.	Responsive		Non- Responsive
			Nil Kamal Ltd.	Non- Responsive	BG Short	Non- Responsive	Not meeting dimension of ice pack	Non- Responsive
			Blow King	Responsive		Responsive		Responsive
			AOV International	Responsive		Non- Responsive	Not meeting dimension of ice pack	Non- Responsive
			Apex International	Non Responsive	End user certificate not submitted	Non- Responsive	Not meeting dimension of ice pack	Non- Responsive

	4	Cold Box (small)	Sintex Industries Ltd.	Responsive		Responsive		Responsive	
			Promens India	Responsive		Responsive		Responsive	
			Nil Kamal	Non- Responsive	BG Short	Responsive		Non- Responsive	
			AOV International	Responsive		Responsive		Responsive	
	5	Cold Box (Large)	Sintex Industries Ltd.	Responsive		Responsive		Responsive	
			Promens India	Responsive		Responsive		Responsive	
			Nil Kamal	Non- Responsive	BG Short	Responsive		Non- Responsive	
			AOV International	Responsive		Responsive		Responsive	
			Apex International	Non- Responsive	End user certificate not as per the required quantity	Non- Responsive		Non- Responsive	
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>	
HLL/PCD/MCD-01/13-14/MT	1	Malathion Technical	No bids Received						
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>	
HLL/PCD/MCD-02/13-14/TG	1	Temephos Granules	Chemet Chemicals P. Ltd.	Responsive		Responsive		Responsive	
			Kilpest India Ltd.	Regret letter submitted					
			Heranba Industries Ltd.	Responsive		Responsive		Responsive	
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>	
HLL/PCD/MCD-03/13-14/SP	1	Alphacypermethrin 5% (synthetic pyrethroid)	Chemet wets & Flows Pvt.Ltd.	Responsive		Responsive		Responsive	
			Kilpest India Ltd.	Responsive		Responsive		Responsive	
			Heranba Industries Limited	Responsive		Responsive		Responsive	

			Tagros chemicals India Ltd.	Responsive		Responsive		Responsive
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
HLL/PCD/MCD-04/13-14/TE	1	Temephos 50% EC	Coromandal International Limited	Responsive		Responsive		Responsive
			Chemet wets & Flows Pvt.Ltd	Responsive		Responsive		Responsive
			Heranba Industries Limited	Responsive		Non-Responsive	The submitted ISI certificate is valid upto 23.05.13 which is before due date of tender opening	Non-Responsive
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
HLL/PCD/MCD-05/13-14/PE	1	Pyrethrum Extract 2%	Bombay Chemicals Private Ltd.	Responsive		Responsive		Responsive
<b>Tender Ref No</b>	<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
HLL/PCD/MCD-06-RT/13-14/MT	1	Malathion Technical	Coromandel International Ltd.	Responsive		Responsive		Responsive