

**Amendment No. 4**

Date: 12.05.2015

**Subject: Amendment no. 04 to the Tender Enquiry Document****Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY-II/NAGPUR/04/14-15 dated 06.02.2015 and subsequent amendments published thereafter.****Due to certain technical reasons in the e-tendering portal, the bidders are requested to submit a hardcopy of their technical bids and all related documents (uploaded in the portal) EXCEPT for the Price bid before the mentioned "Closing date & time for physical submission of bids".**

The due date for submission of bids is revised as follows:

**Section I**  
**Notice Inviting Tenders(NIT)****(1) For:-**

Sl.	Description	Schedule
c	Closing date & time for submission of <b>tender fee and EMD in physical form</b>	13-May-2015, 1700 hrs (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	15-May-2015, 1300 hrs IST
e	Time and date of opening of online bids	15-May-2015, 1400 hrs IST

**Read as:-**

Sl.	Description	Schedule
c	Closing date & time for submission of <b>tender fee and EMD in physical form</b>	<b>28-May-2015, 1400 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	i) Closing date & time for submission of online bids	<b>29-May-2015, 1300 hrs IST</b>
	ii) Closing date & time for physical submission of bids (Price bid shall only be uploaded online; All other technical and commercial documents uploaded in the e-portal are to be submitted manually)	<b>30-May-2015, 1300 hrs IST</b>
e	Time and date of opening of bids	<b>30-May-2015, 1400 hrs IST</b>

**Note:****(i) If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e., upto 11.11.2015.**

**(ii) Tender fee (Rs.50,000/-) and EMD ( As applicable) should be deposited in the Tender Box located at HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh on or before 28-May-2015, 1400 hrs (IST) . Submission beyond stipulated date & time would result in REJECTION of BID.**

**Section VII**  
**Technical Specifications**

**Schedule No. 1**  
**Modular Operation Theatre**

**1. Existing Specification**

**Para 6a.16:** The complete door assembly should be CE marked.

**Read as**

**Para 6a.16 - The door controller should be CE marked.**

**Schedule No. 2**  
**INTEGRATION AND DATA MANAGEMENT SYSTEM FOR MODULAR OT**

**1. Existing Specification**

**Para 1.3:** The monitors should be mounted on a boom arm and separate ceiling suspended boom arm for each monitor should supplied Patch panel for power & signal to be laid down for 32" LCD Monitor at Wall of MOT.

**Read as**

**Para 1.3:** The monitor 1 should be on ceiling suspended boom arm and monitor 2 on top shelf of surgical pendant (Incase any attachment will be needed to mount the monitor2, the same is to be supplied by the integration vendor). Should be supplied Patch panel for power & signal to be laid down for 32" LCD Monitor at Wall of MOT.

**2. Existing Specification**

**Para 2.4:** Two loudspeakers shall be installed within the Operating room. **3-Channel Loudspeaker** with Digital volume control and Audio mixer and Audio equalizer should be installed at a most suitable place. Suitable cable material and a patch panel should be offered as per the position of the Loudspeaker

**Read as**

**Para 2.4:** Two loudspeakers shall be installed within the Operating room. **Loudspeaker** with Digital volume control and Audio mixer and Audio equalizer should be installed at a most suitable place. Suitable cable material and a patch panel should be offered as per the position of the Loudspeaker

**3. Existing Specification**

**Para 3.5:** It should have atleast 1 TB internal Hard Disk Drive (HDD) for in-system archiving. Also, it should have a feature of real time in-procedure DVD burning besides at-the end procedure DVD burning.

**Read as**

**Para 3.5** - It should have **(Internal + External) storage capacity of 1 TB or more** for in-system archiving. Also, it should have a feature of real time in-procedure DVD burning besides at-the end procedure DVD burning

#### 4. Existing Specification

**Para 4.1:** Should have individual **high definition** medical grade 32" LED monitor, wall mounted (to display images from PACS)

**Read as**

**Para 4.1:** Should have individual **full high definition** medical grade 32" LED monitor, wall mounted (to display images from PACS)

#### 5. Existing: BOQ for Integration and Data Management System

##### BOQ Para 2

S.No	NAME OF THE ITEMS	UNIT	Qty
2	Ceiling arm to mount 32" monitor as per tender specifications	Nos	6

**Read as**

##### BOQ Para 2

S.No	NAME OF THE ITEMS	UNIT	Qty
2	Ceiling arm to mount 32" monitor as per tender specifications	Nos	3

**All other terms and conditions of the tender enquiry remain unaltered.**