

## **Clarification**

Date: 19/04/2016

**Sub: Clarification to the Tender Enquiry Document**

**Ref: NIT No.: HLL/PCD/PMSSY/AIIMS-II/15-RT-01/15-16 dated 31.12.2015**

### **Section – VII** **Technical Specifications**

#### **Schedule no. 2** **Integration and Data Management System for Modular OT**

##### **1. Para 1. DIGITAL DISPLAY MONITOR**

###### **Added Para under Para1:**

26" Full High Definition(1920X1080p) medical grade monitor, flat panel LED Colour screen to display both 2D & 3D images. **(Optional – Price to be quoted separately) – This is in addition to the monitor asked as standard.**

###### **2. Added Para:**

All the necessary equipment required in the Auditorium/Conference Room (like room camera, Video Projector, PA system etc) for Video Conferencing should be quoted. **(Optional – Price to be quoted separately)**

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

###### **Note:**

- (i) There will not be any extension of due dates for submission of bids(Due date & time for submission of tender in e-Portal is 27.04.2016, 1800 hrs)**
- (ii) Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**