

AMENDMENT NO.3 Dated 20.03.2020

Ref No: HLL/SD/RBD/2019-20/TENDER/15 Dt: 05.03.2020

Title: Supply of Personal Protective Equipments - Coverall (Garments), Goggles, N95 Face Mask with Expiratory Valve, Nitrile Gloves, Face shield and Triple layer surgical mask.

The following amendment has been incorporated to the bid document for the above tender;

1. Following products has been included;

**Section 1 - NOTICE INVITING TENDER, Point no II: Product requirement with Quantity
Annexure -2, PRICE SCHEDULE
Annexure -7, TECHNICAL COMPLIANCE SHEET**

Sr. No.	Item	Total tentative Quantity required	UoM
19	Digital Infrared Thermometer (any make)	10,000	Nos
20	Digital Thermal Scanner - Full Body (with sensing to 3m and beyond)	50	Units
21	Examination Glove (Medium and large Size)	70 Lakh	Pairs
22	Spray Machine	150	Nos
23	Hair Cover (Surgeon Cap)	2000	Nos
24	Medical mask - Cloth Reusable	1000	Nos
25	Paper Towel	100	Nos

2. Annexure 5 - Specification of the following products added;

19. Digital Infrared Thermometer

- Measures the body temperature by determining the infrared reflected from the front radiation (no contact).
- Measuring range of the infrared temperature (body) from 32°C to 42.5°C
- Resolution 0.1°C
- High accuracy +/-0.2°C
- Obtaining quick results: Time response 1 second
- Indication of measured value in °C or °F
- Limit value for the adjustable alarm as programmable
- Distance of measuring: at least 15 cm
- Memory to store (expandable memory desirable)
- Easy to use, robust
- Automatic switch off
- Display backlight glow
- Waterproof lens easy to clean and disinfect
- Batteries: Which are replaceable/chargeable
- High degree of health and safety
- Warranty: At least 3 years
- Accessories:
 - I. Carrying case
 - II. Data transfer software
 - III. User manual

20. Digital Thermal Scanner - Full Body

- Compact and light weight design
- Tripod mountable
- Able to measure temperature from a distance of at least 3-5 mtrs
- Process single and multiple hotspots tracing and alarm
- Able to capture both thermal and digital camera image
- Will be able to measure temperature from 32⁰C to 42.5⁰C with a sensitivity of at least 0.08⁰C
- Images should have high resolution and high sensitivity
- Can be connected to a external monitor through video output
- Built in colour screen for thermal images 5-8 inch
- Software compatible with windows. Provide the outputs of thermal data and temperature data
- Operating temperature -10 +50⁰C
- Memory for data storage (expandable)
- Battery operated and works on external power supply
- Provide all configurations as per standard required
- Onsite installation and hands on training
- Warranty for 3 years and assured comprehensive onsite service for atleast 5 years
- Annual maintenance support for minimum 5 years.

21. Examination Gloves:

- **Size: Medium and Large**
- **Powdered**
- **Latex - ASTM D-3578**

3. Last date of submission of bids

FOR

Last date for submission of bidding Document: 20.03.2020, 11.30 hrs

MAY BE READ AS

Last date for submission of bidding Document: 25.03.2020, 14.30 hrs

FOR

Last date for submission of bidding Document: 20.03.2020, 14.30 hrs

MAY BE READ AS

Last date for submission of bidding Document: 25.03.2020, 15.30 hrs

Date of Bid opening

FOR

Time and date of opening of bids: 20.03.2020, 15.30 hrs

MAY BE READ AS

Time and date of opening of bids: 25.03.2020, 15.30 hrs

All relevant clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Senior Manager
Sourcing Division - RBD
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.
Ph.no: 0471 2353932.