

## **Clarification Regarding Technical Specifications (Part III)**

**Date: 17/01/2014**

**Subject: Clarification regarding Technical Specifications**

- Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/02/13-14 dated 02/12/2013**  
**(ii) Amendment No.1 dated 20/12/2013.**  
**(iii) Amendment No.2 dated 10/01/2014.**  
**(iv) Clarification regarding technical specifications published on 11/01/2014.**  
**(v) Clarification regarding technical specifications (Part II) published on 15/01/2014.**

**The following clarifications regarding technical specifications are made further to those referred above.**

### **Section – VII** **Technical Specifications**

#### **Schedule no. 2** **Technical specification of a new, state of the art 3T MRI scanner**

**Existing Specification:**

Para 8: Workstation

Two workstations (thin client server architecture based) with post processing tools as specified below. The server (single/dual configuration) should have image storage capacity of 3 Tera bytes, **minimum 4,000 concurrent slice processing power** and at least 64GB RAM. Necessary Workstation hardware to be included with 18" or more TFT/LCD Medical Grade (2 megapixel or more resolution) monitor with dual processor, separate hard drive with image storage of at least 2.5 lakh images (256 x 256 matrix) with CD RW or DVD RW. DICOM 3.0 compatibility and interfacing with other modalities must be possible.

**Read as:**

Para 8: Workstation

Two workstations (thin client server architecture based) with post processing tools as specified below. The server (single/dual configuration) should have image storage capacity of 3 Tera bytes, **minimum 40,000 concurrent slice processing power** and at least 64GB RAM. Necessary Workstation hardware to be included with 18" or more TFT/LCD Medical Grade (2 megapixel or more resolution) monitor with dual processor, separate hard drive with image storage of at least 2.5 lakh images (256 x 256 matrix) with CD RW or DVD RW. DICOM 3.0 compatibility and interfacing with other modalities must be possible.

**All other terms and conditions of the tender enquiry remain unaltered.**