

## **Clarification**

**Date: 25/02/2016**

**Sub: Amendment to the Tender Enquiry Document**

**Ref: (i) NIT No.: HLL/PCD/PMSSY/AIIMS-II/25/15-16 dated 14/10/2015  
(ii) Amendment No. 07 dated 19-02-2016  
(iii) Amendment No. 08 dated 22-02-2016**

The following clarifications are being issued against the above referred tender.

### **Item No. 56 ICU ventilator**

#### **1. Existing Specification:**

Para: 21. Expiratory valve/cassette should be autoclavable and supply 2 nos with each unit.

#### **Read as:**

Para: 21. Expiratory valve/cassette/**block** should be autoclavable and supply 02 nos with each unit.

#### **2. Existing Specification: (as per in amendment no. 07 added para (point-9):**

iii. One no. of expiratory valve/block should be provided with each ventilator

#### **Read as:**

**Deleted**

### **Item No. 57 View Box**

#### **1. Existing Specification:**

Para: 1. Specialty X- Ray Film LED based View Box for viewing CT/ MRI/ Diagnostic X-ray films **with External Electrodes Fluorescent Lamp Technology.**

#### **Read as:**

Para: 1. Specialty X- Ray Film LED based View Box for viewing CT/ MRI/ Diagnostic X-ray films.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**