

07-09-2019

Amendment No. 4**Sub: Amendment to the referred tender enquiry**

Ref.: Tender Enquiry No.HLL/PCD/IMMU-14/2019-20 dated 29/06/2019; Clarification 1 dated 09-07-2019, Amendment No. 1 dated 25-07-2019, Amendment No. 2 dated 14-08-2019 and Amendment No. 3 dated 29-08-2019

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION – IV
GENERAL CONDITIONS OF CONTRACT (GCC)

15. Warranty**For:**

15.7 During Warranty period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the installation for preventive maintenance of the goods.

Read As:

15.7 During Warranty period, the supplier is required to visit at each consignee's site at least once in 12 months commencing from the date of the installation for preventive maintenance of the goods.

Section VII
Technical Specification

Schedule No.1 to 4 - Ice Line Refrigerator- ILR (Large)			
Sl. No	Tender Page No. & Para	TENDER SPECIFICATION	READ AS (AMENDED SPECIFICATION)
1	Para 2.6 page 57	Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.	The unit should have ground clearance from 70 mm to 100 mm.
2	Para 2.7 page 57	The unit should have ground clearance of minimum 100 mm.	
3	Para 7.2 Page 58	Voltage stabilizer as per the MOHFW approved specifications and standard enclosed as Annexure-2 (applicable for external stabilizer)	Either external voltage stabilizer as per specifications and standard enclosed as Annexure-2 or internal stabilizer as per rating of the equipment.

Schedule No. 5 to 8 - Ice Line Refrigerator- ILR (small)			
Sl. No	Tender Page No. & Para	TENDER SPECIFICATION	READ AS (AMENDED SPECIFICATION)
1	Para 2.6 page 60	Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.	The unit should have ground clearance from 70 mm to 100 mm.
2	Para 2.7 page 60	The unit should have ground clearance of minimum 100 mm.	
3	Para 7.2 Page 61	Voltage stabilizer as per the MOHFW approved specifications and standard enclosed as Annexure-2(applicable for external stabilizer)	Either external voltage stabilizer as per specifications and standard enclosed as Annexure-2 or internal stabilizer as per rating of the equipment.

Schedule No. 9 to 12- Deep freezer (Large)			
Sl. No	Tender Page No. & Para	TENDER SPECIFICATION	READ AS (AMENDED SPECIFICATION)
1	Para 2.6 page 63	Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.	The unit should have ground clearance from 70 mm to 100 mm.
2	Para 2.7 page 63	The unit should have ground clearance of minimum 100 mm.	
3	Para 7.2 Page 64	Voltage stabilizer as per the MOHFW approved specifications and standard enclosed as Annexure-2(applicable for external stabilizer)	Either external voltage stabilizer as per specifications and standard enclosed as Annexure-2 or internal stabilizer as per rating of the equipment.

Schedule No. 13 to 16 - Deep freezer (small)			
Sl. No	Tender Page No. & Para	TENDER SPECIFICATION	READ AS (AMENDED SPECIFICATION)
1	Para 2.5 page 65	Should have legs in the base with rotating screw type height adjustments to balance the weight on uneven floor.	The unit should have ground clearance from 70 mm to 100 mm.
2	Para 2.6 page 65	The unit should have ground clearance of minimum 100 mm.	
3	Para 7.2 Page 67	Voltage stabilizer as per the MOHFW approved specifications and standard enclosed as Annexure-2(applicable for external stabilizer)	Either external voltage stabilizer as per specifications and standard enclosed as Annexure-2 or internal stabilizer as per rating of the equipment.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.