

## **Amendment No. 2**

**Date: 28.04.2015**

**Subject: Amendment no. 2 to the Tender Enquiry Document**

**Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/12/RT-01/14-15 dated 16.03.2015 and subsequent amendments published thereafter.**

The pre-bid meeting for the referred tender enquiry was held on 08/04/2015. Based on pre-bid discussions following amendments are being incorporated in the referred tender enquiry document.

### **Section VII** **Technical Specifications**

#### **Schedule 01** **ETO Sterilizer**

**1. Existing Specification:**

Para: 9.b Packaging material (1 roll of each size) with chemical & biological indicator should be supplier

**Read as:**

Para: 9.b Packaging material (1 roll each of different standard sizes) may be quoted and 50 units of chemical & biological indicator should be supplied.

**2. Added Para:**

**If any machine required compressed air plant for running the machine it should be supplied as standard.**

**3. Added Para:**

**All listed accessories and parts (eg.Vent hoods and exhaust hoods) as per manufacturer's guidelines required for running ETO should be supplied as standard.**

**4. Existing Specification:**

Para: 12 - The ETO sterilizer should have compliance to **BS EN ISO 9001:2000**. Certificate to be provided by equipment manufacturer.

**Read as:**

Para: 12 - The ETO sterilizer should have compliance to **BS EN ISO 9001**. Certificate to be provided by equipment manufacturer.

**5. Existing Specification:**

Para: 13- The ETO sterilizer should have compliance to **ISO13485:1996**. Certificate to be provided by equipment manufacturer.

**Read as:**

Para: 13- The ETO sterilizer should have compliance to **ISO 13485**. Certificate to be provided by equipment manufacturer.

**All other terms and conditions of the tender enquiry remain unaltered.**