

Tender No. : HLL/CRD/PUR/TENDER/2014-15/02

Amendment No .02

05-11-2014

SUB: Amendment in Supply installation and commissioning of various equipments.

TENDER NO. HLL/CRD/PUR/TENDER/2014-15/02

With reference to the above we hereby incorporate the following amendment to the tender documents as detailed below.

Last date and time for receipt of bid: 13/11/2014 4:30 PM

Date and time of opening of Technical Bid: 14/11/2014 10:30 AM

Specification of equipment, Item No E 13 is amended as follows

Sl. No.	Tender No.	Name of the Equipment	URS (User Requirement Specifications)	Qty Required	EMD amount (Rs.)	EMD amount (USD)
13	HLL/CRD/PUR/TENDER/2014-15/02- E13	Antibody dispensing machine	<ul style="list-style-type: none">No. of Reagents/pumps: dispense 4 different reagents simultaneously with 1 Air jet pumpDispensing rate : 0.3 to 10 μL/cmCarriage minimum Speed : 10 to 100mm/sMinimum Membrane Length : 300mmMinimum Membrane width : 125mmMinimum spacing between the nozzles: \geq 5mmPower Requirement : Approx. 100-240VAC, 50/60Hz.Operation: Completely digital (Touch screen is preferable)Should able to dispense the gold conjugate by spraying method (Air Jet pump)Should provide with one extra set of Nozzle, plunger, syringe, Holder & Power cable etc.Machine should be table topShould able to retrieve/recollect the reagents from the syringe and tube.Should have the calibration and volume adjustment functionsMachine should be durable and required less maintenanceShould have easy programming for speed, volume, spacing and other functionsSystem requires vacuum pump, must provide with machine.Should include installation and training	01	30,000	492.00

Tender No. : HLL/CRD/PUR/TENDER/2014-15/02