

Amendment No. 2

06.09.2018

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/22/18-19 dated 26.07.2018 read with its Amendment No. 1 dated 28.08.18

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION - VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Rfx/ Event number 3000003176)****Micro Vascular Instruments Set**

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 46, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Deleted
2	page 46, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
3	page 46, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Dedicated Slotted sterilization container with lock and silicon mat
4	page 46, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ S.No 3, 6, 11, 13, 15 and 19 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page,46 BOQ, point 4	Vein hook	T Vein hook:- 11-13 cm
6	Page,47 BOQ, point 13	Debakey vascular scissors	Debakey vascular scissors 25deg, 15-17cm & 45 deg, 15-17cm: 1 each

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
7	Page,47 BOQ, point 17	Satinsky vascular clamps medium	Satinsky vascular clamps medium : 24-26cm
8	Page,47 BOQ, point 18	Satinsky vascular clamps small	Satinsky vascular clamps small :- 18-22cm
9	Page,47 BOQ, point 19	Debakey's vascular clamps medium	Debakey's vascular clamps medium 16-18cm
10	Page,47 BOQ, point 20	Debakey vascular clamp small	Debakey vascular clamp small 10-12cm
11	Page,47 BOQ, point 21	Bull dog clamps 5.5 cm	Bull dog clamps 5.5-7.5cm

Item No. 2 (Rfx/ Event number 3000003177)

Major Vascular Instruments Set

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	Page 48, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Deleted
2	Page 48, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
3	page 48, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Deleted
4	Page 48, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ S..No 2,5,9,15,21,40,41,52,58,66 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 48, BOQ point 1	Adson non-toothed micro dressing forceps	Adson non-toothed micro dressing forceps: 12-14 cm
6	Page 48, BOQ point 3	Artery forceps spencer wells	Artery forceps spencer wells 14-16 cm
7	Page 48, BOQ point 5	Artery forceps-Halsted(Adson) curved	Artery forceps-Halsted(Adson) curved 17-19cm
8	Page 48, BOQ point 6	Artery forceps-Halsted(Adson) straight	Artery forceps-Halsted(Adson) straight 17-19cm

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
9	Page 48, BOQ point 7	Baby satinsky	Baby satinsky 14-18cm
10	Page 49, BOQ point 9	Bozmann Needle holder serrated	Bozmann Needle holder serrated 20-22cm
11	Page 49, BOQ point 10	Cooley atraumatic forceps	Cooley atraumatic forceps 12-14cm
12	Page 49, BOQ point 11	Cooley vascular clamps	Cooley vascular clamps 16-18cm
13	Page 49, BOQ point 12	Crile wood Needle holder serrated	Crile wood Needle holder serrated 14-16cm
14	Page 49, BOQ point 13	Cushing dressing forceps (curved)	Cushing dressing forceps (curved) 17-19cm
15	Page 49, BOQ point 14	Cushing nerve hook	Cushing nerve hook 18-20cm
16	Page 49, BOQ point 15	Debakey atraumatic vascular clamp	Debakey atraumatic vascular clamp 15-17cm
17	Page 49, BOQ point 16	Debakey multipurpose vascular clamps	Debakey multipurpose vascular clamps 14-16cm
18	Page 49, BOQ point 17	Debakey vascular scissors	Debakey vascular scissors 25 deg & 45 deg & length 15-17cm – 1 each
19	Page 49, BOQ point 23	Debaky atraumatic forceps (angled)	Debaky atraumatic forceps (angled) 23-25cm
20	Page 49, BOQ point 31	Debaky Needle holder serrated	Debaky Needle holder serrated 17-19cm
21	Page 49, BOQ point 32	Debakysatinsky	Debaky satinsky 14-18cm
22	Page 49, BOQ point 33	Debaky vascular scissors angled	Debaky vascular scissors angled 45 deg 15-17cm
23	Page 49, BOQ point 34	Golden Handle small needle holder	Golden Handle small needle holder 6-8"
24	Page 49, BOQ point 35	Halsted mosquito hemostatic forceps	Halsted mosquito hemostatic forceps 12cm
25	Page 49, BOQ point 37	Kelley clamp long curved	Kelley clamp long curved 14-16cm
26	Page 49, BOQ point 38	Lahey gall duct forceps	Lahey gall duct forceps 18-20cm
27	Page 49, BOQ point 41	metzenbaum dissecting scissors	Metzenbaum dissecting scissors 15-17cm
28	Page 49, BOQ point 42	metzenbaum dissecting scissors curved	Metzenbaum dissecting scissors curved 14-16cm
29	Page 50, BOQ point 46	Mosquito Artery forceps(curved)	Mosquito Artery forceps(curved) 12-14cm
30	Page 50, BOQ point 47	Mosquito Artery forceps(straight)	Mosquito Artery forceps(straight) 12-14cm
31	Page 50, BOQ point 52	Pott's Smith dissecting forceps	Pott's Smith dissecting forceps 17-19cm

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
32	Page 50, BOQ point 53	Reynolds scissors fine curved	Reynolds scissors fine curved 17-19cm
33	Page 50, BOQ point 57	S.S Tray (Big)	Deleted
34	Page 50, BOQ point 58	Satinsky blood vessel clamp	Satinsky blood vessel clamp medium 22-24cm
35	Page 50, BOQ point 59	Satinsky blood vessel clamp (Big)	Satinsky blood vessel clamp (Big) 26-28cm
36	Page 50, BOQ point 60	Satinsky blood vessel clamp (Small)	Satinsky blood vessel clamp (Small) 20-24cm
37	Page 50, BOQ point 64	Vein Hook	Vein Hook 12-14cm
38	Page 50, BOQ point 68	Eyelid retractor	Eyelid retractor 12-14cm

Item No. 3 (Rfx/ Event number 3000003178)

Thoracic vascular Instruments Set

Sl. No.	Ref. to the Bidding Document	Existing specification/clause in the Bidding Document	Amended as
1	Page 51, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Deleted
2	Page 51, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	All needle holders & Metzenbaum scissors should be Tungsten carbide insert
3	page 48, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Deleted
4	Page 51, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ S. No 4,11,17,21,23,29 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval
5	Page 51, BOQ point 1	S.S Tray Large	Deleted
6	Page 51, BOQ point 9	Satinsky blood vessel clamp big	Satinsky blood vessel clamp big 24-27

7	Page 52, BOQ point 14	Adson hemostat forceps straight	Adson hemostat forceps straight 14-16 cm
8	Page 52, BOQ point 15	Adson hemostat forceps curved	Adson hemostat forceps curved 14-16 cm
9	Page 52, BOQ point 29	Pott's smith dressing forceps TC	Pott's smith dressing forceps TC 17-19 cm
10	Page 52, BOQ point 32	Cushing nerve hook medium	Cushing nerve hook medium 18-20 cm

Item No. 4 (Rfx/ Event number 3000003179)

Plastic Surgical Instruments Set

Sl. No.	Ref. to the Bidding Document	Existing specification/ clause in the Bidding Document	Amended as
1	Page 53, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property	Deleted
2	Page 53, point 4	All needle holders & Metzenbaum scissors should be Tungsten Carbide coated	All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert
3	page 53, point 6	Instruments should be supplied with autoclave compatible container and preferably of same make of Instruments	Deleted
4	Page 53, point 8	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ Sl. No 4, 9,12,16,19 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 53, BOQ point 5	Non toothed dissecting forceps golden TC tip	Non toothed dissecting forceps golden TC tip 17-19 cm
6	Page 53, BOQ point 8	Toothed dissecting forceps	Toothed dissecting forceps 17-19 cm
7	Page 54, BOQ point 12	Iris scissor angled golden	Iris scissor angled golden 10-12 cm
8	Page 54, BOQ point 14	Castro-veijo needle holder golden	Castro-veijo needle holder golden 14-16 cm

9	Page 54, BOQ point 15	Needle holder golden	Needle holder golden medium
10	Page 54, BOQ point 16	Needle holder small	Needle holder small 6-8"
11	Page 54, BOQ point 19	Muller retractor	Muller retractor 15-17 cm
12	Page 54, BOQ point 20	Adson retractor	Adson retractor 23-25 cm
13	Page 54, BOQ point 21	S.S Tray (medium)	Deleted

All other contents of the Bidding Document including terms & conditions remain unaltered.