

Amendment No. 2

06.09.2018

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/23/18-19 dated 26.07.2018 read with its Amendment No. 1 dated 28.08.18

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION - VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Rfx/ Event number 3000003180)****Abdominal Instruments Set**

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | Page 46, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 46, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 46, point 4 | All needle holders & Metzenbaum scissors should be Tungston Carbide | All needle holders & Metzenbaum scissors should be Tungston Carbide Insert |
| 4 | Page 46, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ Sl. No 11, 34a, 34f, 34m, 34o are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 46, BOQ point 1 | SS tray large | Deleted |
| 6 | Page 46, BOQ point 2 | Towel Clips Large | Deleted |
| 7 | Page 46, BOQ point 3 | Towel Clips medium | Deleted |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 8 | Page 46, BOQ point 4 | Towel Clips small | Deleted |
| 9 | Page 47, BOQ point 20 | SS bowl large | Deleted |
| 10 | Page 47, BOQ point 21 | SS bowl medium | Deleted |
| 11 | Page 47, BOQ point 22 | Sponge holder | Deleted |
| 12 | Page 47, BOQ point 23 | Suction tip | Suction tip 18-20 cm |
| 13 | Page 47, BOQ point 24 | Self retaining retractor with 3rd blade long | Gosset or Similar type three blade self retaining abdominal wall retractor 17-20 cm |
| 14 | Page 47, BOQ point 25 | Self retaining retractor with 3rd blade small | Gosset or Similar type three blade self retaining abdominal wall retractor 12-15 cm |
| 15 | Page 47, BOQ point 29 | Deaver's retractor large | Deaver's retractor 33-36 cm |
| 16 | Page 47, BOQ point 30 | Deaver's retractor medium | Deaver's retractor 24-26 cm |
| 17 | Page 47, BOQ point 34 | In Kidney Tray (Big) | Deleted |
| 18 | Page 47, BOQ point 34.d | Debaky's vascular forceps | Debaky's vascular forceps 16-18cm |

Item No. 2 (Rfx/ Event number 3000003181)

Thoracotomy Instruments Set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| 1 | Page 49, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 49, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 49, point 4 | 4. All needle holders & Metzenbaum scissors should be Tungston Carbide | All needle holders & Metzenbaum scissors should be Tungston Carbide Insert |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 4 | Page 49, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ S.No 10,13,23,31,34,37,44,50 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval |
| 5 | Page 49, BOQ point 1 | SS tray large | Deleted |
| 6 | Page 49, BOQ point 2 | Towel clips large | Deleted |
| 7 | Page 49, BOQ point 3 | Towel clips medium | Deleted |
| 8 | Page 49, BOQ point 4 | Towel clips small | Deleted |
| 9 | Page 49, BOQ point 16 | SS bowl large | Deleted |
| 10 | Page 50, BOQ point.17 | SS bowl medium | Deleted |
| 11 | Page 50, BOQ point.18 | Sponge holder | Deleted |
| 12 | Page 50, BOQ point.19 | Suction tip | Suction tip 18-20cm |
| 13 | Page 50, BOQ point.22 | Chest wall retractor with key large blade | Caspar or Similar large blade 15-17cm chest wall retractor with key |
| 14 | Page 50, BOQ point.23 | Chest wall retractor with key normal blade | Caspar or Similar medium size blade 11-13cm chest wall retractor with key |
| 15 | Page 50, BOQ point.24 | Chest wall retractor with key double hook | Caspar or Similar Chest wall retractor with key double hook with 11-13cm blade |
| 16 | Page 50, BOQ point.25 | Rib shear | Rib shear 24-26cm |
| 17 | Page 50, BOQ point.28 | Rib raspatory left | Rib raspatory left 20-22cm |
| 18 | Page 50, BOQ point.29 | Rib raspatory right | Rib raspatory right 20-22cm |
| 19 | Page 50, BOQ point.30 | Periosteum elevator | Periosteum elevator 16-18cm |
| 20 | Page 50, BOQ point.31 | Lung holding forceps | Lung holding forceps 19-21cm |
| 21 | Page 50, BOQ point.32 | Semb dissecting forceps | Semb dissecting forceps 24-26cm |
| 22 | Page 50, BOQ point.33 | Nelson long dissecting forceps | Nelson lung grasping forceps 20-22 cm |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|---------|------------------------------|--|--|
| 23 | Page 50, BOQ point.35 | Debakey's atraumatic non toothed dissecting forceps | Debakey's atraumatic non toothed dissecting forceps 17-19cm |
| 24 | Page 50, BOQ point.36 | Davidson scapular retractor | Davidson scapular retractor 15-17cm |
| 25 | Page 50, BOQ point 42 | Bozemom needle holder | Bozemom needle holder 19-21cm |
| 26 | Page 51, BOQ point 53 | Sternal needle holder | Sternal needle holder 14-16cm |
| 27 | Page 51, BOQ point 54 | Pliar | Pliar 14-16cm |
| 28 | Page 51, BOQ point 55 | Wire cutter | Wire cutter 17-19cm |

Item No. 3 (Rfx/ Event number 3000003182)

Large Bone Set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|---------|------------------------------|---|--|
| 1 | page 52 , point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 52, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 52, point 4 | All needle holders & Metzenbaum scissors should be Tungsten Carbide coated | All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 52, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ Sl. No 9,12,19,22,32 & 34 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 52 , BOQ point 1 | SS tray big | Deleted |
| 6 | Page 52 , BOQ point 2 | Amutation knife medium | Amputation knife liston/catlin 22-24cm |
| 7 | Page 52 , BOQ point 3 | Amutation knife small | Amputation knife liston/catlin/langenbeck 13-16cm |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| 8 | Point 52 BOQ point 4 | Bone cutter angled | Bone cutter angled 14-16cm |
| 9 | Point 52 BOQ point 5 | Bone cutter double action | Bone cutter double action 25-27cm |
| 10 | Point 52 BOQ point 6 | Bone file big | Bone file 22-24cm |
| 11 | Point 52 BOQ point 7 | Bone file small | Bone file 16-18cm |
| 12 | Point 53 BOQ point 13 | Bone lever lancer type plain | Bone lever lancer type plain 15-17cm |
| 13 | Point 53 BOQ point 14 | Bone lever lancer type serrated | Bone lever lancer type serrated 15-17cm |
| 14 | Point 53 BOQ point 15 | Bone nibbler double action curved | Bone nibbler double action curved 22-24cm |
| 15 | Point 53 BOQ point 16 | Bone nibbler double action straight | Bone nibbler double action straight 22-24cm |
| 16 | Point 53 BOQ point 17 | Boyle mouth gag with 5 plates | MOLT or Similar Mouth gas 10-12 cm & 14-16 cm - 1 each size |
| 17 | Point 53 BOQ point 21 | Curette medium | Curette medium 12-15 cm |
| 18 | Point 53 BOQ point 22 | Curette with handle big | Curette with handle big 17-20 |
| 19 | Point 53 BOQ point 24 | Hammer/Mallet medium | Hammer/Mallet small 22-24 cm length, head dia. 25-27 on one side & 13-15 on other side |
| 20 | Point 53 BOQ point 25 | Hammer/Mallet large | Hammer/Mallet small 22-24 cm length, head dia 38-40 on one side & 28-30 cm on other side |
| 21 | Point 53 BOQ point 26 | Hand saw medium | Charriere or Similar hand saw length 24-27cm |
| 22 | Point 53 BOQ point 27 | Hand saw large | Satterlee or Similar hand saw length 30-33 cm |
| 23 | Point 53 BOQ point 28 | HM wire cutter | HM wire cutter 16-18 cm |
| 24 | Point 53 BOQ point 29 | Mouth gag (15 cm) | MOLT or Similar Mouth gas 10-14-16 cm - 1 no. |
| 25 | Point 53 BOQ point 34 | Periosteum elevator | Periosteum elevator 16-18cm |

Item No. 4 (Rfx/ Event number 3000003183)**Head & Neck and Orbitotomy Bone Set**

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | page 54, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 54, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 54, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 54, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ S.No 8,11,15,19,29,60 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 54 , BOQ point 1 | SS tray big | Deleted |
| 6 | Page 54 , BOQ point 2 | Amutation knife medium | Amputation knife liston/catlin 22-24cm |
| 7 | Page 54 , BOQ point 3 | Amutation knife small | Amputation knife liston/catlin/langenbeck 13-16cm |
| 8 | Page 54 , BOQ point 4 | Bone cutter angled | Bone cutter angled 14-16cm |
| 9 | Page 54 , BOQ point 5 | Bone cutter double action | Bone cutter double action 25-27cm |
| 10 | Page 54 , BOQ point 6 | Bone file big | Bone file 22-24cm |
| 11 | Page 54 , BOQ point 7 | Bone file small | Bone file 16-18cm |
| 12 | Page 55 , BOQ point 12 | Bone lever lancer type plain | Bone lever lancer type plain 15-17cm. |
| 13 | Page 55 , BOQ point 13 | Bone lever lancer type serrated | Bone lever lancer type serrated 15-17cm |
| 14 | Page 55 , BOQ point 14 | Bone nibbler double action curved | Bone nibbler double action curved 22-24cm |
| 15 | Page 55 , BOQ point 15 | Bone nibbler double action straight | Bone nibbler double action straight 22-24cm |
| 16 | Page 55 , BOQ point 16 | Boyle mouth gag with 5 plates | MOLT or Similar Mouth gas 10-12cm & 14-16cm - 1 each |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| 17 | Page 55 , BOQ point 20 | Curette small | Curette medium 12-15cm |
| 18 | Page 55 , BOQ point 21 | Curette with handle medium | Curette with handle big 17-20 |
| 19 | Page 55 , BOQ point 23 | Hammer/Mallet medium | Hammer/Mallet small 22-24cm length, head dia 25-27 on one side & 13-15 on other side |
| 20 | Page 55 , BOQ point 24 | Hammer/Mallet small | Hammer/Mallet small 22-24cm length, head dia 38-40 on one side & 28-30cm on other side |
| 21 | Page 55 , BOQ point 25 | Hand saw medium | Charriere or Similar hand saw length 24-27cm |
| 22 | Page 55 , BOQ point 26 | Hand saw small | Langenbeck or Similar hand saw length 20-22 cm |
| 23 | Page 55 , BOQ point 27 | HM wire cutter | HM wire cutter 17-19cm |
| 24 | Page 55 , BOQ point 28 | Mouth gag (15 cm) | MOLT or Similar Mouth gas 14-16cm - 1 no. |
| 25 | Page 55 , BOQ point 33 | Periosteum elevator | Periosteum elevator 16-18cm |
| 26 | Page 55 , BOQ point 36 | Tooth extractor forceps | Tooth extractor set for different teeth |
| 27 | Page 55 , BOQ point 37 | Malliable retractors set | Deleted |
| 28 | Page 55,BOQ point 38 | Dental Extractor | Deleted |
| 29 | Page 55,BOQ point 40 | Bone - saw, file, nibbler | Bone saw Bone file 16-18cm Bone nibbler double action curved 14-16cm - 1 each |
| 30 | Page 55,BOQ point 41 | Drill set | Deleted |
| 31 | Page 55,BOQ point 42 | Malliable retractors Small | Deleted |
| 32 | Page 55,BOQ point 43 | Malliable retractors medium | Deleted |
| 33 | Page 55,BOQ point 44 | Malliable retractors large | Deleted |
| 34 | Page 55 , BOQ point 45 | Hammer mallet DF H4 | Hammer Mallet 200-250g (total weight) 18-20cm |
| 35 | Page 56 , BOQ point 46 | kockers clamp BH 614 | kockers clamp 14-16cm |
| 36 | Page 56 , BOQ point 47 | Cheasle FK 318 | Deleted |
| 37 | Page 56 , BOQ point 48 | Bone Gauge straight | Bone Gauge straight 8-10mm, 20cm |
| 38 | Page 56 , BOQ point 49 | Bone Gauge curved | Bone Gauge curved 8-10mm, 20cm |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| 39 | Page 56 , BOQ point 50 | Sharp two prong retractor BT 242 | Sharp two prong retractor 20-22cm |
| 40 | Page 56 , BOQ point 51 | Freer bone elevator OL-165 | Freer bone elevator 16-19cm |
| 41 | Page 56 , BOQ point 52 | periosteal knife OF-700 | periosteal knife |
| 42 | Page 56 , BOQ point 53 | Perkin antrum self retaining retractor | Perkin antrum self retaining retractor 125mm |
| 43 | Page 56 , BOQ point 54 | Langenbeck retractor | Langenbeck retractor 20-22cm, 10x40mm |
| 44 | Page 56 , BOQ point 55 | Hartman dressing forceps OK 242 nasal | Hartman dressing forceps 8-10cm |
| 45 | Page 56 , BOQ point 56 | Butterfly probe | Butterfly probe 12-14cm |
| 46 | Page 56 , BOQ point 57 | Retractor small | Retractor small 15-17cm |
| 47 | Page 56 , BOQ point 58 | Elevator big FK-133 | Elevator big 18-20cm |
| 48 | Page 56 , BOQ point 59 | Elevator small OM-307 | Elevator small 14-16cm |
| 49 | Page 56 , BOQ point 61 | S.S.Tray (medium) | Deleted |

Item No. 5 (Rfx/ Event number 3000003184)

Gynae Instrument Set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| 1 | page 57, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 57, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 57, point 4 | All needle holders & Metzenbaum scissors should be Tungsten Carbide coated | All needle holders & Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 57, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ S..No 8,9,10,14,15 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|---------|------------------------------|--|--|
| 5 | Page 57 , BOQ point 1 | SS tray medium | Deleted |
| 6 | Page 57 , BOQ point 2 | Sims' vaginal speculum big 409 | Sims' vaginal speculum size 1 |
| 7 | Page 57 , BOQ point 3 | Sims' vaginal speculum medium 408 | Sims' vaginal speculum size 2 |
| 8 | Page 57 , BOQ point 4 | Sims' vaginal speculum small 407 | Sims' vaginal speculum size 3 |
| 9 | Page 57 , BOQ point 5 | Sponge holder | Deleted |
| 10 | Page 57 , BOQ point 6 | Sims uterine probe malleable curved | Sims uterine probe malleable curved 30-32cm |
| 11 | Page 57 , BOQ point 7 | Bladder sound | Bladder sound 3mm, 4mm, 5mm, 6mm & 7mm – 1 no. each |
| 12 | Page 57 , BOQ point 8 | Endometrial biopsy curette | Endometrial biopsy curette 22-24cm |
| 13 | Page 57 , BOQ point 9 | Hysterectomy clamp St | Hysterectomy clamp St 20-22cm |
| 14 | Page 57 , BOQ point 10 | Cervical punch biopsy forceps shaft length 631 | Cervical punch biopsy forceps shaft length 20-22cm |
| | | Female pelvic Set | |
| 15 | Page 58, BOQ point 14 | Mangeshkar retractor | Mangeshkar Or Similar retractor |

Item No. 6 (Rfx/ Event number 3000003185)

Thyroidectomy Instruments set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|---------|------------------------------|---|---|
| 1 | page 59, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 59, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 59, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 59, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation | items mentioned at BOQ S..No 10,15,27g,27h,27k are to be demonstrated and feedback of |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|---|
| | | committee regarding satisfactory performance will be taken into account for technical approval. | evaluation committee regarding satisfactory performance will be taken into account for technical approval |
| 5 | Page 59 , BOQ point 1 | SS tray medium | Deleted |
| 6 | Page 59 , BOQ point 2 | Towel clips large | Deleted |
| 7 | Page 59 , BOQ point 3 | Towel clips medium | Deleted |
| 8 | Page 59 , BOQ point 4 | Towel clips small | Deleted |
| 9 | Page 60 , BOQ point 14 | Kelly's clamp | Kelly's clamp 12-14cm |
| 10 | Page 60 , BOQ point 15 | Kocher's clamp curved (medium | Kockers clamp curved 14-16cm |
| 11 | Page 60 , BOQ point 16 | Kocher's curved | Deleted |
| 12 | Page 60 , BOQ point 19 | Langenbeck retractor medium | Langenbeck retractor 20-22cm, 10-11x40mm |
| 13 | Page 60 , BOQ point 20 | Langenbeck retractor small | Langenbeck retractor 20-22cm, 15-16x30mm |
| 14 | Page 60 , BOQ point 21 | S.S.Bowl (large) | Deleted |
| 15 | Page 60 , BOQ point 22 | S.S.Bowl (medium) | Deleted |
| 16 | Page 60 , BOQ point 23 | Sponge holder | Sponge holder 22-24cm |
| 17 | Page 60 , BOQ point 27 | In Kidney tray (big) | Deleted |
| 18 | Page 60 , BOQ point 27.g | Adson vascular forceps | Adson vascular forceps medium |
| 19 | Page 60 , BOQ point 27.k | Metzenbaum scissors small | Metzenbaum scissors medium |

Item No. 7 (Rfx/ Event number 300003186)**General Surgery Instruments Set**

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | page 61, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 61, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 61, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 61, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | items mentioned at BOQ S..No 11,27b,27e,27g,27k,27n, are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 61, BOQ point 1 | SS tray large | Deleted |
| 6 | Page 61, BOQ point 2 | Towel clips large | Deleted |
| 7 | Page 61, BOQ point 3 | Towel clips medium | Deleted |
| 8 | Page 61, BOQ point 4 | Towel clips small | Deleted |
| 9 | Page 62 , BOQ point 20 | SS bowl large | Deleted |
| 10 | Page 62 , BOQ point 21 | SS bowl medium | Deleted |
| 11 | Page 62 , BOQ point 22 | Sponge holder | Deleted |
| 12 | Page 62 , BOQ point 23 | Suction tip | Suction tip 16-18cm |
| 13 | Page 62 , BOQ point 27 | In Kidney Tray (Big) | Deleted |
| 14 | Page 62 , BOQ point 27.n | Debakeys's vascular forceps | Debakeys's vascular forceps medium |

Item No. 8 (Rfx/ Event number 3000003187)**Mastectomy Instruments Set**

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | page 63, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | Page 63, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | Page 63, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 63, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ S..No 11,17,25b,25d,25f,25l are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 63, BOQ point 1 | SS tray medium | Deleted |
| 6 | Page 63, BOQ point 2 | Towel clips large | Deleted |
| 7 | Page 63, BOQ point 3 | Towel clips medium | Deleted |
| 8 | Page 63, BOQ point 4 | Towel clips small | Deleted |
| 9 | Page 64 , BOQ point 22 | S.S Bowl medium | Deleted |
| 10 | Page 64 , BOQ point 23 | S.S Bowl small | Deleted |
| 11 | Page 64 , BOQ point 24 | Sponge holder | Deleted |
| 12 | Page 64 , BOQ point 25 | In kidney tray (big) | Deleted |
| 13 | Page 64 , BOQ point 25.h | Debakey's vascular forceps | Debakey's vascular forceps 18-20 cm |
| 14 | Page 64 , BOQ point 25.o | Skin hooks | Skin hooks 14-16 cm |

Item No. 9 (Rfx/ Event number 3000003188)
Clip Applicator Set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | page 65, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | page 65, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | page 65, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 65, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Demonstration of medium & medium large clip applicators to be given and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 65, BOQ point 1 | SS tray large | Deleted |
| 6 | Page 65, BOQ point 2 | Towel clips large | Deleted |
| 7 | Page 65, BOQ point 3 | Towel clips medium | Deleted |
| 8 | Page 65, BOQ point 4 | Towel clips small | Deleted |
| 9 | Page 65, BOQ point 5 | Allis tissue holding forceps large | Deleted |
| 10 | Page 65, BOQ point 6 | Allis tissue holding forceps medium | Deleted |
| 11 | Added Para | | Added Para: Open Surgery clip applicator for use with small size clips – 1no. |
| 12 | Added Para | | Added Para: Open Surgery clip applicator for use with medium size clips – 1no. |
| 13 | Added Para | | Added Para: Open Surgery clip applicator for use with medium large size clips – 1no. |
| 14 | Added Para | | Added Para: Open Surgery clip applicator for use with large size clips – 1no. |

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 15 | Added Para | | Added Para: Titanium clips small, medium, medium large and large size for use with clip applicators mentioned |

Item No. 10 (Rfx/ Event number 3000003189)

Manual Bone Drill Set

| Sl. No. | Ref. to the Bidding Document | Existing specification/clauses in the Bidding Document | Amended as |
|----------------|-------------------------------------|---|--|
| 1 | page 66, point 2 | The Instruments Should be US-FDA or European CE with 4 digit notified body certified | The Instruments Should be US-FDA or European CE certified |
| 2 | page 66, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free and have antiglare property | Deleted |
| 3 | page 66, point 4 | All needle holders &Metzenbaum scissors should be Tungsten Carbide coated | All needle holders &Metzenbaum scissors should be Tungsten Carbide Insert |
| 4 | Page 66, point 8 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Items mentioned at BOQ S..No 2,3 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 5 | Page 66, BOQ point 1 | S.S Tray (medium) | Deleted |

All other contents of the Bidding Document including terms & conditions remain unaltered.