

Amendment No. 1**08.06.2018****Sub: Amendment to the Bidding Document****Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/19/18-19 dated 25.05.2018**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION - VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:**

**Item No. 1 (Rfx/ Event number 3000002860)
Transport Stretcher**

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amended as
1	Page 45 Point no 2	The mattress base should be made of compressed fiber sheet to make it durable, X-ray translucent and light weight. It should be atleast 4 inches thick with anti-static hygienic matters with pull straps.	The mattress base should be made of compressed fiber sheet or High pressure laminate to make it durable, X-ray translucent and light weight. It should be atleast 4 inches thick with anti-static hygienic matters with pull straps. It should be fitted in such a way so as to allow for easy cleaning and disinfection of the x-ray platform
2	Page 45 point no 9.4	Trendelenburg : 14 deg or better	Trendelenburg : 12 deg or more
3	Page 45 Point no. 10	High pressure laminate mattress base should be fitted in such a way so as to allow for easy cleaning and disinfection of the x-ray platform	DELETED

4	Page 45 Point no 18	Trolley should meet international quality directives body i.e. US-FDA/ European CE issued by notified bodies under directive 93/42/EEC Medical Devices. The O.E.M. should also have relevant BIS certifications.	Trolley should meet international quality directives body i.e. BIS/ US-FDA/European CE issued by notified bodies under directive 93/42/EEC Medical Devices. '
---	------------------------	--	--

**Item No. 2 (Rfx/ Event number 3000002861)
Paediatric Bed**

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amended as
1	Page 46 Point 3	Fixed Height: 600 mm (+/- 10%)	Minimum Height from base of the mattress should be between 540 mm to 660 mm.
2	Page 46 Point 8	Trendelenburg - 10° or more	Trendelenburg - 9° or more
3	Page 46 Point 9	Reverse Trendelenburg - 10° or more	Reverse Trendelenburg - 9° or more

**Item No. 4 (Rfx/ Event number 3000002863)
Chemotherapy Chair**

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amended as
1	Page no 48 Point no 16	Manufacturer should have ISO 9001 and ISO 13485 certification.	Manufacturer should have ISO 9001 or ISO 13485 certification.

**Item No. 8 (Rfx/ Event number 3000002867)
Wall Mounted Integrated OPD Diagnostic Set**

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amended as
1	Page 52 point no 2	Should be capable of measuring NIBP, Temperature, along with physical assessment like EYE (Ophthalmoscope) and EAR (Otoscope) with wall mounting option to save valuable floor space.	Should be capable of measuring NIBP, Temperature, along with physical assessment like EYE (LED Ophthalmoscope) and EAR (LED Otoscope) with wall mounting option to save valuable floor space.

2	Page 52 point no 5	Should have electronic thermometry with disposable probe cover to avoid cross contamination (Oral, Auxillary and optional rectal measurement).	Should have electronic thermometry with disposable probe cover to avoid cross contamination (Oral, Axillary and optional rectal measurement).
3	Page no 52	ADDED PARA -	BP Apparatus, Thermometer, Ophthalmoscope and Otoscope should have European CE or USFDA or BIS certification.

Item No. 9 (Rfx/ Event number 3000002868)

View Box

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amended as
1	Page no 53 Point no 7	It should have an on-off switch along with digital feather touch dimmer and a button to set the intensity	It should have an on-off switch along with digital feather touch /rotary dimmer to set the intensity.
2	Page no 53 Point no 12	10 step Digital dimmer facility with step up/step down intensity of 400 LUX.	DELETED

All other contents of the Bidding Document including terms & conditions remain unaltered.