

Amendment No. 3

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-III/10/URLG/17-18 dated 10-07-2017 read with Amendment No. 01 dated 26-07-2017 & Amendment No. 02 dated 26-07-2017

The following changes are being incorporated in the above referred Tender Enquiry Document

**Section VII
Technical Specification**

SCH 5. Laparoscopic Set		
Para	Existing Specification	Read As
6.m	CAMERA CONTROL UNIT & CAMERA Facility to directly record HD (High definition) 1920x1080p quality photos and HD (High definition) 1920x1080p quality videos directly through the hub into an USB device.	The Camera should be capable of directly recording full HD quality photos (1920x1080p) and full HD quality videos (1920x1080p) into an internal memory of 300 GB or more / into an external hard drive of 300GB or more. If not, a medical grade HD recorder capable of directly recording full HD quality photos(1920x1080p)and full HD quality videos(1920x1080p) with internal memory of 300GB or more /external hard drive of 300 GB or more must be supplied."
Sch 13 Camera system for Endourology		
1.m	CAMERA CONTROL UNIT & CAMERA Facility to directly record HD (High definition) 1920x1080p quality photos and HD (High definition) 1920x1080p quality videos directly through the hub into an USB device.	The Camera should be capable of directly recording full HD quality photos (1920x1080p) and full HD quality videos (1920x1080p) into an internal memory of 300 GB or more / into an external hard drive of 300GB or more. If not, a medical grade HD recorder capable of directly recording full HD quality photos(1920x1080p)and full HD quality videos(1920x1080p) with internal memory of 300GB or more /external hard drive of 300 GB or more must be supplied."

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**