

Amendment No. 10

Date: 26.02.2015

Subject: Amendment no. 10 to the Tender Enquiry Document

Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/20/14-15 dated 20.08.2014 and subsequent amendments published thereafter.

The pre-bid meeting for the referred tender enquiry was held on 27/08/2014. Based on pre-bid discussions following amendments are being incorporated in the referred tender enquiry document.

Section VII
Technical Specifications

Schedule No.2
Radiofrequency ablation (RFA) machine

1. Existing Specification:

Para: 2. iv. Output Power: **50 to 120 W**

Read as:

Para: 2. iv. Output Power: **50W or more**

2. Existing Specification:

Para: 3. i. Neutral plates (3 different sizes)

Read as:

Para: 3. i. **Single standard size neutral plate with one spare plate.**

3. Existing Specification:

Para: 3.iii. RF surgipens

Read as:

Para: 3.iii. RF surgipens (**Optional**)

Schedule No.3
SPECIFICATIONS FOR MOTOR (ELECTRIC) DERMABRADER

1. Existing Specification:

Para: 2. Speed dialling systems

Read as:

Para: 2. Speed dialling system **is required for adjusting required speed for specific procedure.**

2. Existing Specification:

Para: 4. Forward & Reverse switch

Read as:

Para: 4. **It should have both forward and reverse control.**

3. Existing Specification:

Para: 8. Electric Power: about 50 W

Read as:

Para: 8. **Deleted**

Schedule No.5
OPERATING BINOCULAR MAGNIFIER

1. Existing Specification:

Para: 1. Binocular loupe with **4x-10x** magnification.

Read as:

Para: 1. Binocular loupe **upto 6x** magnification.

2. Existing Specification:

Para: 2. Exceptionally bright and cool light with slim, light weight for coaxial, shadow free illumination of the working area.

Read as:

Para: 2. Exceptionally bright and cool **LED** light with slim, light weight for coaxial, shadow free illumination of the working area.

3. Existing Specification:

Para: 8. Should be BIS approved product.

Read as:

Para: 8. Should be **USFDA or European CE approved** product.

Schedule No. 6
Trinocular Microscope with camera attachment

1. Existing Specification:

Para: Observation tube **45 degrees**, IPD adjustment 50mm - 75 mm

Read as:

Para: Observation tube **30/45 degrees**, IPD adjustment 50mm - 75 mm

2. Existing Specification:

Para: High resolution Digital CCD Camera with resolution: **12.0 mega pixels**

Read as:

Para: High resolution Digital CCD Camera with resolution: **5.0 mega pixels** and **storage card of at least 8GB.**

3. Existing Specification:

Para: Objectives – Nose Piece **5 Position** to hold **5 objectives**

Read as:

Para: Objectives – Nose Piece **4 Position** to hold **4 objectives**

Schedule No.8
Equipment specification for hand foot UV phototherapy device

1. Existing Specification:

Para: 9. Should be FDA or CE approved product

Read as:

Para: 9. Should be **US FDA** or **European CE** approved product

Schedule No. 9
Vertical Autoclave

Read the Schedule name as “Horizontal Autoclave”

All other terms and conditions of the tender enquiry remain unaltered.