

## S.No:- 1. Full HD Video Endoscope System (2 nos.) with Endoscopic Ultrasound System (1 no.) (Rfx 3000003729)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para I. Full HD Gastro videoscope with accessories Page 45	Should have real time optical Chrome endoscopy imaging such as NBI/BLI/OE or equivalent	Karl Storz	S-technology should be added	No Change
Para I. Full HD Gastro videoscope with accessories Page 45	Should have forward/Auxiliary water jet for mucosal cleaning	Karl Storz	point should be amendmed as: Should have forward/Auxiliary water jet for mucosal cleaning or provision for water jet for mucosal cleaning	No Change
Para I.1.I. Full HD Gastro videoscope with accessories, Page 45	Total length: 1400mm or more	Karl Storz	Point should be deleted	No Change
Para2:Full HD colonovideoscope with accessories, Page 46	Should have real time optical Chrome endoscopy imaging such as NBI/BLI/OE or equivalent	Karl Storz	S-technology should be added	No Change
Para2.H:Full HD colonovideoscope with accessories, Page 46	Total length: 1900mm or more	Karl Storz	Point should be deleted	No Change
Para 3. Duodenovideoscope, Page 47	Suitable for real time optical Chrome endoscopy imaging such as NBI/BLI/OE or equivalent	Karl Storz	S-technology should be added	No Change
Para 5. Full HD video processor Page 49	should be compatible HD plus video scopes with real time optical Chrome should be endoscopy imaging such as NBI/BLI/OE or equivalent	Karl Storz	S-technology should be added	No Change
Para 5. Full HD video processor Page 49	The system should be compatible with Endoscopic ultrasound	Karl Storz	Point should be deleted	No Change

## S.No:- 1. Full HD Video Endoscope System (2 nos.) with Endoscopic Ultrasound System (1 no.) (Rfx 3000003729)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 5. Full HD video processor Page 49	Automatic IRIS Control & automatic white balance	Olympus	Kindly delete Automatic word from the statement	To be Amended as: Automatic IRIS Control & white balance
Para 5. Full HD video processor Page 49	Should have inbuilt or seperate light source with real time optical chrome endoscopy imaging such as NBI/BLI/OE or equivalent	Karl Storz	S-technology should be added	No Change
Para 5. Full HD video processor Page 49	Real time optical chrome endoscopy imaging such as NBI/BLI/OE or equivalent capability LED or 300W xenon light source with halogen/LED back up. Two xenon spare lamps to be supplied along with the machine in-case of xenon light source	Karl Storz	Point should be amended as "Real time optical chrome endoscopy imaging such as NBI/BLI/OE or equivalent capability <b>LED with a minimum lamp life of 25000 or more/</b> 300W xenon light source with halogen/LED back up. Two xenon spare lamps to be supplied along with the machine in-case of xenon light source	No Change
Para 5. Full HD video processor Page 49	Compatibility with all endoscopes and Endoscopic ultrasound system	Karl Storz	Point should be amended as: System should be compatible with all GI endoscopes	No Change
Para. II. Page 49	Technical Specifications of Endoscopic Ultrasound	Karl Storz	As EUS system is itself an independent system and always kept in a seperate suite hence we request to seperate this schedule from above GI endoscopy schedule. It will help the tender authorities to have maximum participation in GI schedule	No Change

S.No:- 1. Full HD Video Endoscope System (2 nos.) with Endoscopic Ultrasound System (1 no.) (Rfx 3000003729)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
BOQ S.no.A.5, Page 51	Full HD VIDEO PROCESSOR as per technical specifications - 2nos.	Karl Storz	Quantity of video processor & light source should be increased to atleast 3, as there should be minimum 2 cedicated suits are required to perform ERCP & Normal UGI/LGI/Sigmoidoscopy, EUS suite required a seperate independent tower including video processing, light source and monitor	No Change
BOQ S.no.A.6, Page 51	Full HIGH DEFINITION LCD MONITOR as per technical specifications- 2nos.	Karl Storz	Should be increased to 3 nos.	No Change
BOQ S.no.B.10, Page 51	Trolley as per technical specifications - 2nos.	Karl Storz	Should be increased to 3 nos.	No Change
BOQ S.no.B.11, Page 51	Computer system as per technical specifications-2nos.	Karl Storz	Should be increased to 3 nos.	No Change

S.No:- 2. Full HD Video Bronchoscope System with Endobronchial Ultrasound System (Rfx 3000003730)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para A Page52	The flexible bronchoscope should be supplied complete with light source and trolley and <b>minimun 15" LCD</b> monitor			To be Amended as: The flexible bronchoscope should be supplied complete with light source, trolley and monitor
Page 53	Radial probe	as per pre bid discussion	Request to delete	No Change
Under FULL HD Video Processor & Light source Page 52	Combined or separate LED/Xenon light source. Emergency backup facility in case xenon light source.	Fujifilm	Combined or separate <b>Multi LED/ 300W Xenon light source with halogen/LED backup</b> . Emergency backup facility in case xenon light source.	No Change
Under Para2. Linear probe Page.53	<b>Linear Probes</b> Depth of field : Approximately 2-50 mm	Fujifilm	Depth of field : Approximately 3-50 mm	To be Amended as: Depth of field : Approximately <b>3-50 mm</b>

Handwritten signatures and initials in blue ink, including names like 'Suresh', 'Jee', and 'Surya', along with various scribbles and checkmarks.

S.No:- 4. Rigid Cysto-Resectoscope (Rfx 3000003732)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 1.C Page 56	Should have Robust ceramic beak with guarantee during warranty period	Karl Storz	Kindly remove the line "guarantee during warranty period"	<b>To be Amended as:</b> Should have Robust ceramic beak to be covered under warranty. Replacement if needed to be done free of cost during the warranty period
Para 7 Page 57	ESU with 300W or more (of any reputed make) output power	Olympus	Kindly amend it from <b>SAME MAKE</b>	<b>To be Amended as:</b> ESU with 300W or more output power. <b>ESU should be of same make of cysto-resectoscope</b>
Para 7f Page 57	All compatible accessories for therapeutic & diagnostic hysteroscope work should be supplied along with ESU unit	as per pre bid discussion		To be Amended as: All compatible accessories for cysto-resectoscopy work should be supplied along with ESU unit
BOQ Sl.no.33 Page 58	Disposable patient plate pediatric	as per pre bid discussion		Item Deleted from BOQ

Handwritten signatures and initials in blue ink, including a large signature at the top center, several smaller initials, and a signature at the bottom right.

Response To Pre-Bid Queries (Pre-Bid date: 09.10.2018)

NIB Ref:HITES/PCD/NCI-AIIMS/39/18-19

S.No:- 5. Micro Vascular Instruments Set (Rfx 3000003733)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 5 Page 59	Page 59-Dedicated Slotted sterilization container with lock and silicon mat	Link Medical	Need to specify the size of the container and its quantity	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accomodate all the instruemnets for sterilization.
Para 8 Page 59	Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer	as per pre bid discussion		To be Amended as: Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Decleration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
Added Para BOQ Sl.no.22				Added BOQ Item at Sl.no.22: Sterilization container as per technical specificaltions - 1 set

Handwritten signatures and initials in blue ink at the bottom of the page, including a large checkmark and several illegible signatures.

S.No:- 6. Major Vascular Instruments Set (Rfx 3000003734)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 7 Page 61	Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer	as per pre bid discussion		To be Amended as: Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Decleration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
BOQ Sl.no..64 Page 63	Reynolds' vascular forceps small	Jolly Technocrats	vascular forceps small	To be Amended as: Reynolds' or similar vascular forceps small
BOQ Sl.no..65 Page 63	Reynolds' vascular forceps medium	Jolly Technocrats	vascular forceps medium	To be Amended as: Reynolds' or similar vascular forceps medium
BOQ Sl.no..66 Page 63	Reynolds' vascular forceps long	Jolly Technocrats	vascular forceps large	To be Amended as: Reynolds' or similar vascular forceps large

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
				Added Para: Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accomodate all the instruemnets for sterilization.
				Added BOQ Item: Sterilization container as per technical specificaitons - 1 set

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*



S.No:- 7. Thoracic Vascul ar Instruments Set (Rfx 3000003735)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 7 Page 64	Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer	as per pre bid discussion		To be Amended as: Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Decleration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
				Added Para: Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accomodate all the instruemnets for sterilization.
				Added BOQ Item: Sterilization container as per technical specifcaitons - 1 set

S.No:- 8. Plastic Surgical Instruments Set (Rfx 3000003736)

Tender Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
Para 7 Page 64	Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer	as per pre bid discussion		To be Amended as: Bidder should quoted for all the instruments of the set. Atleast 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Decleration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
				Added Para: Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accomodate all the instruemnets for sterilization.
				Added BOQ Item: Sterilization container as per technical specificalitons - 1 set

*[Handwritten signatures and initials in blue ink]*