

Amendment No.7

Date: 21.10.2014

Subject: Amendment no. 7 to the Tender Enquiry Document

- Ref:**
- (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/18/14-15 dated 20.06.2014**
 - (ii) Amendment No. 1 dtd. 30.07.2014**
 - (iii) Amendment No. 2 dtd. 16.08.2014**
 - (iv) Amendment No.3 dtd. 15.09.2014**
 - (v) Amendment No.4 dtd. 26.09.2014**
 - (vi) Amendment No. 5 dtd. 13.10.2014.**
 - (vii) Amendment No. 6 dtd. 18.10.2014.**

In subsequent to the Amendment No.6 published on 18.10.2014, the technical amendment of Schedule No. 16, Pentahead Microscope and Schedule No.17, Cytospin/Cytocentrifuge are as below:

Section VII **Technical Specifications**

Read as:

Para 5 : Vertical specimen stroke length : 60 - 70 mm

3. Added Para:

1. Demonstration of the quoted model is must.
2. Instrument should be supplied with minimum 150 high profile and 100 low profile disposable blades and 2 sets of brushes.

**Item No: 15
Cryomicrotome/Cryostat**

1. Existing Specification:

Para: 21. System should be quoted with Disposable Blade system:

Read as:

System should be quoted with Disposable Blade system of both high and low profile. Instrument should be supplied with minimum 150 high profile and 100 low profile disposable blades and 2 sets of brushes.

2. Added para:

1. Demonstration of the quoted model is must.
2. Embedding medium of 500ml should be supplied.
3. Facility of UV lamp decontamination should be available.

**Item No: 16
Pentahead Microscope**

1. Existing Specification:

Para: Illuminator : Built-in-Koehler illumination for transmitted light LED bulb (pre-centered) Light Intensity adjustment Centrally located so both hand can be used to increase And decrease light and with auto light intensity Adjustment with change of objective lens

Read as:

Para: Illuminator : Built-in-Koehler illumination for transmitted light LED bulb (pre-centered) and with auto light intensity Adjustment with change of objective lens

2. Existing Specification:

Para: Revolving nosepiece : Interchangeable/Removable Reversed Coded Quintuple Nosepiece for auto light adjustment



10/11/14
29/1/14
27
0/1/14

Read as:

Revolving nosepiece : Interchangeable/Removable Reversed Coded Quintuple Nosepiece for auto light adjustment with facility of toggle button to enable swiping of objectives (i.e Revolving nose piece should be automated/ motorized)

3. Existing Specification:

Para: Objectives : Plan 2,4x, 10X, 40X, & 100XOil

Read as:

Objectives : Plan 2/2.5x, 4/5x, 10X, 40X, & 100XOil

4. Existing Specification:

Para: Condenser : Swing out condenser (N.A 1.1), for 2X -100X

Read as:

Condenser: APPROPRIATE SWING OUT CONDENSER FOR 2X - 100X

5. Added Para:

Demonstration of the equipment is must.

**Item No: 17
Cytospin/Cytocentrifuge**

1. Existing Specification:

Para: Speed 100 to 4,000 rpm

Read as:

Para: Speed upto-3000rpm or more

2. Existing Specification:

Para: Processes about 80 samples per cycle with automatic chain-of-custody verification of patient samples


Read as:

Deleted

3. Added Para:

1. Demonstration of the quoted model is must.


DIRECTOR
All India Institute of Medical Sciences
Raipur 492019


Office
27/07/14
Dr. Anant Kungali
Asst. Professor
Pathology & Lab Medicine
AIIMS Raipur

Assoc. Professor,
Pathology & Lab Medicine
AIIMS, Raipur
Dr. [Signature]

All other terms and conditions of the tender enquiry remain unaltered.