

## **Amendment No. 3**

**17.05.2018**

### **Sub: Amendment to the Bidding Document**

**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/15/18-19 dated 13.04.2018 read with its Amendment No 1 and 2 dated 02.05.18 and 11.05.18 respectively.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

### **SECTION – VII**

#### **TECHNICAL SPECIFICATION AND GENERAL POINTS**

##### **Item Sl. No. 01 (Rfx/Event number 3000002805) State of Art Linear Accelerator**

- 1. Existing: Page 63, Para 15.5:** Quote the rates of necessary consumables valid for 5 years block of CMC period

**Amended as:** Deleted

- 2. Existing: Page 74, Para 12.4:** All the participating firms should quote the price of all required spares for upkeep and smooth functioning of the equipment for a period of 5 years.

**Amended as:** Deleted

- 3. Existing: Page 74, Para 1.3:-**Spare parts kit should be available for minimum of 10 years and price must be included in the offer

**Amended as:** Spare parts kit should be available for minimum of 10 years

- 4. Existing: Page 74, Para 1.4:-** During the warranty period, all the software updates and upgradation should be provided without asking for free of cost.

**Amended as:** During the warranty period, all the **software updates should be provided free of cost by the bidder.**

- 5. Existing: page 74, Para 1.5:-** Please quote the rates of necessary consumables recommended valid for 5 years block

**Amended as:** Deleted

**Item Sl. No. 02 (Rfx/Event number 3000002806)  
HDR Brachytherapy System**

- 1. Existing: Page 81, Para 7.1:** The vendor must quote for five years comprehensive warranty (including all spares and labour from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. The vendors must submit their quote (Rate) also for subsequent five years comprehensive AMC (including all Spares and labor) in the price bid, failure to comply this condition will entail the rejection of the bids.

**Amended as: Para 7.1:** The vendor must provide five years comprehensive warranty of the entire system (including all spares and labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. The vendors must submit their quote for subsequent five years comprehensive AMC (CAMC) (including all Spares and labor) in the price bid, failure to comply this condition will entail the rejection of their bid.

- 2. Existing: Page 82, Para 7.6:-**Quote the rates of consumables recommended valid for 5 years block.

**Amended as:-** Deleted

- 3. Existing:- Page 82, Para 5:** The bidder may quote the unit rates of any other site modification work activity which is not mentioned in the list below.

**Amended as:-** Deleted

**Item Sl. No. 03 (Rfx/Event number 3000002807)  
4D CT – Simulator**

- 1. Existing: Page 95, Para 23.17.iv:-** Lead Glass: 200 x 150 cm or more with Lead equivalent to meet safety requirements of AERB to be supplied.

**Amended as: -** Lead Glass: 200 x **120 cm** or more with Lead equivalent to meet safety requirements of AERB to be supplied

**All other contents of the Bidding Document including terms & conditions remain unaltered.**