

09-07-2018

Amendment No. 1**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/01-RT/18-19 dated: 23.06.2018**

The pre bid meeting of the above referred tender enquiry was held on 29.06.2018. Based on pre-bid discussions following amendments are being incorporated in the tender enquiry document.

Section VII
Technical Specifications

Schedule No. 01 - Balance Analytical capacity 200 gms and sensitivity 1/10 mg (Rfx no. 3000002947)		
Tender Page & Para	TENDER SPECIFICATION	READ AS
Page 42 Para 13	Added Para	Accuracy - 5%
Page 42 Para 14	Added Para	Minimum weight to be measured - 10mg

Schedule No. 03 - Centrifuge machine 5000 RPM (Rfx no. 3000002949)		
Tender Page & Para	TENDER SPECIFICATION	READ AS
Page 43 Para 3 3.6	Technical Specifications Maintenance free brushless drive motor with exact speed pre selection and display. Speed range 300 to 6000 rpm and above, accuracy 20 to 30 rpm.	Technical Specifications Maintenance free brushless drive motor with exact speed pre selection and display. Speed Range – 300-5000 RPM and above, accuracy 20 to 30 rpm.
Page 43 Para 3 3.7	Technical Specifications RPM: Up to 6500-7000	Technical Specifications Maximum RPM: at least 5000 RPM
Page 43 Para 4 4.1	System Configuration Accessories, spares and consumables Centrifuge complete with Swig and basic rotors and four buckets- 01 set.	System Configuration Accessories, spares and consumables Centrifuge complete with Swing out Bucket (04 Nos.) and basic rotors - 01 set.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**