

AMENDMENT IV
07.01.2014

SITC OF EQUIPMENT (PART A-Medical Equipments and PART B-Non-Medical Equipments) AT EMERGENCY MEDICAL SERVICES / CASUALTY DEPARTMENT IN JIPMER, PUDHUCHERRY

SITC OF EQUIPMENT (PART A-Medical Equipments and PART B-Non-Medical Equipments) AT EMERGENCY MEDICAL SERVICES / CASUALTY DEPARTMENT IN JIPMER, PUDHUCHERRY

Equipment 20 – Laparoscopy set					
Tender clause No.	Original tender specification	Amended tender clause No.	Reference Amendment No III:-	Amended tender clause No.	Latest Amendment:-
6.5	LCD based central display monitor with multilingual text & graphics	6.5	Standard medical grade monitor.	6.5	Standard LCD/LED screen to display various insufflator parameters