

Amendment No:1

to

Tender no. HLL/ID/17/03/TMC/MC Dated 27.06.2017

for

Procurement of mortuary cabinets

at

Polytrauma Care Building, Government Medical College, Thiruvananthapuram

The following amendments are incorporated in the referred tender enquiry document:

In Section – VII Technical Specifications		
SN	Existing Specification	Amended Specification
2.4	Exterior finish should be 0.5 mm SS Sheet or Pre-painted GI sheet.The material should be corrosion resistant and rust proof.	Exterior finish should be 0.5 mmSS.The material should be corrosion resistant and rust proof.
2.5	Insulation should be 0.5 mm polyurethane foam (CFC free) of density 40 kg/m ³ and minimum thickness of 2-4 inches.	Insulation should be 0.5 mm polyurethane foam (CFC free) of density 40 kg/m ³ and minimum thickness of 3-4 inches
2.15	Refrigeration unit failure alarm with battery operated remote unit should be supplied	Deleted
2.16	Strip chart recorder or circular chart recorder for recording 7 days temperature data should be supplied with 100 charts/equivalent rolls of paper	Deleted
2.17	Suitable 440 V,50 Hz electrical supply will be provided by the consignee at the site. All Internal electrical works , plumbing works or any other works required for smooth functioning of the mortuary cabinet has to be provided by the bidder.	Suitable 440 V,50 Hz electrical supply will be provided by the consignee at the site. All Internal electrical works including taking power supply from electrical room with safety accessories (3 phase DB,isolators,earthingetc),electrical cabling, plumbing works or any other works required for smooth functioning of the mortuary cabinet has to be provided by the bidder.

	Added para	3 inch drain should be provided by the bidder to drain out dirty water and condensate .The drain should be covered by suitable grating.
--	------------	---

END